



Date: - 16<sup>th</sup> January, 2018

Corrigendum  
For  
Tender for  
Paediatric Ventilator for the Department of Paediatric  
Surgery

NIT Issue Date	: 15 <sup>th</sup> September, 2017
NIT No.	: Admn/Tender/180/2017-AIIMS.JDH
Pre-Bid Meeting	: 27 <sup>th</sup> September, 2017 at 04:30 PM
Earlier Last Date of Submission	: 18 <sup>th</sup> January, 2018 at 03:00 PM
Extended Last Date of Submission	: 30 <sup>th</sup> January, 2018 at 03:00 PM
Bid opening	: 31 <sup>st</sup> January, 2018 at 03:30 P.M

**The following revised and additional specification will be added:-**

- 1. Page 10, Annexure-I, Point No. 3, Technical specification, Sub Point No. b:**  
**For**  
Should have colored Touch Screen, 10 Inch or more, should be able to turn and tilt the display monitor  
**Read**  
Should have colored Touch Screen, 17 Inch or more, should be able to turn and tilt the display monitor
- 2. Page 10, Annexure-I, Point No. 3, Technical specification, Sub Point No. cm In Point No. ii:**  
**For**  
2 loops P-V, F-V.  
**Read**  
3 loops P-V, F-V, F-P.
- 3. Page 10, Annexure-I, Point No. 3, Technical specification, Sub Point No. f, In Point No. i:**  
**For**  
Tidal Volume 5 ml to 2000 ml in volume control  
**Read**  
Tidal Volume 2 ml to 2000 ml in volume control
- 4. Page 12, Annexure-I, Point No. 10:**  
**For**  
Medical Air compressor USFDA and European CE Certified  
**Read**  
Medical Air compressor US FDA/ European CE Certified of same make.