



Date: - 27th November, 2017

Corrigendum
For
Tender for Microsurgery Drill System for the
Department of Neurosurgery.

NIT Issue Date	: 06 th October, 2017
NIT No.	: Admn/Tender/188/2017-AIIMS.JDH
Pre-Bid Meeting	: 16 th October, 2017 at 03:00 PM
Earlier Last Date of Submission	: 04 th December, 2017 at 03:00 PM
Extended Last Date of Submission	: 18 th December, 2017 at 03:00 PM
Bid opening	: 19 th December, 2017 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page 9, point 1

For

High Speed Electrical Drill system with Variable speed setting from 10,000 to $\geq 80,000$ rpm..

Read

High Speed Electrical Drill system with Variable speed setting with maximum speed more than 60000 rpm.

Page 9, two more points to be added after point 1.

- Minimum torque should be 40 mNm.
- Drill should be light-weight (less than 150 gms) and ergonomically designed for long duration surgeries.

2. Page 10, Heading II. Accessories, A. Perforator Set, point b

For

Reusable cranial perforators for adult and pediatric sizes (10 each) to be supplied.

Read

Disposable cranial perforators for adult and pediatric sizes (10 each) to be supplied

3. Page 10, Heading II. Accessories, B. Craniotomy Set, point c, item iii

For

Large: 05 (for c)

Read

Deleted

4. Page 10, Heading II. Accessories, Heading of C.

For

Osteotomy saw set

Read

Osteotomy Reciprocating saw Hand piece

5. Page 10, Heading II. Accessories, subheading C, point b.

For

It should also have tools to harvest bone grafts from ribs or iliac crest.

Read

Delete

6. Page 10, Heading II. Accessories, subheading C, point c

For

Two set of necessary blades to be supplied.

Read

Two saw blades each of short and long sizes (Total 04 in numbers) to be supplied

7. Page 10, Heading II. Accessories, subheading E, point b

For

Should have angled long (approx. 14 to 16 cm +/-1) and short (approx. 10 to 12 cm +/-1) attachments of less than 5mm outer thickness for use in trans-sphenoidal, trans-oral, micro-spinal and endoscopic surgeries (preferably of variable length) – one each,

- With corresponding cutting burrs of 1mm, 2mm and 4mm heads – 5 each
- With corresponding diamond burrs of 1mm, 2mm and 4mm – 5 each

Read

Should have angled long (approx. 14 to 16 cm +/-1) and short (approx. 10 to 12 cm +/-1) attachments of less than 5mm +/- 1mm outer thickness for use in trans-sphenoidal, trans-oral, micro-spinal and endoscopic surgeries (preferably of variable length) – one each,

- With corresponding cutting burrs of 2 mm, 3 mm and 4mm heads – 5 each
- With corresponding diamond burrs of 2mm, 3mm and 4mm – 5 each”

8. Page 10, Heading III. Other essentials: 4 points to be added

- a. Hand piece lubrication and cleaning solution- 6 in number
- b. Irrigation Tubing- 20 in number
- c. Imported cart to be supplied with the system.
- d. Demonstration of the product (part/full) may be demanded before approval.