



Date: - 27<sup>th</sup> November, 2017

Corrigendum  
For  
Tender for  
Transcranial Doppler for the  
Department of Neurosurgery

NIT Issue Date	: 06 <sup>th</sup> October, 2017
NIT No.	: Admn/Tender/190/2017-AIIMS.JDH
Pre-Bid Meeting	: 26 <sup>th</sup> October, 2017 at 03:00 PM
Earlier Last Date of Submission	: 04 <sup>th</sup> December, 2017 at 03:00 PM
Extended Last Date of Submission	: 18 <sup>th</sup> December, 2017 at 03:00 PM
Bid opening	: 19 <sup>th</sup> December, 2017 at 03:15 P.M

**The following revised and additional specification will be added:-**

- 1. Page 10, point 1:**  
**For**  
Four channel Transcranial Doppler capable of Intracranial; extra cranial, peripheral and Intraoperative use.  
**Read**  
Two channel Transcranial Doppler capable of Intracranial; extra cranial, peripheral and Intraoperative use
- 2. Page 10, point 6:**  
**For**  
Should have multi-gating.  
**Read**  
Should have multi-gating with at least 64 gates
- 3. Page 10, point 21**  
**For**  
Should have provision of connecting 7 external analog channels  
**Read**  
Delete.
- 4. Page 10, point 26:**  
**For**  
Should have power back up feature of more than 2 hours

**Read**

Should have built in Power back up feature of more than 1.5 Hours with recharging facility, Complete system should run on built in Battery, No battery outside the unit will be acceptable

**5. Following two (02) points to be added as points in the specifications**

- a. Tablet PC with below mentioned as minimum configuration: 13 inch or more HD Touchscreen, x360 Convertible, Intel core i7 Processor, 8 GB Memory / 256 GB SSD, 1 TB Hard Disk, at least 6 hours battery backup, Licensed Windows 10”
- b. “List of installation in India during last 5 years with contact details from all vendors for the similar brand and model of equipment as quoted in tender to verify the past performance”