



Date: - 27th November, 2017

Corrigendum
For
Tender for
Noiseless Suction Machine for the
Department of Neurosurgery

NIT Issue Date	: 06 th October, 2017
NIT No.	: Admn/Tender/192/2017-AIIMS.JDH
Pre-Bid Meeting	: 26 th October, 2017 at 03:00 PM
Last Date of Submission	: 27 th November, 2017 at 03:00 PM
Extended Last Date of Submission	: 18 th December, 2017 at 03:00 PM
Bid opening	: 19 th December, 2017 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page 10, point 1

For

Suction capacity - up to 60 lit/minute

Read

Suction capacity - up to 50 lit/minute

2. Page 10, point 2

For

Should have at least three variable flow rates so that machine can be used for various purposes in brain surgery

Read

To be deleted

3. Page 10, point 3

For

Integrated Foot on/off switch and vacuum regulator for hands free operation

Read

Integrated foot regulator and vacuum regulator for hands free operation