



Date: - 03<sup>rd</sup> August 2016

## Corrigendum

for

### Fowler Bed with Rails

NIT No.	: Admn/Tender/20/2016-AIIMS.JDH
NIT Issue Date	: 23 <sup>rd</sup> June, 2016
Pre Bid Meeting	: 04 <sup>th</sup> July, 2016 at 03:00 PM
Last Date of Submission	: 20 <sup>th</sup> July, 2016 at 03:00 PM
Extended Last Date of Submission	: 08 <sup>th</sup> August, 2016 at 03:00 PM
Revised Last Date of Submission	: 22 <sup>nd</sup> August, 2016 at 03:00 PM

#### 1. The following revised and additional specification will be added:-

**1. Page No. 10, Chapter-II Specification and Allied Technical Specification, Parameters and Technical Specifications for Executing the Work:, Point No. 01:**

**For**

All dimensions shall be read in metric unit only. Tolerance in overall dimensions upto  $\pm 5\%$ . No other deviations from specifications shall be allowed.

**Read**

All dimensions shall be read in metric unit only. Tolerance in overall dimensions upto **-2% to + 5%**. No other deviations from specifications shall be allowed.

**2. Page No. 10, Chapter-II Specification and Allied Technical Specification, Parameters and Technical Specifications for Executing the Work:, General Eligible Criteria, Point No. 02:**

**For**

Manufacturer should have valid European CE, ISO 9001:2008, ISO 14001:2004, and OHSAS 18001:2007 certificate. Dealers participating should enclose certificate from their parent manufacturer company.

**Read**

Manufacturer should have valid CE, ISO 9001:2008, ISO 14001:2004, and OHSAS 18001:2007 certificate. Dealers participating should enclose certificate from their parent manufacturer company.

**3. Page No. 14, Chapter-V, TECHNICAL SPECIFICATION FOR FOLWER BED WITH RAILS, Bullet No. 08:**

**For**

Should have easily removable head and foot panels made up of SS with four corner buffers.

**Read**

Should have easily removable head and foot panels made up of moulded polymer panels mounted in SS frame four corner buffers.

**4. Page No. 14, Chapter-V, TECHNICAL SPECIFICATION FOR FOLWER BED WITH RAILS, Bullet No. 13:**

**For**

Each bed must be quoted with one no, four section mattress of dimension (2000 mm X 900 mm) with washable cover of lasting quality. The mattress should be made of high density PU foam of 100 mm thickness.

**Read**

Each bed must be quoted with one no, four section mattress of dimension (2000 mm X 900 mm) with washable cover of lasting quality. The mattress should be made of 40 density and PU foam of 100 mm thickness. The mattress cover should be water resistant and fire retardant.

**Administrative Officer  
AIIMS, Jodhpur**