



Date: - 31st January, 2018

Corrigendum
For
Tender for
Decahead Microscope for the Department of
Microbiology

NIT Issue Date	: 09 th October, 2017
NIT No.	: Admn/Tender/207/2017-AIIMS.JDH
Pre-Bid Meeting	: 24 th October, 2017 at 03:30 PM
Earlier Last Date of Submission	: 05 th February, 2018 at 03:00 PM
Extended Last Date of Submission	: 20 th February, 2018 at 03:00 PM
Bid opening	: 21 st February, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Para No. 02:

For

Focus: Vertical stage movement 25mm per coarse stroke Vertical stage movement 1micron per fine stroke Stage rotation of 270 degrees with Stage Lock and Stage Tension adjustment.

Read

Focus: Vertical stage movement 24mm per coarse stroke Vertical stage movement 1micron per fine stroke Stage rotation of 270 degrees with Stage Lock and Stage Tension adjustment.

2. Page No. 10, Para No. 05:

For

Objectives: Plan 2x, 4x, 10X, 20X, 40X, & Plan Fluor 100XOil

Read

Objectives: Plan 2.5x, 5x, 10X, 20X, 40X, & Plan Fluor 100XOil

3. Page No. 10, Para No. 06:

For

Observation tube: Wide field Trinocular head with Field no. 22 mm or more with three Light path selection of 100:0, 20:80 and 0:100

Read

Observation tube: Wide field Trinocular head with Field no. 22 mm or more with three Light path selection of 100:0 and 0:100

4. Page No. 10, Para No. 09:

For

Teaching Attachment: For 1+ 9 persons Head with eyepiece of Field no. 22 or more, LED arrow pointer with variable intensity and with Green / Red colour selection. The equipment should be USA- FDA/European- CE approved

Read

Teaching Attachment: For 1+ 9 persons Head with eyepiece of Field no. 22 or more, LED arrow pointer with variable intensity and with Green, Red & white or more colour selection.