



Date: - 31st January, 2018

Corrigendum
For
Tender for
Shaker water Bath (Digital) for the Department of
Microbiology

NIT Issue Date	: 09 th October, 2017
NIT No.	: Admn/Tender/213/2017-AIIMS.JDH
Pre-Bid Meeting	: 24 th October, 2017 at 03:30 PM
Earlier Last Date of Submission	: 05 th February, 2018 at 03:00 PM
Extended Last Date of Submission	: 20 th February, 2018 at 03:00 PM
Bid opening	: 21 st February, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Point No. 5

For

External Construction MS Powder Coated (304 SS Optional)

Read

Epoxy powder coated Cold Rolled Steel

2. Page No. 10, Point No. 9

For

Insulation- Glass Wool

Read

Insulation Glass Wool or better

3. Page No. 10, After Point No. 17

Added Point No. 18:

Optimize scheduling with auto- on and auto- off timers

Help prevent bath overheating and damage with low- Fluid protection

Help protect your work with audible alarms.

Easily operate and monitor with icon-based graphically display