



Date: -15<sup>th</sup> December, 2017

Corrigendum  
For  
Tender for  
Pediatric Multipara Monitor for the Department of  
Pediatrics

NIT Issue Date	: 09 <sup>th</sup> October, 2017
NIT No.	: Admn/Tender/220/2017-AIIMS.JDH
Pre-Bid Meeting	: 23 <sup>rd</sup> October, 2017 at 03:00 PM
Earlier Last Date of Submission	: 11 <sup>th</sup> December, 2017 at 03:00 PM
Extended Last Date of Submission	: 01 <sup>st</sup> January, 2018 at 03:00 PM
Bid opening	: 02 <sup>nd</sup> January, 2018 at 03:15 P.M

**The following revised and additional specification will be added:-**

**1. Page No 10 , Point No 1 line 3**

**For**

Micro stream end tidal CO2 module in four monitors

**Read**

Mainstream end tidal CO2 module in four monitors

**2. Page No 10, Point No 3 , line No 5**

**For**

Should have both remote/ mouse control and knob control/ touch screen

**Read**

Should have both remote/ mouse control and knob control and touch screen

**3. Page No 13 Point No 23**

**For**

Quote separate price for web browsing facility in price bid. Also mention that how many beds can be connected with a single server.

**Read**

The web browsing should be a standard feature and the price should also be included at the time of price comparison after price bid opening.