



Date: - 22nd December, 2017

Corrigendum
For
Tender for
3 -D C Arm with Image Intensifier for the
Department of Neurosurgery

NIT Issue Date	:	06 th November, 2017
NIT No.	:	Admn/Tender/227/2017-AIIMS.JDH
Pre-Bid Meeting	:	17 th November, 2017 at 03:30 PM
Earlier Date of Submission	:	18 th December, 2017 at 03:00 PM
Extended Last Date of Submission	:	01 st January 2018 at 03.30 PM
Bid opening	:	02 nd January 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page 1, Heading of the tender.

For

Mobile 3-D C-Arm with Image Intensifier for the Department of Neurosurgery

Read

Mobile 3 D C- Arm with Flat Panel Detector for the Department of Neurosurgery

2. Page 10, table: 01 Item Heading

For

Mobile 3 D-Arm with Image Intensifier

Read

“Mobile 3 D C- Arm with Flat Panel Detector”

3. Page 10, table: Point 1.8:

For

Versatile, user friendly 3D recon facility with the compatibility to present image guidance system

Read

Versatile user friendly 3D recon facility with the compatibility to present image guidance system with 512 voxel resolution.

4. Page 10, table: Point 2:

For

Image Intensifier

Read

Flat Panel Detector