



Date: - 21st June, 2018

Corrigendum
For
High End Operating Microscope for the Department of
Neurosurgery

NIT Issue Date	: 07 th May, 2018
NIT No.	: Admn/Tender/232-2/2017-AIIMS.JDH
Pre-Bid Meeting	: 17 th May, 2018 at 03:00 PM
Earlier Last Date of Submission	: 28 th June, 2018 at 03:00 PM
Extended Last Date of Submission	: 04 th July, 2018 at 03:00 PM
Bid opening	: 05 th July, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 11, 8th point:

For

Coaxial illumination system with variable spot illumination.

Read

Coaxial illumination system with adjustment for illumination field.

2. Page No. 11, 18th point:

For

Should have Integrated Digital Video image recording, editing and archiving system mounted on the floor stand.

Read

Should have Integrated Digital Video image recording, editing and archiving system.

3. Page No. 11, 20th point:

For

Should have the inbuilt hardware and software for ICG fluorescence for vascular surgeries and ALA fluorescence for tumors (with YELLOW 560nm filter)

Read

Should have the inbuilt hardware and software for ICG fluorescence for vascular surgeries with YELLOW 560nm filter.

4. Page No. 11, addition of 26th Point:

All the accessories should be from the same manufacturer.