



Date: -19th January, 2018

Corrigendum
For
Tender for
Full HD Endoscopy System for the Department of
ENT

NIT Issue Date	: 30 th November, 2017
NIT No.	: Admn/Tender/235/2017-AIIMS.JDH
Pre-Bid Meeting	: 12 th January, 2017 at 03:30 PM
Last Date of Submission	: 30 th January, 2018 at 03:00 PM
Bid opening	: 31 st January, 2018 at 03:30 P.M

The following revised and additional specification will be added:-

1. Page No 10, Point No. 01, Sub Point No. B:

For

USB Port for Capturing HD Stills/ FULL HD Videos (desirable) in External USB drive and direct interface of USB Printer to facilitate direct printouts.

Read

USB Port for Capturing HD Stills as well as FULL HD Videos in External USB drive and direct interface of USB Printer to facilitate direct printouts.

2. Page No 10, In Technical Specifications, Para No. 04:

For

Lens:	Integrated Zoom Lens f = 15-31 mm (1.8x or above optical zoom)
--------------	--

Read

Lens:	Integrated Zoom Lens f = 13-16 to 29-32 mm (1.8x or above optical zoom)
--------------	---