



Date: - 12th February, 2018

Corrigendum
For
Tender for
Mechanical Ventilator for the Department of
Cardiology

NIT Issue Date	: 04 th December, 2017
NIT No.	: Admn/Tender/236/2017-AIIMS.JDH
Pre-Bid Meeting	: 14 th December, 2017 at 03:45 PM
Earlier Last Date of Submission	: 15 th February, 2018 at 03:00 PM
Last Date of Submission	: 27 th February, 2018 at 03:00 PM
Bid opening	: 28 th February, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 10, Point No. 02:

For

Compressed air / oxygen driven.

Read

Ventilator should have standalone compressor based system from same manufacturer (Turbine/ piston or inbuilt air-source within ventilators will not be accepted)

2. Page No. 11, Point No. 23:

For

Compressor should be US-FDA or European CE approved.

Read

Compressor should be US-FDA approved.

3. Page No. 11, After Point No. 24:

For

Anesthesia Machine with Integrated Monitor & Ventilator

The Machine should have the following:

Read

Deleted

4. Page No. 12, After Point No. 15.h:

Added Point No. i:

Upgradable to any one of the Advance modes like ASV/ Auto flow/ PAV+/NAVA/ smart care.

Price of upgradation should be quoted separately.

The whole section "Anesthesia Machine with Integrated Monitor & Ventilator

The Machine should have the following: "should be deleted.