



Date: - 30th April, 2020

Corrigendum
for
Specialised Open Instruments for the Department of
Obstetrics and Gynaecology

NIT Issue Date	: 24th February, 2020
NIT No.	: Admn/Tender/31-2/2019-AIIMS.JDH
Pre-Bid Meeting	: 02nd March, 2020 at 03:00 PM
Last Date of Submission	: 30th April, 2020 at 03:00 PM
Last Date of Submission	: 18th May, 2020 at 03:00 PM
Bid opening	: 19th May, 2020 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No. 10, S. No. 1 : ITEM: Specialized open instruments, point no. 4:**
For
Clip Applicator: Medium and large clips
Read as:
Clip Applicator: small, medium and large size compatible with 100mm, 200 mm, and 300mm Ethicon clips
- 2. Page No. 10, S. No. 1 : ITEM: Specialized open instruments, point no. 5:**
For
Vascular forceps 25 cm
Read as
Debakey's vascular forceps with 2mm tip
- 3. Page No. 10, S. No. 1 : ITEM: Specialized open instruments, point no. 7:**
For
Mosquito forceps:
Read as:
Curved Mosquito forceps
- 4. Page No. 10, S. No. 1 : ITEM: Specialized open instruments, point no. 8:**
For
Balfour's Retractor; Quantity 2
Read as;
Instrument Name: Balfour's Retractor with -
Specifications: Suprapubic abdominal retractor with locking mechanism and compass (self-holding), compass is a support with hinged arms. The valves move independently and engages with the compass. Lateral spread of fenestrated blades: 200mm. One each of following should be included:
Central Blade 75x 50m, Central Blade 70x100mm
Lateral Blade 60mmdeep, Lateral Blade 110mm deep
Quantity : 1

5. Page No. 10, S. No. 1 : ITEM: Specialized open instruments, point no. 10:

For

Omnitract retractor {pelvic and abdominal}

Read as:

Vein Retractor

Specifications: 21cm, 23 cm, 25 cm. Quantity: 2 each

6. Page No. 10, After Point No. 10:

Added Point No. 11:

All Instruments must be CE certified by a four numbered authority or USFDA approved.