



Date: -11<sup>th</sup> May, 2017

Corrigendum  
For  
Tender for  
CPAPBipap Machine for the Department of  
Cardiothoracic Surgery

NIT Issue Date	: 03 <sup>rd</sup> April, 2017
NIT No.	: Admn/Tender/46/2017-AIIMS.JDH
Pre-Bid Meeting	: 18 <sup>th</sup> April, 2017 at 03:00 PM
Earlier Last Date of Submission	: 18 <sup>th</sup> May, 2017 at 03:00 PM
Extended Last Date of Submission	: 08 <sup>th</sup> June, 2017 at 03:00 PM
Bid opening	: 09 <sup>th</sup> June, 2017 at 03:45 P.M

**The following revised and additional specification will be added:-**

**1. Page No. 12, In Technical Specification, Point No. 1:**

**For**

Usable for both adults and pediatric patients above 20kg body weight

**Read**

Usable for both adults and pediatric patients above 5kg body weight

**2. Page No. 12, In Technical Specification, Point No. 3:**

**For**

Should have multiple therapy modes i.e. CPAP, AUTOSET,BILEVEL, PAC (PRESSURE ASSIST CONTROL), SPONTANEOUS, S/T (SPONTANEOUS/TIMED), VAP (VOLUME ASSURED PRESSURE SUPPORT) etc

**Read**

Should have multiple therapy modes i.e. CPAP, AUTOSET,BILEVEL, PAC (PRESSURE ASSIST CONTROL), SPONTANEOUS, S/T (SPONTANEOUS/TIMED), VAP (VOLUME ASSURED PRESSURE SUPPORT), Auto E-PAP etc

**3. Page No. 12, In Technical Specification, Point No. 6:**

**For**

Pressure output should be upto 30cmH<sub>2</sub>O

**Read**

Pressure output should be 30 cmH<sub>2</sub>O or higher

**4. Page No. 12, In Technical Specification, Point No. 8:**

**For**

Should have humidification facility (with humidifier)

**Read**

Should be supplied with Integrated Humidifier

**5. Page No. 12, In Technical Specification, Point No. 10:**

**For**

LED display screen displaying all the set parameters like RR,EPAP,IPAP<Exhaled VT, Leak, mode of ventilation etc.

**Read**

**Colored** LED display screen displaying all the set parameters like RR,EPAP,IPAP<Exhaled VT, Leak, mode of ventilation etc.

**6. Page No. 12, In Technical Specification, Point No. 12:**

**For**

Accessories: Power cord, silicon Reusable (Autoclavable) 2 small size Full face masks, 3 medium size Full face masks, 2 large size Full face mask and 3 medium size nasal masks.

**Read**

Accessories **to be supplied with each unit:**

- (1.) Power cord,
- (2.) silicon Reusable (Autoclavable) **Two** small size Full face masks,
- (3.) silicon Reusable (Autoclavable) **Three** medium size Full face masks,
- (4.) silicon Reusable (Autoclavable) **Two** large size Full face mask and
- (5.) silicon Reusable (Autoclavable) **Three** medium size nasal masks.
- (6.) **One Wheel based stand (by same manufacturer will be preferred over other branded)**

**Cover for the Machine**

**7. Page No. 12, In Technical Specification, Point No. 16:**

**For**

Should have Spo2 Module and data Module.

**Read**

Should have Spo2 Module **with one adult and one pediatric probe** and data Module.

**8. Page No. 12, In Technical Specification, Point No. 18:**

**For**

Weight of the machine should be less than 2kgs (light weight preferred) and should be user friendly

**Read**

Weight of the machine should be less than **5** kgs (light weight preferred) and should be user friendly