



Date: -23<sup>rd</sup> June, 2017

Corrigendum  
For  
Tender for  
Rapid Infuser for the Department of Anaesthesiology

NIT Issue Date	:	11 <sup>th</sup> April, 2017
NIT No.	:	Admn/Tender/53/2017-AIIMS.JDH
Pre-Bid Meeting	:	24 <sup>th</sup> April, 2017 at 02:15 PM
Earlier Last Date of Submission	:	27 <sup>th</sup> June, 2017 at 03:00 PM
Extended Last Date of Submission	:	12 <sup>th</sup> July, 2017 at 03:00 PM
Bid opening	:	13 <sup>th</sup> July, 2017 at 03:30 P.M

**The following revised and additional specification will be added:-**

- 1. Page No. 11, In technical specification, Point No. 2:**  
**For**  
It should have automated pressure chambers (Minimum 02) providing constant pressure for rapid infuser.  
**Read**  
It should have at least one high speed peristaltic pump or pressure chambers for rapid infuser
- 2. Page No. 11, In technical specification, After Point No. 12:**  
**Added Point No. 13:**  
It should be USFDA approved
- 3. Page No. 11, In technical specification, After Point No. 13:**  
**Added Point No. 14:**  
10 Disposable kits to be provided with each equipment
- 4. Page No. 11, In technical specification, After Point No. 14:**  
**Added Point No. 15:**  
Rate should be quoted separately for disposable kits