



Date: - 15th December 2016

Corrigendum

for

Tender for Video assisted thoracoscopic surgery for the Department of General Surgery

NIT Issue Date	: 03 rd October, 2016
NIT No.	: Admn/Tender/57/2016-AIIMS.JDH
Pre-Bid Meeting	: 17 th October, 2016 at 03:00 PM.
Last Date of Submission	: 21 st November, 2016 at 03:00 PM.
Extended Last Date of Submission	: 03 rd January, 2017 at 03:00 PM.
Bid Opening	: 03 rd January, 2017 at 03:15 PM.

1. The following revised and additional specification will be added:-

1. Page No. 10, Annexure-I, Technical Specification, Point No. 01:

For

Should be Thoracoscopic Chitwood DeBakey Clamp (10mm), ring handle, 2x3 wide curved left DeBakey jaws, 10 cm jaw length, 10" /25.4 cm working length, 14 1/4" / 36 cm.

Read

Parenchymal Forceps, atraumatic, straight jaws, single action jaws, size 5 mm, length 28 cm, consisting of: Metal Handle with 4 locking positions, with ratchet, Outer Tube with Working Insert–Qty 1.

2. Page No. 10, Annexure-I, Technical Specification, Point No. 02:

For

Should be Thoracoscopic Chitwood DeBakey Clamp (10 mm), ring handle, 1x2 curved left DeBakey jaws, 10 cm jaw length, 10" / 25.4 cm working length, 14 1/4" / 36 cm.

Read

Parenchymal Forceps, atraumatic, double curved jaws, single action jaws, size 5 mm, length 28 cm, consisting of: Metal Handle with 4 locking positions, with ratchet, Outer Tube with Working Insert–Qty 2.

3. Page No. 10, Annexure-I, Technical Specification, Point No. 03:

For

Should be Thoracoscopic DeBakey Clamp (10mm), Harken #1, ring handle, 1x2

wide DeBakey jaws, without ratchet, 9.5 cm jaw length, 9-7/8" / 24.8 cm working length, 14" / 35.5 cm.

Read

Lung Forceps, atraumatic, fenestrated, curved jaws, single action jaws, size 5 mm, length 28 cm, consisting of: Metal Handle with 4 locking positions, Outer Tube with Working Insert–Qty 2.

4. Page No. 10, Annexure-I, Technical Specification, Point No. 04:

For

Should be Thoracoscopic DeBakey Clamp (10mm), ring handle, 1x2 wide straight tapered DeBakey jaws, with ratchet, 11.5 cm jaw length, 10-1/2" / 27 cm working length, 14-3/4" / 37.5 cm.

Read

Dissecting and Grasping Forceps, curved jaws, double action jaws, size 5 mm, length 28 cm, consisting of: Metal Handle with 4 locking positions, Outer Tube with Working Insert–Qty 1.

5. Page No. 10, Annexure-I, Technical Specification, Point No. 05:

For

Should be Thoracoscopic DeBakey Clamp (10mm), ring handle, 1x2 wide curved left tapered DeBakey jaws, with ratchet, 11 cm jaw length, 10" / 25.4 cm working length, 14-1/2" / 37 cm.

Read

Lung Nodule Forceps, atraumatic, fenestrated, curved jaws, single action jaws, size 5 mm, length 28 cm, consisting of: Metal Handle with 4 locking positions, Outer Tube with Working Insert–Qty 1.

6. Page No. 10, Annexure-I, Technical Specification, Point No. 06:

For

Should be Thoracoscopic Jacobson DeBakey Forceps (5 mm), spring style, round handle, 1 mm straight jaws, without ratchet, 7-3/4" / 19.5 cm working length, total length 16" / 41 cm.

Read

Grasping Forceps, Cobra-Jaws, 1 x 2 teeth, straight jaws, single action jaws, size 5 mm, length 28 cm, consisting of: Metal Handle with 4 locking positions Outer Tube with Working Insert–Qty 1.

7. Page No. 10, Annexure-I, Technical Specification, Point No. 07:

For

Should be Thoracoscopic Jacobson DeBakey Forceps (5 mm), spring style, round handle, 2 mm straight jaws, without ratchet, 7-3/4" / 19.5 cm working length, total length 16" / 41 cm.

Read

Grasping Forceps, straight jaws, single action jaws, size 5 mm, length 28 cm, consisting of: Metal Handle with 4 locking positions, Outer Tube with Working Insert–Qty 1.

8. Page No. 10, Annexure-I, Technical Specification, Point No. 08:

For

Should be Thoracoscopic Needle Holder (10mm), (for 2-0 and smaller needles), gold ring handle, straight carbon-inlay jaws, 8- 1/4" / 21 cm working length, 13-1/2" / 34 cm.

Read

Dissecting Forceps, insulated, curved jaws, double action jaws, size 5 mm, length 28 cm, with connector pin for unipolar coagulation, consisting of: Insulated Metal Handle with 4 locking positions, Insulated Outer Tube with Working Insert–Qty 1.

9. Page No. 10, Annexure-I, Technical Specification, Point No. 09:

For

Should be Thoracoscopic style, round handle, with ratchet, curved left fine DIAMOND DUST™ jaws, 6-1/2” / 16.5 cm working length, 15-3/8” / 38.5 cm Firm Sale.

Read

Scissors, insulated, straight jaws, curved scissor-blades, double action jaws, size 5 mm, length 28 cm, with connector pin for unipolar coagulation, consisting of: Insulated Metal Handle with 4 locking positions, Insulated Outer Tube with Working Insert–Qty 2.

10. Page No. 10, Annexure-I, Technical Specification, Point No. 10:

For

Should be Thoracoscopic Duval Clamp (10 mm), ring handle, curved left 1 cm width triangular ring jaws, with ratchet, 9-1/4” / 23.5 cm working length, 13-1/2” / 34 cm.

Read

Scissors, insulated, distally angled outer sheath, curved scissor-blades, scissor blades open horizontally to angulation, double action jaws, size 5 mm, length 28 cm, with connector pin for unipolar coagulation, consisting of: Insulated Metal Handle with 4 locking positions, Insulated Outer Tube with Working Insert–Qty 2.

11. Page No. 10, Annexure-I, Technical Specification, Point No. 11:

For

Should be Thoracoscopic Foerster Clamp (10 mm), ring handle, curved left 20 mm oval ring jaws, 9-1/2” / 24 cm working length, 13-1/4” / 33.5 cm.

Read

Scissors, insulated, distally angled outer sheath, straight scissor-blades, scissor blades open parallelly to angulation, single action jaws, size 5 mm, length 28 cm, with connector pin for unipolar coagulation, consisting of: Insulated Metal Y-Handle with 4 locking positions, Insulated Outer Tube with Working Insert–Qty 2.

12. Page No. 10, Annexure-I, Technical Specification, Point No. 12:

For

Should be Thoracoscopic Chitwood Dissecting Scissors (5 mm), curved left, blunt Super Cut™ blades, 3 cm blade length, black platinum ring handle, 7-1/4” / 18.5 cm working length, 11-1/2” / 29cm.

Read

Biopsy Forceps, insulated, distally angled outer sheath, single action jaws, size 5 mm, length 28 cm, with connector pin for unipolar coagulation, consisting of: Insulated Metal Y-Handle with 4 locking positions, Insulated Outer Tube with Working Insert–Qty 1.

13. Page No. 10, Annexure-I, Technical Specification, Point No. 13:

For

Should be Thoracoscopic Micro Scissors (5 mm), straight blunt Super Cut™ blades, 1 cm blade length, black platinum ring handle, 16.5 cm working length, 10-1/2” / 27 cm.

Read

Biopsy Forceps, insulated, straight, single action jaws, size 5 mm, length 28-35 cm, with connector pin for unipolar coagulation, consisting of: Insulated Handle, Insulated Outer Tube with Working Insert–Qty 1.

14. Page No. 11, Annexure-I, Technical Specification, Point No. 14:

For

Should be Thoracoscopic Metzenbaum Scissors (10 mm), curved left Super Cut™ blades, 3.5 cm blade length, black platinum ring handle, 7-3/4” / 19.5 cm working length, 11-1/2” / 29 cm.

Read

Surgical Sponge Holder, for atraumatic dissection of tissue layers, size 10 mm, length 30 cm, consisting of: Handle, Outer Sheath, insulated Sponge Holder Insert–Qty 1.

15. Page No. 11, Annexure-I, Technical Specification, Point No. 15:

For

Should be Thoracoscopic Needle Holder (7 mm), (for 4-0 and smaller needles), gold ring handle, straight carbon-inlay jaws, 8-1/2” / 21.5 cm working length, 13-1/2” / 34 cm.

Read

Dissecting forceps, , rotating, dismantling, with connector pin for bipolar coagulation, distally angled outer sheath, double action jaws, curved jaws, jaws open horizontally to angulation, size 5 mm, length 20 cm, color code: light blue consisting of: Outer Sheath with Forceps Insert qty-1.

16. Page No. 11, Annexure-I, Technical Specification, Point No. 16:

For

Should be Thoracoscopic Chitwood Knot Pusher (5 mm) (“trap model”), Jacobson spring style, round handle, for 2-0 and smaller suture, 6-1/8” / 15 cm working length, 12-1/2” / 32 cm.

Read

Dissecting Forceps, model, dismantling, with connector pin for bipolar coagulation, distally angled outer sheath, double action jaws, curved, jaws open horizontally to angulation, size 5 mm, length 28 cm, color code: light blue, consisting of: P Plastic Handle, with 4 ratchet positions Forceps Insert with Outer Sheath qty-1.

17. Page No. 11, Annexure-I, Technical Specification, Point No. 17:

For

Should be Thoracoscopic DeBakey Cooley Forceps, (10 mm), ring handle, 90° angled jaw, without ratchet, 9-1/2” / 24 cm working length, 13-1/2” / 34 cm.

Read

Scissors, model, dismantling, with connector pin for bipolar coagulation, distally angled outer sheath, single action jaws, straight scissor blades, scissor blades open vertically to angulation, size 5 mm, length 28 cm, color code: light blue consisting of: Plastic Handle, with 4 ratchet positions Scissors Insert with Outer Sheath qty-1.

18. Page No. 11, Annexure-I, Technical Specification, Point No. 18:

For

Should be Thoracoscopic Dennis Dissector (7 mm), ring handle, curved left serrated jaws, with ratchet, 9” / 23 cm working length, 13-1/4” / 33.5 cm.

Read

METZENBAUM Scissors, Model, rotating, dismantling, with connector pin for bipolar coagulation, curved jaws, more slender scissor blades, double-action jaws,

size 5 mm, length 20 cm, color code: light blue consisting of: Plastic Handle Metal Outer Sheath, insulated MW Scissors Insert qty-1.

19. Page No. 11, Annexure-I, Technical Specification, Point No. 19:

For

Should be US FDA APPROVED

Read

All items should be US FDA and European CE approved.

20. Page No. 11, Annexure-I, Technical Specification, After Point No. 19:

Added Point No. 20:

Telescope, 10 mm, length 32 cm, variable direction of view from 0° - 120°, twisting controller to select the desired view of direction, fibre optic light transmission incorporated . Qty 1.

21. Page No. 11, Annexure-I, Technical Specification, After Point No. 20:

Added Point No. 21:

Thoracopore size 11 mm, consisting of: Trocar only with blunt tip, Cannula without valve, with insufflation stop- cock, length 10.5 cm Automatic valve – qty1.

22. Page No. 11, Annexure-I, Technical Specification, After Point No. 21:

Added Point No. 22:

Thoracopore size 11 mm, autoclavable consisting of: Trocar, with blunt tip, Trocar Cannula, flexible, without valve, length 8.5 cm-Qty 2.

23. Page No. 11, Annexure-I, Technical Specification, After Point No. 22:

Added Point No. 23:

Thoracopore 6 mm, consisting of: Trocar with blunt tip, Cannula with thread, length 4 cm –Qty 4.

**Administrative Officer
AIIMS, Jodhpur**