



Date: - 07<sup>th</sup> September, 2018

Corrigendum  
For  
Fibre Optics Video Rhino Laryngoscope for the  
Department of Surgical Oncology

NIT Issue Date	: 25 <sup>th</sup> July, 2018
NIT No.	: Admn/Tender/67/2018-AIIMS.JDH
Pre-Bid Meeting	: 06 <sup>th</sup> August, 2018 at 04:00 PM
Earlier Last Date of Submission	: 04 <sup>th</sup> September, 2018 at 03:00 PM
Extended Last Date of Submission	: 19 <sup>th</sup> September, 2018 at 03:00 PM
Bid opening	: 20 <sup>th</sup> September, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No. 11, S. No. 1 In heading:**  
**For**  
Video Bronchoscope  
**Read**  
Flexible Video Bronchoscope
- 2. Page No. 11, S. No. 1, Video Bronchoscope, After Point No. 11:**  
**Add Point No: 12**  
Scope should be compatible with Narrow Band Imaging/Iscan-SE/BLL.
- 3. Page No. 11, S. No. 1, Video Bronchoscope, After Point No. 12:**  
**Add Point No: 13**  
Product should be European CE/ US FDA approved.
- 4. Page No. 11, S. No. 2, In heading:**  
**For**  
Video Processor/Light source:  
**Read**  
Video Processor & Light source:
- 5. Page No. 11, S. No. 2 Video Processor & Light source, Point No. 6:**  
**For**  
Facility to connect ancillary equipment like VCR, Video printer, Image filing system and recording system.  
**Read**  
Deleted

**6. Page No. 11, S. No. 2, Video Processor & Light source, After Point No. 10:  
Added Point No. 11:**

Product should be European CE/ US FDA approved.

**7. Page No. 12, S. No.3, High definition medical grade Monitor, Point No. 04:  
For**

All quoted product should be US FDA European CE approved.

**Read**

All quoted product should be US FDA/ European CE approved.