



Date: - 19th August, 2019

Corrigendum
For
Micro Drill and Saw System for the Department of
Surgical Oncology

NIT Issue Date	: 01 st July, 2019
NIT No.	: Admn/Tender/70-3/2018-AIIMS.JDH
Pre-Bid Meeting	: 11 th July, 2019 at 03:15 PM
Earlier Date of Submission	: 13 th August, 2019 at 03:00 PM
Extended Date of Submission	: 29 th August, 2019 at 03:00 PM
Bid opening	: 30 th August, 2019 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No. 10, S. No. 3(1)**
For
Maximum speed not less than 60000 RPM
Read
Maximum speed not less than 50000 RPM
- 2. Page No. 11, S. No. 6(c(1)):**
For
Maximum speed of 15000 RPM or more
Read
Maximum speed of 14000 RPM or more
- 3. Page No. 11, S. No. 14(iii):**
For
All attachments, handpiece etc. should be sterilisable through Steam and Flask Autoclave Prior demo to be provided if needed.
Read
All attachments, handpiece etc. should sterilisable through Steam and Flask Autoclave.
- 4. Page No. 11, S. No.15 : Addition of the following specification:**
Read
Generator Cart/ Trolley:
Generator cart or Trolley compatible with the above system from the same manufacturer should be provided.
- 5. Page No. 11, S. No.16: Addition of the following specification :**
Read
Prior Demonstration of the system is necessary.