



Date: - 12th November, 2018

Corrigendum -2
For
Tender for
Ultrasonic Cutting Device with Combined Generator
for the Department of Surgical Oncology

NIT Issue Date	: 25 th July, 2018
NIT No.	: Admn/Tender/73/2018-AIIMS.JDH
Pre-Bid Meeting	: 06 th August, 2018 at 04:30 PM
Earlier Last Date of Submission	: 15 th November, 2018 at 03:00 PM
Extended Last Date of Submission	: 29 th November, 2018 at 3:00 PM
Bid opening	: 30 th November, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. In Corrigendum, Page No. 3, Point No. 11:

For

17 / 20 cm shaft, curved/ straight, tapered tip for precise dissection, seals 7 mm vessels, as well as lymphatic with 16 mm active blade & 240-360 degree activation, triggers support multiple hand positions (Quantity-2 Nos.)

Read

17 / 20 cm shaft, curved/ straight, tapered tip for precise dissection, seals 5 mm or more vessels, as well as lymphatic with 16 mm active blade & 240-360 degree activation, triggers support multiple hand positions (Quantity-2 Nos.)