



Date: - 03<sup>rd</sup> January, 2016

## Corrigendum

for

### TMJ Arthroscopy unit with high definition camera for the Department of Dentistry

NIT Issue Date	:	27 <sup>th</sup> October 2016.
NIT No.	:	Admn/Tender/74/2016-AIIMS.JDH
Pre Bid Meeting held on	:	10 <sup>th</sup> November, 2016 at 03:00 PM
Last Date of Submission	:	30 <sup>th</sup> November, 2016 at 03:00 PM
Revised Last Date of Submission	:	17 <sup>th</sup> January, 2017 at 03:00 PM

#### **The following revised and additional specification will be added:-**

**1. Page No. 10, Point No. 2, ENDOVISION SYSTEM, Sub Point No. d:**

**For**

Should have High definition information management system. USB Port for Capturing FULL HD Videos/ HD Stills in External USB drive and direct interface of USB Printer to facilitate direct printouts.

**Read**

Separate FULL HD RECORDER SYSTEM with a minimum storage of 750 GB, USB DRIVE for copy function. Input –DVI, SDI, S-VIDEO. Output- HDMI. With a LED screen incorporated into the recorder.

**2. Page No. 10, Point No. 2, ENDOVISION SYSTEM, Sub Point No. e:**

**For**

Multiple Video output: Two DVI-D output, one 3G-SDI output, three camera input for communication with compatible camera modules, LAN connection and four USB connection.

**Read**

TWO DVI output, one S-Video output, Direct fiber output and keyboard input for character generator. Preferably USB connection should be provided.

**3. Page No. 10, Point No. 4, LED LIGHT SOURCE WITH FIBER OPTIC CABLE, Sub Point No. d:**

**For**

LED lamp life of minimum 50,000 Hrs or more.

**Read**

LED lamp life of minimum 20,000 Hrs or more.

**Administrative Officer  
AIIMS, Jodhpur**