



Date: - 19<sup>th</sup> December 2016

## Corrigendum

for

### Co2 Surgical Laser System for the Department of ENT

NIT Issue Date	:	08 <sup>th</sup> November 2016.
NIT No.	:	Admn/Tender/75/2016-AIIMS.JDH
Pre Bid Meeting held on	:	21 <sup>st</sup> November, 2016 at 03:15 PM
Last Date of Submission	:	20 <sup>th</sup> December, 2016 at 03:00 PM
Revised Last Date of Submission	:	18 <sup>th</sup> January, 2017 at 03:00 PM

#### 1. The following revised and additional specification will be added:-

1. Page No. 10, Technical Specification, Point No. 5:

**For**

IT SHOULD HAVE 5 MW RED DIODE AIMING BEAM, 635 NM, ADJUSTABLE INTENSITY

**Read**

IT SHOULD HAVE 3-5 MW RED DIODE AIMING BEAM, 635-650 NM, ADJUSTABLE INTENSITY.

2. Page No. 10, Technical Specification, Point No. 6, Sub Point No. 2:

**For**

A LIGHT WEIGHT CARBON DIOXIDE GLASS HOLLOW FIBRE, CO2 FIBRE SHOULD BE 2 METER LONG 1.04 MM OUTSIDE DIAMETER STERILE, SINGLE / MULTIPLE USE,

**Read**

A LIGHT WEIGHT CARBON DIOXIDE GLASS HOLLOW FIBRE, CO2 FIBRE SHOULD BE 2 METER LONG LESS THAN 1.1 MM OUTSIDE DIAMETER STERILE, SINGLE / MULTIPLE USE.

3. Page No. 10, Technical Specification, Point No. 7:

**For**

SPOT SIZE: 350 MM AT FIBER OUTPUT UP TO 40 WATT

**Read**

SPOT SIZE: 300  $\mu$ M OR LESS AT FIBER OUTPUT UP TO 40 WATT.

4. **Page No. 10, Technical Specification, Point No. 16:**  
**For**  
**For**  
IT SHOULD HAVE A 0.2 MM FOCUSED HAND PIECE.  
**Read**  
IT SHOULD HAVE A LESS THAN 0.4 MM FOCUSED HAND PIECE.
5. **Page No. 11, Technical Specification, ACCESSORIES, Point No. 2, ORAL, PHARYNGEAL AND NASAL HANDPIECE SET WHICH INCLUDES, Toggle Button No. 1:**  
**For**  
230MM HANDPIECE UNIT (CVD OPTICAL UNIT, PORTS HOLDER, M CONICAL MAIN EXTENDER, CONTAMINATION COLLECTOR), EXTRA CONICAL MAIN EXTENDER, BACKSTOP EXTENDER – 3 NOS, TIP EXTENDER – 3 NOS, STRAIGHT TIP, KAMAMI NASAL TIP – 3 NOS, KAMAMI TONSIL TIP – 3 NOS, 90 DEGREE ANGLED MIRROR TIP EXTENDER, CLEANING BRUSH, TYGON TUBE(8MM ID,1.5M LONG) W/ REDUCER FITTING.  
**Read**  
230MM HANDPIECE UNIT (CVD OPTICAL UNIT, PORTS HOLDER, CONICAL MAIN EXTENDER, CONTAMINATION COLLECTOR), EXTRA CONICAL MAIN EXTENDER, BACKSTOP EXTENDER – 3 NOS, TIP EXTENDER – 3 NOS, STRAIGHT TIP, KAMAMI NASAL TIP – 3 NOS, KAMAMI TONSIL TIP – 3 NOS, 90 DEGREE ANGLED MIRROR TIP EXTENDER, CLEANING BRUSH, TYGON TUBE(8MM ID,1.5M LONG) W/ REDUCER FITTING.
6. **Page No. 11, Technical Specification, ACCESSORIES, Point No. 4, FIBER ACCESSORIES, Toggle Button No. 1:**  
**For**  
REUSABLE CO2 FIBER- 01 NOS  
**Read**  
REUSABLE CO2 FIBER WITH MULTIPLE NUMBER OF SURGERIES - 05 NOS.
7. **Page No. 11, Technical Specification, In bottom:**  
**Added Point No. 1:**  
MINIMUM FIVE PAIRS OF LASER SAFE GLASSES SHOULD BE PROVIDED WITH THE UNIT.
8. **Page No. 11, Technical Specification, In bottom:**  
**Added Point No. 2:**  
PRICE OF REUSABLE CO2 FIBER AND LASER SAFE GLASSES SHOULD BE QUOTED SEPARATELY.

**Administrative Officer**  
**AIIMS, Jodhpur**