



Date: - 27th August, 2018

Corrigendum
For
Digital Dental Radiography, Radiovisiography (RVG)
for the Department of Dentistry

NIT Issue Date	: 25 th July, 2018
NIT No.	: Admn/Tender/76/2018-AIIMS.JDH
Pre-Bid Meeting	: 08 th August, 2018 at 03:30 PM
Earlier Last Date of Submission	: 04 th September, 2018 at 03:00 PM
Extended Last Date of Submission	: 11 th September, 2018 at 03:00 PM
Bid opening	: 12 th September, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

1. Page No. 11, Point No. 6 :

For

Resolution should be minimum 18 lp/mm.

Read

True Resolution should be minimum **18 lp/mm**

2. Page No. 11, Point No. 4 :

For

Sensor to directly connect to USB Port without any interface with 2- 3 meter cable length.

Read

Sensor to directly connect to USB Port without or with any interface with 2- 3 meter cable length