



अखिल भारतीय आयुर्विज्ञान संस्थान, जोधपुर
ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

Date: - 15th December 2016

Corrigendum

for

**Open Surgical Instruments for the Department of General
Surgery**

NIT Issue Date	:	27 th October 2016.
NIT No.	:	Admn/Tender/78/2016-AIIMS.JDH
Pre Bid Meeting held on	:	08 th November, 2016 at 04:00 PM
Last Date of Submission	:	30 th November, 2016 at 03:00 PM
Revised Last Date of Submission	:	20 th December, 2016 at 03:00 PM
Extended Last Date of Submission	:	05 th January, 2017 at 03:00 PM

1. The following revised and additional specification will be added:-

1. Page No. 10, Technical Specification, Point No. 7, Sub Point No. g:

For

Must supply suitable Autoclavable plastic double tray for sterilization and storage of at least 20 hand instruments preferably from OEM 1 in number.

Read

Must supply suitable Autoclavable mesh tray and wire mesh tray (full size 53x25x10 cm) with sterilization container and storage of at least 20 hand instruments preferably from OEM 10 in number.

**Administrative Officer
AIIMS, Jodhpur**