



Date: - 14th February 2017

Corrigendum
for
Laparoscopic Set for the Department of Urology

NIT Issue Date	:	18 th November 2016.
NIT No.	:	Admn/Tender/95/2016-AIIMS.JDH
Pre Bid Meeting held on	:	16 th January, 2017 at 03:30 PM
Earlier Last Date of Submission	:	16 th February, 2017 at 03:00 PM
Extended Last Date of Submission	:	01 st March, 2017 at 03:00 PM
Bid Opening	:	02 nd March, 2017 at 04:00 PM

The following revised and additional specification will be added:-

1. Page No. 11, In technical specification, Point No. 1, Line No. 03:

For

HD Optics for better contrast & colour reproduction.

Read

Should be compatible with full HD camera system

2. Page No. 11, In technical specification, Point No. 1, Line No. 06:

For

Autoclavable type and should be supplied with autoclavable tray.

Read

Autoclavable type and should be supplied with autoclavable tray. **Tray should be from original equipment manufacturer**

3. Page No. 11, In technical specification, Point No. 1.1, Line No. 1,2&3:

For

Other features:

System Should have compatibility with various Video scopes for future upgradability.

- System should have compatibility with flexible tip telescopes for Single port surgeries.
- System should be compatible with telescopes with Chip-on tip technology to provide bright & clear image.

Read

Deleted

4. Page No. 11, In technical specification, Point No. 2, Line No. 3:

For

HD Optics for better contrast & colour reproduction.

Read

Should be compatible with full HD camera system

5. **Page No. 11, In technical specification, Point No. 2, Line No. 6:**
For
Autoclavable type and should be supplied with autoclavable tray.
Read
Autoclavable type and should be supplied with autoclavable tray. **Tray should be from original equipment manufacturer**
6. **Page No. 11, In technical specification , Point No. 3, In Heading:**
For
Light Guide Cable, Must be Autoclavable: 2 No.
Read
Light Guide Cable, Must be Autoclavable: 2 No. (Diameter 4.5-5 mm)
7. **Page No. 11, In technical specification, Point No. 3, Line No. 6:**
For
Should be ROHS compliant.
Read
Deleted
8. **Page No. 11, In technical specification, Point No. 4, Line No. 3&4:**
For
- Automatic feedback control for any malfunction.
- Automatic Smoke evacuation facility with adjustable modes
Read
Deleted
9. **Page No. 12, In technical specification, Point No. 5, Sub Point No. m:**
For
Fan retractor 5mm 2 no
Read
Fan retractor 5mm 2 no (**CAN BE OF OTHER MANUFACTURER BUT SHOULD BE US FDA/ EUROPEAN CEA**)
10. **Page No. 12, In technical specification, Point No. 5, Sub Point No. n:**
For
Haem O'-lock clip endo-applier for large clips 10mm 2 no
Read
a) 1. Haem O'-lock clip endo-applier for large clips 10mm 1 no (**Should be imported**)
2. **Haem O'-lock clip endo-applier for extra-large clips 10mm 1 no (should be imported)**
11. **Page No. 12, In technical specification, Point No. 5, Sub Point No. r:**
For
a) Laparoscopic right angle dissecting forceps 5mm 2 no
Read
a) Laparoscopic right angle dissecting forceps 5mm 2 no (**CAN BE OF OTHER MANUFACTURER BUT SHOULD BE US FDA/ EUROPEAN CEA**)
12. **Page No. 12, In technical specification, Point No. 5, Sub Point No. v:**
For
a) Tray to accommodate all hand instruments 2 no
Read

- a) Tray to accommodate all hand instruments. Tray **should be original equipment manufacturer** 2 no.

13. Page No. 2, In technical specification, Point No. 6, Sub Point No. c:

For

- a. Laparoscopic extra peritoneal Hassan's balloon port: **2**

Read

- a. Laparoscopic extra peritoneal Hassan's balloon port: (**CAN BE OF OTHER MANUFACTURER BUT SHOULD BE US FDA/ EUROPEAN CEA**) **2**

**Administrative Officer
AIIMS, Jodhpur**