



Date: - 18<sup>th</sup> December, 2018

Corrigendum  
For  
Faculty Microscopes for the Department of  
Microbiology

NIT Issue Date	: 22 <sup>nd</sup> October, 2018
NIT No.	: Admn/Tender/95/2018-AIIMS.JDH
Pre-Bid Meeting	: 30 <sup>th</sup> October, 2018 at 04:00 PM
Earlier Last Date of Submission	: 13 <sup>th</sup> December, 2018 at 03:00 PM
Extended Last Date of Submission	: 27 <sup>th</sup> December, 2018 at 03:00 PM
Bid opening	: 28 <sup>th</sup> December, 2018 at 03:15 P.M

The following revised and additional specification will be added:-

- 1. Page No. 10, Point No. 4, equipment –Faculty Microscope , Under heading “Technical”:**  
**For**  
Revolving nose piece: Reversed sextuple revolving nose piece should be upgradable to DIC in future.  
**Read**  
Revolving nose piece: Reversed sextuple revolving nose piece with DIC slot.
- 2. Page No. 10, Point No. 7, equipment –Faculty Microscope , Under heading “Technical”:**  
**For**  
Stage: Ceramic coated surface mechanical stage with right hand low drive controlled with left hand for two specimens.  
**Read**  
Stage: Ceramic/Anodized coated surface mechanical stage with right hand low drive controlled with left hand for two specimens