



Date: - 14th February 2017

Corrigendum

for

Lower Tract Instruments for the Department of Urology

NIT Issue Date	:	09 th January 2017
NIT No.	:	Admn/Tender/96/2016-AIIMS.JDH
Pre Bid Meeting held on	:	16 th January, 2017 at 03:15 PM
Earlier Last Date of Submission	:	16 th February 2017 at 03:00 PM
Extended Last Date of Submission	:	01 st March, 2017 at 03:00 PM
Bid Opening	:	02 nd March, 2017 at 04:00 PM

The following revised and additional specification will be added:-

1. Page No. 11, In technical specification, Point A:

For

A. wo working channels with luer lock adaptors –Two

Read

A.1 Adult Telescope 4mm, 0⁰, enlarged view, autoclavable, fiber optic transmission incorporated, length 28-30 cm - One

A.2 Adult Telescope 4mm, enlarged view, autoclavable, fiber optic transmission incorporated 30⁰, length 28-30 cm - Two

A.3 Adult Telescope 4mm, 70⁰, enlarged view, autoclavable, fiber optic transmission incorporated, length 28-30 cm – One

2. Page No. 11, In technical specification, Point D:

For

Cystoscope sheath with obturator 14-15 fr with one working channels and bridge of one working channel with luer lock adaptor – one

Read

Cystoscope sheath with obturator **19-21 Fr** with one working channels and bridge of one working channel with luer lock adaptor – one

3. Page No. 11, In technical specification, Point T:

For

OTIS-MAUERMAYER Urethrotome, parallel expanding, length of dilating surface 16 cm- TWO

Read

OTIS-Mauermayer Urethrotome, parallel expanding, length of dilating surface **16-20 cm**- TWO

4. Page No. 11, In technical specification (accessories), Point 7:

For

Cold cup biopsy forceps compatible with 21-22 fr cystoscopic sheath- One

Read

Cold cup **Rounded Shaped** biopsy forceps compatible with 21-22 Fr cystoscopic sheath- One

- 5. Page No. 11, In technical specification (accessories), Point 7:**
Added Point No. 8:
OTIS Blades 6 in number

**Administrative Officer
AIIMS, Jodhpur**