



Date: -24th July, 2017

Corrigendum
For
Tender for
Green Diode Laser Console with Slit Lamp for the
Department of Ophthalmology

NIT Issue Date	: 11 th May, 2017
NIT No.	: Admn/Tender/97/2017-AIIMS.JDH
Pre-Bid Meeting	: 29 th May, 2017 at 03:45 PM
Earlier Last Date of Submission	: 24 th July, 2017 at 03:00 PM
Extended Last Date of Submission	: 11 th August, 2017 at 03:00 PM
Bid opening	: 12 th August, 2017 at 04:00 P.M

The following revised and additional specification will be added:-

1. Page No. 11, In In Technical Specification, Line No. 3:

For

Output power	50 to 1700 mW
--------------	---------------

Read

Output power	50 to 2500 mW
--------------	---------------

2. Page No. 11, In In Technical Specification, Line No. 7:

For

Aiming laser	Red diode, 635 nm, max. 0.2 to 1.0 mW
--------------	---------------------------------------

Read

Aiming laser	Diode Laser, 635 nm, max. 0.2 to 1500mW
--------------	---