



All India Institute of Medical Sciences Jodhpur

Dated: 24th December, 2013

CORRIGENDUM-2

Tender

For

Equipments required for Department of Pathology at AIIMS, Jodhpur.

NIT Issue Date : September 02, 2013.
Initial Date of Submission : October 10, 2013 at 03:00 PM
Revised Last date of Submission : 13th January, 2014 at 03:00 PM

List of Required Items:

S.No	NIT No.	Item Description	EMD (Rs)	Delivery Period (in Days)
1	Admin/General/146/2013-AIIMS.JDH	Storage Cabinet for Specimens	5,000	30
2	Admin/General/147/2013-AIIMS.JDH	Bone Cutter	9,000	30
3	Admin/General/164/2013-AIIMS.JDH	Centrifuge for Vacutainers- Floor Mounted	9,000	30

Technical Specifications

In original tender and corrigendum-I

Item No.2 Storage Cabinet for Specimen

Corrigendum-I, Under Technical Specification on Page No.: 12.

Point No.: 2

For

It should be possible to store a minimum of 1200 specimens.

Read

It should be possible to store a minimum of 900-1200 specimens.

Item No.3 Bone Cutter

Corrigendum-I, Under Technical Specification on Page No.: 12

Point No. 2:

For

Should be fitted with a large stationary table and extension table operated on four ball-bearing rollers.

Read

DELETED

Point No. 4:

Original tender, under technical Specification on Page No.9

For

Size of cutting table should be about 78 × 58 cms

Read

DELETED

Corrigendum-I, under technical specification on Page No.12, After Point No.9

ADDED Point No. 9:

The weight of the machine approx. 350-450Kg and working surface is divided into two halves with one half fixed and the other is capable of sliding to and fro.

Item No. 19 Centrifuge for Vacutainers- Floor Mounted

Under Technical Specification on Page No.: 22-23

Point No.1:

For

Depth: 80 cms, Height: 90 cms.

Read

Depth: 60-80cms, Height: 60-90cms.

Point No. 6:

For

Speed range: 300 to 21000 rpm

Read

Speed range: Upto 4000-5000 rpm.

Point No. 9:

For

The centrifuge rotor should be able to hold a minimum of 36 vacutainers at one time.

Read

Rotor head should be able to hold a tube of 15X75mm height and 28-36 vacuum tubes at one time.

**Administrative Officer
AIIMS Jodhpur**