



All India Institute of Medical Sciences Jodhpur

Dated: 05th July, 2016

Corrigendum & Date Extension

For

Tender for Supply of LED Double Film X-Ray View Box

NIT No.	:	Admn/Tender/18/2016-AIIMS.JDH.
NIT Issue Date	:	14 th June, 2016
Pre Bid Meeting held on	:	24 th June, 2016
Last Date of Submission	:	05 th July, 2016 at 03:00 PM.
New Last Date of Submission	:	19th July, 2016 at 03:00 PM.

Page No. 09:- Annexure - 1

Technical Specification for LED Double Film X-Ray View Box.

For Serial No.:07 -

It should have on-off switch along with digital feather touch dimmer and a button to set the intensity.

Changes:-

It should have on-off switch.

For Serial No.:08 -

It should have fully electronic continuous brightness control, with adjustment range of approximately 90%.

Changes:-

It should have individual panel for On/Off control and dimmer control.

For Serial No.:12 -

10 step Digital dimmer facility with step up/step down intensity of 400 LUX.

Changes:-

Deleted.

**Administrative Officer
AIIMS, Jodhpur**