



**ALL INDIA INSTITUTE OF MEDICAL SCIENCES,
JODHPUR

Dated: 24th September, 2021

CORRIGENDUM

Rate Contract for supply of Ophthalmology Consumables

At

AIIMS, Jodhpur

Tender No. : Admin/RC/03/2021-AIIMS.JDH
NIT Issue Date : 11th August, 2021
Last Date of Submission : 13th September, 2021 up to 03:00 PM
Date of Opening : 14th September, 2021 at 03:00 PM
Revised Last Date of Submission : 30th September, 2021 at 03:00 PM
Revised Date of Opening : 01st October, 2021 at 03:00 PM

Refer Page No. 27 > Annexure – VII > List of Ophthalmology Consumables:

T. R. No.	FOR		READ AS	
	Item Name	Specification	Item Name	Specification
6	Vicryl 6-0 reverse cutting suture (spatulated)	Material/Composition- polyglactin 910 Needle point type - cutting edge prime reverse	6-0 absorbable suture	Polyglactin 910 absorbable suture, double armed, 1/4 circle spatulated micro point, needle dimension 8.0mm, suture length 45 cm
8	Vicryl 7-0 reverse cutting suture		7-0 absorbable suture	Polyglactin 910 absorbable suture, 3/8 circle spatulated micro point, needle dimension 6.0mm, suture length 30 cm
9	Vicryl 8-0 reverse cutting suture (spatulated micro point)	Material/Composition- polyglactin 910 Needle point type - cutting edge prime reverse	8-0 absorbable suture	Polyglactin 910 absorbable suture, double armed, 3/8 circle spatulated micro point, needle dimension



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

T. R. No.	FOR		READ AS	
	Item Name	Specification	Item Name	Specification
				6.0mm, suture length 30 cm
31	Hydrophobic Aspheric Acrylic IOL Specification: Should be single piece Hydrophobic Acrylic Aspheric Optic with square edge design Haptic should have open loop L/C shaped design Should have UV absorbers, yellow chromophore Refractive index should be more than 1.45 Should have an overall size of 13 mm and optic size 6 mm Cartridge and injector system should be capable of impacting the IOL through 2.2 mm incision Should be US FDA / European CE certified.		Hydrophobic Aspheric Acrylic IOL Single piece Hydrophobic Aspheric, UV Blocking biconvex IOL with monofocal vision. Integrated C Haptics ProTec Frosted 360 degree posterior square edge with surface aberration-0.27, optic size 6mm, overall length 13 mm, Tri Flx Design. Refractive Index-1.47, USFDA and CE approved. Powers- +5D to +13D in 1.0 D increments, and +14D to +34D in 0.5 D increments.	
32	Hydrophobic Aspheric with yellow chromophore (Acrysof) Specification: Optic-haptic angulations of 5 degree, prevention from PCO, better centration, Aspheric, natural yellow, hydrophobic IOL asphericity of negative - 0.14 D better depth perception in mesopic and night condition, natural yellow IOL: blue light filtration, better circadian rhythm, Abbe value of 55, minimum chromatic aberration, incision size of 2.2 mm, approved "GLISTENING FREE" hydrophobic IOL.		Hydrophobic Aspheric Acrylic IOL Hydrophobic Aspheric with yellow chromophore Specification: Optic-haptic angulations of 5 degree, prevention from PCO, better centration, Aspheric, natural yellow, hydrophobic IOL asphericity of negative - 0.14 D better depth perception in mesopic and night condition, natural yellow IOL: blue light filtration, better circadian rhythm, Abbe value of 55, minimum chromatic aberration, incision size of 2.2 mm, approved "GLISTENING FREE" hydrophobic IOL. USFDA and CE approved. Powers- +5D to +13D in 1.0 D increments, and +14D to +34D in 0.5 D increments.	
35	Hydrophobic Multipiece IOL (Acrysof multipiece) Specification Optic Design: Spherical Monofocal Shape of the Optic (D): Anterior Asymmetric Biconvex Lens Toricity: Non Toric Lens Filters: Ultraviolet Light Lens Asphericity: Spherical Optical Principle: Refractive Lens Structure: Design: 3 Piece Lens Material: Hydrophobic Acrylic Haptic Design: Modified C Optic Diameter: 6.0 mm Overall Diameter: 13.0 mm A Constant: 118.4		Hydrophobic Multipiece IOL Hydrophobic acrylic multipiece, UV Blocking foldable IOL, Biconvex, Opti Edge Technology with 360 degree square edge, optic size- 6mm, overall length- 13mm with 5 degree angulation, A- constant- 119.1, USFDA and CE approved. Powers- +10D to +13D in 1.0 D increments, and +14D to +30D in 0.5 D increments.	
37	Micro surgical knives, MVR blade 1 mm (Straight & Angulated) without guard		Micro surgical knives, MVR blade	Straight with or without guard 20G
38	Micro surgical knives, MVR blade 1.2 mm (Straight & Angulated) without guard		Micro surgical knives, MVR blade	Straight with or without guard 19G
38A	Micro surgical knives, MVR blade 1.2 mm		Micro surgical knives, MVR blade	Straight with or without guard 23G



**ALL INDIA INSTITUTE OF MEDICAL SCIENCES,
JODHPUR**

T. R. No.	FOR		READ AS	
	Item Name	Specification	Item Name	Specification
	(Straight & Angulated) without guard			
39	Micro surgical knives, Keratome, 2.2 mm with convex tip design for tactile sensation, low angle (less than 25 degree), depth mark at 2 mm		Micro surgical knives, Keratome, 2.2 mm	Micro surgical knives, Keratome, 2.2 mm, without guard, angled, dual bevel
40	Micro surgical knives, 3.5mm, 3.2 mm, 2.8 mm & 2mm blade (dual bevel/Single Bevel), Blades with and without guard.		Micro surgical knives, Keratome	Straight or angled, with or without guard, micro surgical knives, Keratome 3.2mm, dual bevel
40A	Micro surgical knives, 3.5mm, 3.2 mm, 2.8 mm & 2mm blade (dual bevel/Single Bevel), Blades with and without guard.		Micro surgical knives, Keratome	Straight or angled, with or without guard, micro surgical knives, Keratome 2.8 mm, dual bevel
64	Healon, Ophthalmic Visco Elastic Device	Sodium Hyaluronate (pfs)	Ophthalmic Visco Elastic Device	Sodium Hyaluronate 1% PFS, 0.55 ml, 0.85 ml. USFDA approved
65	Healon GV 0.85ml	Sodium Hyaluronate (pfs)	Ophthalmic Visco Elastic Device	Sodium Hyaluronate 1.8% PFS, 0.85 ml, Molecular Weight- 3.2 million Daltons.
66	Healon 5	Sodium Hyaluronate (pfs)	Ophthalmic Visco Elastic Device	Sodium Hyaluronate 2.3% PFS, 0.60 ml. USFDA approved
67	Viscoat		Ophthalmic Visco Elastic Device	Sodium Hyaluronate 3% W/V + Chondroitin sulphate 4% W/V, 0.35 or 0.5ml prefilled syringe
99	GRIESHABER DSP 45D PIC 23G		DSP 45D PIC 23G	
100	GRIESHABER DSP 45D PIC 25G		DSP 45D PIC 25G	
104	REVOLUT.Autoclavable 25GA+ ENDGRAP		DELETED	



**ALL INDIA INSTITUTE OF MEDICAL SCIENCES,
JODHPUR**

T. R. No.	FOR		READ AS	
	Item Name	Specification	Item Name	Specification
107	REVOLUTION Autoclavable mini 23G CURVED SCISSOR		DELETED	
113	REVOLUTION Autoclavable 23G ENDGRASPING FORCEP		DELETED	
115	REVOLUTION Autoclavable 23G ILM FORCEPS		DELETED	
117	REVOLUTION Autoclavable 23G SERRATED FORCEP		DELETED	
119	REVOLUTION Autoclavable 23GA VERTICAL SCISSOR		DELETED	
121	REVOLUTION Autoclavable 25GA+ CURVED SCISSOR		DELETED	

Note: Above mentioned modifications will not be shown in BoQ, therefore, Bidders are advised to quoted price in front of old specification. **However, modified specifications will be considered for evaluation as per the details mentioned above.** In case, any new item(s) added through this corrigendum, the price for the same should be mentioned through addition of new sheet in BoQ. **Price submitted with the technical bid shall be liable to rejection of entire bid.**

**Deputy Director
(Admin)**