



**All India Institute of Medical Sciences
Jodhpur**

Date: - 22/07/2014

Corrigendum

Tender

For

Atomic Absorption Spectrometer

NIT Issue Date : July 08, 2014.
Pre Bid Meeting held on : July 15, 2014 at 04:00 PM
Last Date of Submission : July 30, 2014 at 03:00 PM

| S.NO | Item Description | NIT No. |
|------|--------------------------------|------------------------------|
| 1. | Atomic Absorption Spectrometer | Admn/Tender/3/2014-AIIMS-JDH |

The following revised and additional terms & conditions will be added:-

Terms & Conditions:

Point No. 6:

For

Equipment should be USFDA and European CE approved / certified.

Read

Equipment should be USFDA or European CE approved / certified.

Point Added After Pont No 37:

Point No. 38:

Sample: AIIMS Jodhpur reserves the right to ask the tenderers for submitting the sample of each item for which rates have been quoted, Technically Qualified Bidders may be asked to submit samples along with their quoted items nos. and their firm name without indicating any prices before opening of Financial Bid to AIIMS, Jodhpur for Inspection.

Technical Specifications, Annexure I

Point 8: Flame System, Sub Point (v)

For

Single element Coded Hollow Cathode Lamp for (Please put elements).

Read

Single element Coded Hollow Cathode Lamp for Pb, As, Hg, Cd, Se, Cu, Zn, Fe, Al.

Point 16:

For

Water Re-circulating Chiller (Imported) of compatible capacity suitable for supplied A. A. S. Unit.

READ

Water Re-circulating Chiller (Imported) of compatible capacity suitable for supplied A.A.S. Unit and UPS 20 KVA with ½hr backup.

Point Added After Point No 21:

Point No 22.

Submit price for all hollow cathode Lamps separately

**Administrative Officer
AIIMS Jodhpur**