



All India Institute of Medical Sciences Jodhpur

Date: - 07/10/2014

Corrigendum for Flash Sterilizer.

NIT Issue Date : September 11th, 2014.
Pre Bid Meeting held on : September 22nd, 2014 at 11:00 AM
Last Date of Submission : October 08th, 2014 at 03:00 PM
Revised Last Date of Submission : October 20th, 2014 at 03:00 PM

1. The following revised and additional terms & conditions will be added:-

1. Page No. 02, Point No. 6:

For

Equipment should be USFDA and European CE Certified/ Approved.

Read

Equipment should be USFDA **or** European CE Certified/ Approved.

2. The following revised and additional specification will be added:-

1. Page No. 10, Annexure-I, Specifications, Point No. 1:

For

It should be a fully automatic table top high pressure steam sterilizer with chamber volume of 20 L for sterilization of wrapped, unwrapped, hollow and porous materials and should have cycles for prions also.

Read

It should be a fully automatic table top high pressure steam sterilizer with chamber volume of **20 L or more** for sterilization of wrapped, unwrapped, hollow and porous materials and should have cycles for prions also.

2. Page No. 10, Annexure-I, Specification, Point No. 11

For

The Sterilizer should come with standard accessories like sterilization trays, holders, start up kit, traps of different sizes, Deionizer for connections to water, support for bags, Tray handles etc.

Read

The Sterilizer should come with standard accessories like sterilization trays, holders, start up kit, traps of different sizes.

3. Page No. 10, Annexure-I, Specification, Point No. 12:

For

It should be CE (European) and FDA (US) certified and should comply with EN 13060 standard for small sterilizers.

Read

It should be CE (European) **or** FDA (US) certified and should comply with EN 13060 standard for small sterilizers.

**Administrative Officer
AIIMS Jodhpur**