



Date: - 07/09/2013

CORRIGENDUM

Subject: - Tender for Anaesthesia Department at AIIMS, Jodhpur.

Sr. No	Name of Tender	Date of Issue.
1.	Tender for Anaesthesia Department (i) Patient Warning System (ii) Crash Cart	02/09/2013

A. **NIT No.** Admin/General/ 179/2013-AIIMS.JDH, **Page No.2**

Item Description

1. **For**
Patient Warning System
Read
Patient Warming System
Page No.22, S.No. 9 Item
2. **For**
Patient Warning System
Read
Patient Warming System

B. **NIT No.** Admin/General/ 180/2013-AIIMS.JDH, **Page No.2**

Item Description Crash Cart

1. **For**
EMD 2,000
Read
EMD 3,000

C. **Annexure-I Technical Specification**

1. Page No. 10, Item no.1, Anaesthesia Work Station,
Point no. 11, Anesthesia Ventilator

Following text to be removed:

Sub Point m: Should be able to ventilate with atmospheric air, in case of missing gases.



All India Institute of Medical Sciences Jodhpur

**2. Page No. 10, Item no.1, Anaesthesia Work Station,
Point no. 12, Airway Monitoring**

Following text to be removed:

Sub Point a: Monitor should be with multi-parameter module with minimum 15 inches colour TFT display with 8 channels.

Sub Point b: The monitor should not require any, lengthy start-up procedure or calibration. It should be ready to monitor as soon as on / off switch is pressed.

Sub Point c: Should have 24 hours graphical and numerical trend with split screen facility of all parameters with at least 15 critical alarms summary.

Sub Point d: Should be able to monitor and display all parameters in single screen.

**Administrative Officer
AIIMS Jodhpur**