



**ALL INDIA INSTITUTE OF MEDICAL SCIENCES,
JODHPUR**

Dated: 09th February, 2018

CORRIGENDUM

Rate Contract

for

Supply of Ward and ICU

Tender No.	:	Admn/RC/08/2017-AIIMS.JDH
NIT Issue Date	:	21st November, 2017
Last Date of Submission	:	26th December, 2017 at 03:00 PM
Date of Opening	:	27th December, 2017 at 03:00 PM
Revised Last date of Submission	:	26th February, 2018 till 03:00 PM
Revised Date of Opening	:	27th February, 2018 at 03:30 PM

Following items are added to above mentioned rate contract:

S. No.	Item Name	Specification
1.	I.V. Set with back check valve	
2.	CSE Kit 18G / 16G	Should have needle to needle technique without back eye on epidural needle. Should have locking mechanism with graduation marking on Epidural and Pencil Point spinal needle. Should have the marking on Epidural needle with 1 cm distance and marking should starts from 3 cm distance from the tip of Epidural needle
3.	Epidural kit with epidural catheter fixation device with locking mechanism 18G / 16 G	
4.	Double lumen closed suction system for endotracheal tube and for tracheostomy tube.	Should have additional accessories like MDI Adaptor, Trac-wedge device for disconnection of closed suction from the tube, Swivel connector, should have Double lumen allows suctioning and irrigation to be performed together.
5.	Tracheostomy tube with subglottic suction catheter along with 2 inner cannula with thin wall technique to reduce the ID by 1 cm of the tube	
6.	F4 Dialyser	
7.	Fx Paeds Dialyser	



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

S. No.	Item Name	Specification
8.	F5 Dialyser	
9.	Pediatric Size Dialysis Tubings	
10.	P1 Plasma Filter	
11.	HD Catheter	Should have Two lumens of Double-D design Should have laser cut side slots to avoid positional occlusion Should have clear silicon extensions tubing in both venous and arterial port Should have size 8 Fr. x 12 cm, 10 Fr. x 12 cm, 8 Fr. x 9 cm Should have 1-2 dilators in the pack Should have Straight or Curved extensions kits Should have USFDA approved
12.	Single lumen Haemodialysis femoral catheters	Straight, 14 FA – 5 ¼ Inches (13.3 cm), single use device, Should have USFDA / CE Approved.
13.	Peritoneal Dialysis Catheter Set – Paediatric Size	Should be 20 cm length, 12 Fr size with 4 cm of catheter tip has 32 lateral eyes. Should be available in adult size also – Length 30 cm, 12 Fr Size with 8 cm of catheter tip having 64 lateral eyes. Should have a stainless steel stylet Catheter should have rounded tip One connecting tube with clamp and injection port for additional medication. Should have USFDA / CE Approval.
14.	Peritoneal Dialysis Fluid 1.7 %	
15.	CTG Paper Roll	Quote all size separately
16.	IUI Cannula	
17.	Condoms	
18.	Safe Kit	Kit used for examination of sexual assault victims
19.	RO Filter	All size quote separately Must be compatible with Fresenius Medical Care WTU 125
20.	Dia safe plus filter for Hemodialysis Machine	Make: Fresenius Medical Care
21.	Skin protectant barrier film	Hexamethyldisiloxane, isooctane, acrylate terpolymer, polyphenylmethylsiloxane
22.	Dressing Kit for Peripheral I. V. Line	1. a Latex-Free Tourniquet 2. a CHG Swab 3. a Dressing 7 x 8.5 cm dressings (+- 3 mm) 4. a Sterile Gauze Swab
23.	Programmable Ventriculo peritoneal shunt	Medicated
24.	Pragrammable Ventriculo peritoneal shunt	Non-Medicated
25.	Medicated Ventriculo peritoneal shunt	Silver impregnated
26.	Fogarty embolectomy catheter for endoscopic Third Ventriculostomy	All size, quoted separately.
27.	Craniotomy Drape	



**ALL INDIA INSTITUTE OF MEDICAL SCIENCES,
JODHPUR**

S. No.	Item Name	Specification
28.	Neurosurgical Patties (Made of non-woven Fabric / Polyester microfiber with horizontal opaque line with black thread)	All size, quote separately.
15.	Silicon patties (Made of silicon sheet, electric shock proof, heat resistant, non-stick and see through) with black silk	All size, quote separately
16.	Dural Substitute	Collagen – based biocompatible dural substitute (Should exhibits tensile strength and wet handling capabilities) of available sizes
17	Troponin T Test	
18	Autoclave Label	<ul style="list-style-type: none">• It should have following feature• Process Date• Expiry Date• Sterilizer Number• Operator Number• Cycle Number
19.	Biological Indicators (Autoclave)	<ul style="list-style-type: none">• It should be Spore Strip, self-contained vial• Made up of large number of resistant bacterial spores.• It should be Malfunctioning sterilizers can be repaired and put back into service quickly.• Provide Auto reader free of cost as per requirement of AIIMS Jodhpur.

Following items are deleted from above said rate contract:

S. No.	Tender Ref. No.	Item Name	Specification
1.	1	Arterial Cannula with seldinger technique	16 G
2.	46	Epidural Catheter Fixation	
3.	135	Indicative Under Glove	Latex Powder Free, Color: Green / Blue
4.	140	Tracheostomy Tube	Cuffed and Uncuffed, All Size, Quote Separately
5.	158	Abdominal Sponge	With X-Ray detectable thread, All sizes, quote separately.

Administrative Officer