



Date: - 12th February 2016

Corrigendum
for
Equipments required for Department of
Forensic Medicine

NIT Issue Date	:	15 th January 2016.
NIT No.	:	Admn/Tender/Forensic Medicine/2015-AIIMS.JDH
Pre Bid Meeting held on	:	02 nd February 2016 at 03:00 PM
Last Date of Submission	:	18 th February 2016 at 03:00 PM
Revised Last Date of Submission	:	26 th February 2016 at 03:00 PM

1. The following revised and additional specification will be added:-

1. Page No. 02, Chapter-I, Air Purifier/ Odour Control System, Qty:

For

Qty- 2

Read

Qty- 3

2. Page No. 02, Chapter-I, Air Purifier/ Odour Control System, EMD (Rs.):

For

Rs. 10,000

Read

Rs. 15,000

3. Page No. 03, Point No. 7, Delivery and Installation, Sub Point No. ii), For goods imported directly from abroad:

For

All the goods ordered shall be delivered and Installed at AIIMS, Jodhpur within **60 days** from the date of opening of Letter of Credit for shipment.

Read

All the goods ordered shall be delivered and Installed at AIIMS, Jodhpur within **120 days** from the date of opening of Letter of Credit for shipment.

4. Page No. 11, Item No. 02, Air Purifier / Odour Control System, Qty:

For

Qty- 2

Read

Qty- 3

5. Page No. 11, Item No. 02, Air Purifier / Odour Control System, Point No. 2:

For

Spraying solution should be environment friendly, non-toxic, ozone safe and biodegradable

Read

Deleted

6. Page No. 11, Item No. 02, Air Purifier / Odour Control System, Point No. 2:

For

Spraying solution should be able to breakdown and neutralize odour causing bacteria and molecules

Read

Deleted

7. Page No. 11, Item No. 02, Air Purifier / Odour Control System, Point No. 4:

For

System should have at least four spraying units.

Read

System should be supplied with **One Unit** of 360 Degree ULV Fogger for Fumigation.

8. Page No. 11, Item No. 02, Air Purifier / Odour Control System, Point No. 5:

For

Spraying solution should be readily available on a recurring basis.

Read

Equipment should be supplied with 1 Liter of Fumigation solution and it should be readily available on a recurring basis.

9. Page No. 11, Item No. 02, Air Purifier / Odour Control System, Point No. 11:

For

Should be USFDA, European CE, UL or BIS approved product

Read

It should be USFDA or European CE approved.

10. Page No. 13, Item No. 04, Oscillating Electric Autopsy Saw (Bone cutting with blades and dust collector), Point No. 10:

For

- Suction performance: 1.380 watts.
- Air flow rate: 57l/min.
- Vacuum: 235 mbar.
- Main supply: 230 volts, 50 Hz.
- Noise level: less than 65 db.

Read

Deleted

11. Page No. 13, Item No. 04, Oscillating Electric Autopsy Saw (Bone cutting with blades and dust collector), BONE VACUUM DUST COLLECTOR SPECIFICATION:, After Point No. 7:

Add point no. 8

Suction performance: 1.380 watts.

Air flow rate: 57l/min.

Vacuum: 235 mbar.

Main supply: 230 volts, 50 Hz.

Noise level: less than 65 db.

12. Page No. 14, Item No. 06, Dissecting Lights (Double Floor Mounts), Point No. 07:

For

Irradiance / Intensity ratio: 3.3.

Read

Irradiance / Intensity ratio: <5.

13. Page No. 14, Item No. 06, Dissecting Lights (Double Floor Mounts), Point No. 08:

For

Average service life of the LEDs (h): > 40,000.

Read

Average service life of the LEDs (h): > 30,000

14. Page No. 14, Item No. 06, Dissecting Lights (Double Floor Mounts), Point No. 12:

For

It should be European CE and UL certified

Read

It should be USFDA and European CE certified.

15. Page No. 15, Item No. 07, Direct Mobile X-ray System, Add after Point No. 2:

Add Point No. 3:

It should be supplied with **4 (four)** plates.

**Administrative Officer
AIIMS, Jodhpur**