



All India Institute of Medical Sciences Jodhpur

Admin/General/105/2013-AIIMS.JDH

Date: - 06/09/2013

CORRIGENDUM

Subject: - Tender for supply of ICU Bed and accessories at AIIMS, Jodhpur.

Sr.No	Name of Tender	Date of Issue.
1	Tender for ICU Bed and accessories	26/08/2013

Following amendments are hereby authorized in the subject tender enquiry

a) CHAPTER VI-TECHNICAL SPECIFICATIONS OF ICU BED AND ACCESSORIES:

For

- Five section Polypropylene detachable tops fitted on four section top bed frame with perforated design for easy breathing of mattress.
- The base frame shall be mounted on 125 mm dia non-rusting twin wheel castor with central & directional locking mechanism.
- Built-in bed extension of approx. 200 mm shall be provided.

Read

- Four/Five section Polypropylene detachable tops fitted on four section top bed frame with perforated design for easy breathing of mattress.
- The base frame shall be mounted on 125 mm/150 mm dia non-rusting twin wheel castor with central & directional locking mechanism.
- Built-in bed extension of not less than 120 mm, but overall length should not be less than 2190 mm (Bed+extension).

Add

- Back rest & leg rest both shall have mattress guard.
- X-ray permeable back-rest with X-ray cassette holder.
- Bed should be having weight bearing capacity of at least 250 kg ($\pm 10\%$)

All other specification will remain same.

**Administrative Officer
AIIMS Jodhpur**