



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

Date: - 27/08/2013

CORRIGENDUM

Subject: - Tender for supply of Patient Care Utility at AIIMS, Jodhpur.

Sr.No	Name of Tender	Date of Issue.
1	Tender for Patient Care Utility	14/08/2013

Following Amendments are hereby authorized in the subject tender enquiry

a) Chapter I- Instruction to bidders :

Name of Furniture :

For Sr. No. 15 Examination couch (same as OPD) read as Examination Couch

b) Quantity required

Attendant Stool :	For 250	Read 320
Patient wheel chair fixed :	For 10	Read 14
Examination Couch:	For 15	Read 75

c) Chapter-VI- Technical Specifications Patient Care Utility

Item No. 01 Attendant stool

Add

Load bearing capacity should be at least 130 kg ($\pm 5\%$ will be acceptable)

Item No. 06 (Oxygen trolley)

For

Frame of the cylinder trolley is made with hi-grade MS tubular steel.

Read

Frame of the cylinder trolley is made with hi-grade SS tubular steel.

Item No. 15 (Examination Couch)

For

It should have sliding foot under the front side of lower middle cabinet.

Read

It should have sliding foot step with anti-skid property under the front side of lower middle cabinet.



All India Institute of Medical Sciences Jodhpur

Item No. 17 (Instrument trolley)

For

Overall approx. size: 680mm L x 450mm W x 900mm H.

Read

Overall approx. size: 680mm L x 450mm W x 900mm H. (+10% will be acceptable)

Add

Load bearing capacity should be at least 20 kg

Item No. 18 (Mayo's Trolley)

Add

+10 % Size will be acceptable.

Load bearing capacity should be at least 20 KG.

Item No. 19 (Cardiac table/Over bed table)

For

Overall size: 112 mm LX 450 mm W X 965 mm H

Read

Overall size: 1120 mm (L) X 450 mm (W) X 965 mm (H)

All other terms and conditions will remain same.

**Administrative Officer
AIIMS Jodhpur**