

Tender
For
Patient Care Utilities
At
All India Institute of Medical Sciences, Jodhpur

NIT No. : Admin/General/98/2013-AIIMS.JDH.
NIT Issue Date : Aug 14, 2013.
Last Date of Submission : Sep 06, 2013 at 03:00 PM.
Pre-Bid Meeting : Aug 22, 2013 at 04:00 PM



All India Institute of Medical Sciences, Jodhpur
Basni Phase - II, Jodhpur – 342005, Rajasthan
Telefax: 0291- 2740532, email: aoadmin@aiimsjodhpur.edu.in
www.aiimsjodhpur.edu.in



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

Chapter I- Instruction to bidders

Notice Inviting Tender

All India Institute of Medical Sciences, Jodhpur, Rajasthan, an apex healthcare institute being established by Parliament of India under aegis of Ministry of Health & Family Welfare, Government of India, invites sealed Tenders for supply & installation of the following item(s) at the institute. You are requested to quote your best offer along with the complete detail of specifications, terms & conditions.

Sr. No.	Name of the furniture	Qty required
1	Attendant stool	250
2	Patient stretcher trolley	20
3	Patient wheel chair Fixed	10
4	Saline/intravenous stand	40
5	Oxygen cylinder	40
6	Oxygen trolley	40
7	Stool with cushion & wheel	55
8	X ray view box	25
9	Glass door almirah	25
10	Iron almirah (closed)	25
11	Commode stool	2
12	Bedpan	40
13	Urinal	40
14	Anti-vibration table	2
15	Examination couch (same as OPD)	15
16	Gynaecology table	1
17	Instrument trolley	36
18	Mayo's Trolley	15
19	Cardiac table /Over bed table	175



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

1. Interested parties may send their tender in sealed cover addressed to the Administrative Officer, All India Institute of Medical Sciences, Basni Phase - II, Jodhpur superscripted with tender number and complete in all respects latest by Sep 06, 2013 up to 03:00 PM. The tender will be opened on the same day at 04.30 PM in the Project Cell, Residential Complex, All India Institute of Medical Sciences, Jodhpur. The tenders received after the scheduled date and time will be rejected out rightly.
2. The tender is in two- bid system i.e. Technical Bid & Financial Bid .The technical bid will be opened on the designated date by the Purchase Committee. The financial bid containing the rate of various item(s) will be opened on a suitable date, to be intimated later by the Purchase Committee in respect of those who qualify the terms and conditions of the technical bid.

The technical bid and the financial bid should be sealed by the bidder in separate covers superscripted "**Technical bid for Tender for Supply of Patient care utilities** "and "**Financial Bid for Tender for Supply of Patient care Utilities** ". Both Sealed Envelopes should be kept in a main/ bigger envelope superscripted as "**Tender for Supply of Patient care Utilities**". The 'Technical Bid" will be analysed and 'Financial Bid' of only those firms who are found eligible in 'Technical Bid' will be opened in due course and the eligible firms would be intimated there of accordingly.



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

Schedule of Tender

Issue Date	:	14 th Aug, 2013
Last date and time of receipt of tender	:	06 th Sep, 2013 at 03:00 PM
Amount of Earnest Money Deposit (EMD)	:	Rs 150,000/- (Rupees One Lakh Fifty Thousand only)
Tender Cost	:	Rs. 2000/- (Rupees Two Thousand only)
Date & time of opening of tender	:	06 th Sep, 2013 at 05:00 PM
Venue	:	Project Cell, AIIMS, Basni Phase-II, Jodhpur-342005.

4. Tender document may be downloaded from this Institute's official website "<http://www.aiimsjodhpur.edu.in>" and the tenderer shall deposit a separate Bank Draft/ Banker's cheque in favour of "All India Institute of Medical Sciences, Jodhpur" worth Rs 2000/- along with tender Document (Technical Bid). **The tenders submitted without tender cost or without EMD shall be liable to be rejected summarily.** The cost of the bid document is non-refundable.

Administrative Officer
Tel No. 0291-2740329.



Admin/General/98/2013-AIIMS.JDH

Chapter-II- Conditions of Contract

General Terms and Conditions

Subject: - Notice Inviting Tender for Supply of Patient care Utilities for All India Institute of Medical Sciences, Jodhpur

1. Earnest Money :

Earnest money by means of a Bank Demand Draft/ Pay Order of Rs 150,000/- (Rs. One Lakh Fifty thousand only) may be enclosed with the quotation (Technical Bid). It is also clarified that the quotations received without earnest money will be summarily rejected. The pay Order/DD may be prepared in the name of "All India Institute of Medical Sciences, Jodhpur".

- a) No request for transfer of any previous deposit of earnest money or security deposit or payment of any pending bill held by the ministry in respect of any previous work will be entertained.
- b) Tenderer shall not be permitted to withdraw his offer or modify the terms and conditions thereof. In case the tenderer fail to observe and comply with stipulations made herein or backs out after quoting the rates, the aforesaid amount of earnest money will be forfeited.
- c) Tenders without Earnest Money or Tender cost will be summarily rejected.
- d) No claim shall lie against the AIIMS in respect of erosion in the value or interest on the amount of EMD.

2. Preparation and Submission of Tender :

The tender should be submitted in two parts i.e. Technical Bid and Financial Bid. The Technical Bid and the Financial Bid should be sealed by the bidder in two separate covers "**Technical Bid for Tender for Supply of Patient care utilities**" and "**Financial Bid for Tender for Supply of Patient care utilities**". Both Sealed Envelopes should be kept in a main/ bigger envelope superscribed as "**Tender for Supply of Patient care utilities** "



Admin/General/98/2013-AIIMS.JDH

3. Signing of Tender :

The individual signing the tender or other documents connected with contract must specify whether he sign as:

- (a) A sole proprietor of the concern or constituted attorney of such sole proprietor ;
- (b) A partner of the firm, if it is a partnership firm in which case he must have authority to execute the contracts on behalf of the firm and to refer to arbitration disputes concerning the business of the partnership either by virtue of the partnership agreement or by a power of attorney duly executed by the partners of the firm.
- (c) Director or a principal officer duly authorized by the Board of Directors of the Company, if it is a company.

N.B.

(1) In case of partnership firms, a copy of the partnership agreement, or general power of attorney duly attested by a Notary Public, should be furnished on stamped paper duly sworn or affirmed by all the partners admitting execution of the partnership agreement or the general power of attorney. The attested copy of the certificate of registration of firm should also be enclosed along with the tender.

(2) In the case of partnerships firm, where no authority to refer disputes concerning the business of partnership firm has been conferred on any partner, the tender and all other related document must be signed by all partners of the firm.

(3) A person signing the tender form or any document forming part of the tender on behalf of another person should have an authority to bind such other person and if, on enquiry it appears that the person so signing had no authority to do so, AIIMS, Jodhpur may without prejudice, cancel the contract and hold the signatory liable for all costs, consequences and damages under the civil and criminal remedies available.

(4) **The tenderer should sign and affix his firm's stamp at each page of the tender and all its annexure as the acceptance of the offer made by tenderer will be deemed as a contract and no separate formal contract will be drawn.** NO PAGE SHOULD BE REMOVED/ DETACHED FROM THIS NOTICE INVITING TENDER.

4. Opening of Tender:

The tenderer is at liberty to present either him or authorize not more than one representative to be present at the opening of the tender. The representative present at the opening of the tender on behalf of the tenderer should bring with him a letter of authority from the tenderer and proof of identification.



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

5. Validity of the bids:

The bids shall be valid for a period of 120 day from the date of opening of the tender. This has to be so specified by the tenderer in the commercial bid.

6. Sample

- i. **AIIMS Jodhpur reserves the right to ask the tenderers for submitting the sample of each item for which rates have been quoted, Technically Qualified Bidders may be asked to submit samples along with their quoted items nos. and their firm name without indicating any prices before opening of Financial Bid to AIIMS, Jodhpur for Inspection.**
- ii. **The sample must confirm to specification given in Chapter-VI of the tender form.**
- iii. **Failure to submit sample on specified date & time will result in rejection of the tender.**

7. Right of acceptance:

AIIMS, Jodhpur reserve the right to accept or reject any or all tenders /quotations without assigning any reason there of and also does not bind itself to accept the lowest quotation or any tender. The tender quantity of any or all the items may be increased or decreased by AIIMS, Jodhpur.

Any failure on the part of the tenderer to observe the prescribed procedure and any attempt to canvass for the work will prejudice the tenderer's quotation or any tender.

8. Communication of Acceptance / Right of Acceptance :

AIIMS, Jodhpur, reserves all right to reject any tender including of those tenderers who fails to comply with the instructions without assigning any reason whatsoever and does not bind itself to accept the lowest or any specific tender. The decision of this Institute in this regard will be final and binding.

9. Performance Security :

The successful tenderer will be required to furnish a Performance Security Deposit of 10% of tender amount in the form of Fixed Deposit Receipt or Bank Guarantee from any Scheduled Bank in the name of "All India Institute of Medical Sciences, Jodhpur".

The security deposit can be forfeited by order of this Institute in the event of any breach or negligence or non-observance of any condition of contract or for unsatisfactory performance or non-observance of any condition of the contract.

Performance Security shall be submitted in the form of Bank Guarantee or Fixed Deposit Receipt issued by a scheduled Bank. Performance Security will be discharged after completion of tender's performance obligations (including Warranty/ Comprehensive maintenance contract period) under the contract.



Admin/General/98/2013-AIIMS.JDH

10. Delivery & Installation :

The items shall be delivered within 30 days of issue of supply order at AIIMS, Jodhpur. Satisfactory installation / commissioning and handover of the items will be completed within two weeks from the date of receipt of the furniture at the AIIMS, Jodhpur premises.

11. Inspecting ,Testing and Quality control

- I. A representative of AIIMS, Jodhpur will inspect and test the ordered goods and the related services to confirm their conformity to the contract specification and other quality control details incorporated.
- II. AIIMS, Jodhpur reserves the right for stage inspection during manufacturing process, pre-dispatch inspection or post-delivery inspection. The Purchaser as deemed fit shall resort to all these inspection or may waive off any of the inspection.
- III. Stage Inspection: The Supplier shall offer semi-finished product without paint for stage inspection supported with invoice & test report of major raw material as per relevant specification standards.
- IV. Before resorting to bulk manufacture, the supplier shall offer/send the prototype/actual finished sample for approval by the Representative of AIIMS, Jodhpur. The approval of samples shall be in respect of workmanship and finishing and shall be without prejudice to the rights of purchaser/user to get random testing of items during pre-dispatch inspection or from the actual lot offered. In the event of the failure of the supplier to deliver the sample by the date specified in the contract or any other date agreed by the purchaser or in the event of rejection of the sample second time, the purchaser shall be entitled to cancel the contract and purchase the stores at the risk and cost of the supplier.
- V. The Representative of AIIMS, Jodhpur shall have full and free access at any time during the execution of the contract to the supplier's work for satisfying himself that the goods are being manufactured in accordance with the specification mentioned in the Contract, and he may require the Supplier to make arrangements of inspection of the goods or any part thereof or any material at his premises or at any other place specified by the Representative of AIIMS, Jodhpur and if the Supplier has been permitted to employ the services of a sub-inspection of the goods after they have been inspected by the Representative of AIIMS, Jodhpur for the purpose aforesaid. The decision of the purchaser/ inspection authority in this regard shall be



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

final and binding on the Supplier. All terms and condition of the contract as they apply to the inspection shall also apply to the re-inspection.

- VI. The Supplier shall provide, without any extra charge, all material, tools, labour and assistance of every kind which the Representative of AIIMS, Jodhpur may demand of him for any test, and examination, which he shall require to be made on the Supplier's premises and the Supplier shall bear and pay all costs attendant thereon. If the Supplier fails to comply with condition aforesaid, the Representative of AIIMS, Jodhpur shall, in his sole judgement, be entitled to remove for test and examination all or any of the goods manufactured by the Supplier to any premises other than his (Supplier's) and in all such cases the Supplier shall bear the cost of transport/and carrying out such tests elsewhere. A certificate in writing of the Representative of AIIMS, Jodhpur, that the Supplier has failed to provide the facilities and the means for test and examination, shall be final.
- VII. The Supplier shall also provide and deliver for test, free of charge, at such place other than his premises as the Representative of AIIMS, Jodhpur may specify such material or goods as he may require for tests for which Supplier does not have the facilities or special/independent tests.
- VIII. The Representative of AIIMS, Jodhpur shall have the right to put all the goods or materials forming part of the same or any part thereof to such tests as he may think fit and proper. The Supplier shall not be entitled to object on any ground whatsoever to the method of testing adopted by the Representative of AIIMS, Jodhpur.
- IX. Goods accepted by the AIIMS, Jodhpur and/or its Representative of AIIMS, Jodhpur at initial inspection and in final inspection in terms of the contract shall in no way dilute AIIMS, Jodhpur's right to reject the same later.
- X. The Representative of AIIMS, Jodhpur reserves the right to draw samples through random sampling method after receiving the finished goods at destination. All the supplies shall be accompanied with the Manufacturer's certificate that the material conforms to the specifications.

Further to above if on the goods being rejected by the Representative of AIIMS, Jodhpur at the destination, the Supplier fails to make satisfactory supply within the stipulated period of delivery the AIIMS, Jodhpur shall be at liberty to:-

- a) Require the Supplier to replace the rejected goods forthwith but in any event not later than a period of 30 days from the date of rejection and the Supplier shall bear all cost of such



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

replacement, including freight, if any, on such replacing and replaced goods but without being entitled to any extra payment on that or any other account.

- b) Purchase or authorize the purchase of quantity of the goods rejected or goods of a similar description when goods exactly complying with particulars are not in the opinion of AIIMS, Jodhpur which shall be final, readily available without notice to the supplier at his risk and cost and without affecting the Supplier's liability as regards the supply of any further instalment due under the contract, or
- c) Cancel the contract and purchase or authorised the purchase of the goods or goods of a similar description at the risk and cost of the Supplier.
- d) The Representative of AIIMS, Jodhpur shall have the power:-
- Before any goods or part thereof are submitted for inspection to certify that they cannot be in accordance with the contract owing to the adoption of any unsatisfactory method of manufacture.
 - To reject any goods submitted as not being accordance with particulars.
 - To reject the whole of the instalment tendered for inspection, if after inspection of such portion thereof as he may in his discretion think fit, he is satisfied that the same is unsatisfactory.
 - To demand all cost incurred in the inspection and/or test. Cost of test shall be assessed at the rate charged by the Laboratory to private persons for similar work However if goods are accepted all cost incurred shall be borne by the AIIMS, Jodhpur.

Representative of AIIMS, Jodhpur's decision as regards the rejection shall be final and binding on the Supplier.

12. Guarantee / Warranty:

- The on-site replacement warrant shall remain for a period of 36 Months from the date of recording of acceptance of goods at site.
- During warranty period, the supplier is required to visit AIIMS, Jodhpur's site at least once in 4 months commencing from the date of delivery of the goods for preventive maintenance of the goods.
- If the supplier having been notified, fails to respond to take action to replace the defect(s) within 10 days the AIIMS, Jodhpur may proceed to take remedial action(s) as deemed fit.



Admin/General/98/2013-AIIMS.JDH

13. Liquidated Damages

If the supplier fails to deliver any or all of the goods or fails to perform the service within the time frame(s) incorporated in the tender, the AIIMS, Jodhpur shall, without prejudice to other right and remedies available to the AIIMS, Jodhpur under the tender, deduct from the quoted price, as liquidated damages, a sum equivalent to 0.5% per week of delay or part thereof on delayed supply of goods and/or services until actual delivery or performance subject to a maximum of 10 % of the quoted price .Once maximum is reached AIIMS, Jodhpur may consider termination of the tender.

14. FALL CLAUSE

If, at any time, during the said period, the supplier reduce the said prices of such Stores/Equipment or sales such stores to any other person/organization/Institution at a price lower than the chargeable, he shall forthwith notify such reduction or sale to the All India Institute of Medical Sciences (AIIMS) Jodhpur and the price payable for the Stores supplied after the date of coming into force of such reduction or sale shall stand correspondingly reduced.

The supplier shall furnish the following certificate along with bill for payment.

“I/We certify that the Stores of description identical to the Stores supplied to the AIIMS, Jodhpur under the contract against Tender herein have not been offered/sold by me/us to any other person/organization/Institution up to date of bill/the date of completion of supplies against all supply orders placed during the currency of the tender/rate contract at the price lower than the institute under contract /against tender”.

15. FORCE MAJEURE :

If, at any time during the subsistence of this contract, the performance in whole or in part by either party of any obligation under this contract is prevented or delayed by reasons of any war or hostility, act of public enemy, civil commotion, sabotage, fire, floods, explosion, epidemics, quarantine restriction, strikers lockout or act of God (hereinafter referred to as events) provided notice of happening of any such eventuality is given by party to other within 21 days from the date of occurrence thereof, neither party shall by reason of such event be entitled to terminate this contract nor shall either party have any claim for damages against other in respect of such non-performance or delay in performance, and deliveries have been so resumed or not shall be final and conclusive.

Further, that if the performance in whole or in part of any obligation under this contract is prevented or delayed by reason of any such event for a period exceeding 60 days, AIIMS, Jodhpur party may, at least have option to terminate the contract.



Admin/General/98/2013-AIIMS.JDH

15. Insolvency etc.:

In the event of the firm being adjudged insolvent or having a receiver appointed for it by a court or any other order under the Insolvency Act made against them or in the case of a company the passing any resolution or making of any order for winding up, whether voluntary or otherwise, or in the event of the firm failing to comply with any of the conditions herein specified AIIMS, Jodhpur shall have the power to terminate the contract without any prior notice.

16. Breach of Terms and Conditions:

In case of breach of any terms and conditions as mentioned above, the AIIMS, Jodhpur will have the right to cancel the work order/ job without assigning any reason thereof and nothing will be payable by AIIMS, Jodhpur. In that event the security deposit shall also stand forfeited.

17 Subletting of Work:

The firm shall not assign or sublet the work/job or any part of it to any other person or party without having first obtained permission in writing of AIIMS, Jodhpur, which will be at liberty to refuse it thinks fit. The tender is not transferable. Only one tender shall be submitted by one tenderer.

18. Right to call upon information regarding status of work:

AIIMS, Jodhpur will have the right to call upon information regarding status of work/ job at any point of time.

19. Terms of payment:

- 90% payment of the order value shall be released after the successful installation of the ordered goods against the submission of the inspection report.
- Balance 10% of the order value shall be released after the submission of the performance security. If the company fail to submit his performance security the 10% payment shall be released after 3 year from the date of opening of tender.

20. Arbitration:

If any difference arises concerning this agreement, its interpretation on payment to be made thereunder, the same shall be settled out by mutual consultation and negotiation. If attempts for conciliation do not yield any result within a period of 30 days, either of the parties may make a request to the other party for submission of the dispute for



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

decision by an Arbitral Tribunal containing Sole Arbitrator to be appointed by the Secretary, Department of Legal Affairs. Such requests shall be accompanied with a panel of names of three persons to act as the sole arbitrator. In case of such arbitrator refusing, unwilling or becoming incapable to act or his mandate having been terminated under law, another arbitrator shall be appointed in the same manner from among the panel of three persons to be submitted by the claimant. The provision of Arbitration and Conciliation Act, 1990 and the rule framed there under and in force shall be applicable to such proceedings.

21. Legal Jurisdiction:

The agreement shall be deemed to have been concluded in Jodhpur, Rajasthan and all obligations hereunder shall be deemed to be located at Jodhpur, Rajasthan and Court within Jodhpur, Rajasthan will have Jurisdiction to the exclusion of other courts.

"PRE –BID CONFERENCE" with the intending bidders shall be held on 22nd Aug 2013 from 04:00 P.M. onwards at AIIMS, Jodhpur.

Administrative Officer
Tel No. 0291-2740329.



Admin/General/98/2013-AIIMS.JDH

Chapter III - Specifications and allied Technical Details

Parameters and Technical Specifications for Executing the Work:

1. All dimensions shall be read in metric unit only. Tolerance in overall dimensions ± 10 mm. No other deviations from specifications shall be allowed.
2. All the mild steel sections shall be finished in epoxy polyester powder coating in broken white colour.
3. The manufacturer shall have In-house powder coating facility, and in-house test laboratory for testing of
 - a. Impact test
 - b. Bend test
 - c. Salt spray chamber test is must.
 - d. Powder Coating - Min 8 dip tanks pre-treatment and powder epoxy coating of minimum 60 microns with phosphate layer underneath for corrosion resistance.
 - e. Phosphate coating with test certificate.
4. Min 12 mm MDF wood with laminate of 1 mm thickness wherever used and seen with normal viewing.
5. Finishing & workmanship in the furniture is of prime importance and must be of high standard.
6. All corners shall be rounded off so that there shall be no sharp corners.
7. The supplier shall ensure that all the fitting/accessories used are of high quality.
8. Rubber/Plastic shoes of good quality shall be used wherever necessary.
9. Stainless steel sheets / tubes mentioned in the specification shall be of SS. 304 grade.
10. Tenderer shall produce test certificate of the SS 304 grade at the time of inspection.
11. Castor wheels shall be as per tender specification & shall have rustproofed.
12. The material shall be packed properly in polythene and then in corrugated boxes.
13. The tenderer has to get the samples approved as per tender specification from the AIIMS, Jodhpur.
14. The participant: manufacture/companies having own manufacturing units or their authorized dealer can only participate. Proof of manufacturing is to be attached. A representative of AIIMS, Jodhpur may also visit and inspect the manufacture set up as deemed fit.



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

15. Company /manufacturer/firm must have local branch office in the city or within Rajasthan. (Enclosed Address details). Copy of shop and establishment act registration certificate to be enclosed for giving proof of establishment.
16. Items should have USFDA/ European CE certified.
17. Manufacturer should have :
 - a) an ISO 9001:2008 certification.
Or
 - b) an ISO 14001:2004 certification.
Or
 - c) an ISO 18001:2007 certification.
Or
 - d) BIFMA Membership/registration.
18. The firm should be registered and should have the average turnover of Rs. 75,00,000/- (Rs. Seventy five lakh) for the last three consecutive years. The firm has to submit a proof of turnover as mentioned above supported by documentary materials.
19. The delivery of the items will have to be made at AIIMS, Jodhpur. No transportation/ carriage charges will be provided for the same.
20. The rate quoted should be firm and final and written in ink or typed against each item and should in no case be overwritten.

Administrative Officer,
AIIMS, Jodhpur.



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

Chapter - IV Contract Form

TENDER FORM - 1 - TECHNICAL INFORMATION AND UNDERTAKING.

(Tenderer may use separate sheet wherever required)

Sr.No	Details of the Firm/Bidder	Page No.	
1.	Name, Address, Phone, Mobile and email address of the Tenderer		
2.	State clearly whether it is Sole proprietor or Partnership firm or a company or a Government Department or a Public Sector Organisation		
3.	Name and address of service centre nearby Jodhpur		
4.	Details of the Earnest Money Deposit (EMD) worth Rs. 1,50,000/- (Rs. One lakh Fifty Thousand only)		
5.	Details of the cost of the Tender documents worth Rs. 2000/- (Rs. Two Thousand only)		
6.	Whether each page of NIT and its annexure have been signed and stamped		
7.	Whether items have USFDA/ European CE certification		
8.	Whether manufacturer have ISO certification		
9.	List of Major Customer may be given on a separate sheet and proof of satisfactory supply, if any		
10.	Proof of the last three year's average turnover of the firm which should not be less than Rs. 75,00,000/- (Rupees Seventy five only) continuously for the preceding three years.		
11.	Permanent Account Number		
12.	TIN No. with Proof		
13.	Whether copies of authenticated balance sheet for the past three years enclosed		
14.	Any other information important in the opinion of the tenderer		



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

Note:

- Page number/serial number may be given to each and every page of Tender Documents and photocopies of the documents attached. Mention Page number, wherever the copy (ies) of the document(s) is kept.
- In case of non-fulfilment of any of the above information/ document(s), the Tender will be summarily rejected without giving any notice.

(Dated Signature of the Tenderer with stamp of firm)

Date:

Place:

Undertaking

1. That I/we have carefully studied all the terms & conditions of NIT and shall abide by it.
2. That I/We shall supply the items of requisite quality.
3. That I/We undertake that the information given in this tender are true and correct in all respect and I/We hold the responsibility for the same.

(Dated Signature of the Tenderer with stamp of firm)

Date:

Place:



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

Chapter-V- Financial Bid

Format for Financial Bid

(To be submitted on the letterhead of the company / firm)

Sr. No.	Name of the furniture	Qty required	Rate	Vat/ Tax	Amount
1	Attendant stool	250			
2	Patient stretcher trolley	20			
3	Patient wheel chair Fixed	10			
4	Saline/intravenous stand	40			
5	Oxygen cylinder	40			
6	Oxygen trolley	40			
7	Stool with cushion & wheel	55			
8	X ray view box	25			
9	Glass door almirah	25			
10	Iron almirah (closed)	25			
11	Commode stool	2			
12	Bedpan	40			
13	Urinal	40			
14	Anti-vibration table	2			
15	Examination couch (same as OPD)	15			
16	Gynaecology table	1			
17	Instrument trolley	36			
18	Mayo'S Trolley	15			
19	Cardiac table/Over bed table	175			

1. I/We have gone through the terms & conditions as stipulated in the tender enquiry document and confirm to accept and abide by the same.
2. No other charges would be payable by the Institute.

(Dated Signature of the Tenderer with stamp of firm)

Date:

Place:



TECHNICAL SPECIFICATIONS PATIENT CARE UTILITIES

Sr.No	General Specification	Qty
1	<p><u>Attendant stool</u></p> <ul style="list-style-type: none">▪ Approximate size 480 mm to 650 mm▪ Tubular four legged SS base of 25.4 x 14 g ERW tube with Legs fitted with thick high quality nylon grommets.▪ Stainless Steel top having 300 mm diameter▪ Height adjustable by rotating SS Top from approx. 480 mm to 650 mm.▪ All SS to be of 304/16 grade/gauge	250
2	<p><u>Patient stretcher trolley</u></p> <ul style="list-style-type: none">▪ Overall Size: 1905mm L x 710mm W x 660mm To 910 mm H.▪ Stretcher dimension approx. 1830 mmL x 555 mm W.▪ Two section top. Height adjusted by foot operated maintenance free hydraulic pump. Height adjustment shall be obtained by hydraulically operated mono block type linear actuator pump foot operated actuation having stroke of 140 +/- 5 mm, push force 10 KN at 270 bars, number of complete pump stroke 22 to 24 for full stroke length. X-ray permeable removable stretcher, Backrest raised on ratchet. Quick trendelenburg as well as reverse trendelenburg positions shall be provided with easily accessible operating handle provided with two gas springs for easy action. S.S. saline rod with 12 mm dia S.S. rod shall telescope in SS socket tube approx. 15.8 mm dia x 18G welded on angular base bracket of 14G SS sheet. Nylon bracket provided to prevent colour damage. It could be placed at four different locations. Complete with sliding X-ray cassette holder, storage tray. Trolley shall be mounted on 125 mm dia non-rusting imported castor wheels two with brakes and two without. Castor housing and wheels made from high grade non floor-staining synthetic materials with integrated thread guards. Wheel centre having precision ball bearing to run smoothly. Complete with corner buffers, one on each corner. Covered handles. Oxygen cylinder arrangement.▪ It shall have a pair of Stainless steel tuck down type railings made of 19 mm dia x 18G tube fitted with M.S. brackets. Effective railing height above main frame is approx. 235 mm & length of the railing is 1175 mm. All MS parts and 8 tank Pre-treated & powder coated & SS parts finished with Matt polish.	20



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

3	<p><u>Patient wheel chair Fixed</u></p> <ul style="list-style-type: none">Overall size: 1120mm D x 670mm W x 920mm H approx.Tubular frame work of 22 mm OD x 18 G tube except for the four widthwise stays of 19 mm OD x 18 G tubesFor self-propulsion, two hoops of 15.8mm OD x 18 G SS tube mounted on rims of solid rubber tyre wheels of approx. 457 mmWheel chair should have cushioned seat and cushioned back of 50 mm thick PU foam mattress covered with rexine.Two foldable case aluminum foot rest and two 200 mm dia strong swivelling castor wheels provided in the frontAll MS parts should be pretreated and powder coated	10
4	<p><u>Saline/Intravenous stand</u></p> <ul style="list-style-type: none">I.V. Stand: Overall approx. size: height – 1360-2440 mm (with telescopic adjustable height)Ht adjustable with top knob (not side knob).Main Frame: Strong & Sturdy five legged neatly made base mounted on five 50 mm diameter swivelling ball castors of good quality.S.S. double hocked I.V. rod & S.S. tube spring loaded.All the Stainless Steel should be 304 grades/ 16 gauge.	40
5	<p><u>Oxygen cylinder</u></p> <ul style="list-style-type: none">It should have Seamless steel cylinderSize : Dia 232X5.4mmX46.7Ltr. Water capacity)It should have approx. nominal length 1370 mm(Without cap and without valve) and Outside diameter : 232 mmNeck ring fitted on it.Test pressure: 250 barWorking pressure should have 150 Bar.It should supply with all the tube ring and other accessories like Pressure Regulator, and other standard accessories including rinch.	40
6	<p><u>Oxygen trolley</u></p> <ul style="list-style-type: none">Frame of the cylinder trolley is made with hi-grade MS tubular steel.Framework of the cylinder trolley mounted on two 125 mm to 150 mm castors.Finish in epoxy powder coated.For small (40cft) Oxygen Cylinder.Pre-treated and Epoxy Powder Coating for Scratch proof Finish	40



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

7	<u>Stool with cushion & wheel</u> <ul style="list-style-type: none">▪ SS frame with five castors▪ Stainless Steel top. With detachable cushion▪ Adjustable height with a pneumatic height adjusting handle.▪ All around circular foot rest rod of solid stainless steel.▪ All SS of 304/16 grade/gauge▪ Thickness of cushion and density of PU foam?	55
8	<u>X ray view box</u> <ul style="list-style-type: none">▪ Thin wall mount type input voltage 240 Volts▪ LED back lit with digital dimmer control and built in ROM for intensity memory should be able to hold two full large size CT/MR films at a time.	25
9	<u>Glass door almirah</u> <ul style="list-style-type: none">▪ Size: 900 mm (W) × 354 mm (D) × 1990 mm (H)▪ Complete MS steel framework▪ Good quality three way key lock system.▪ There should be 5 adjustable shelves made up of acrylic.▪ All the side and front portion are made up of transparent acrylic window	25
10	<u>Iron almirah (closed)</u> <ul style="list-style-type: none">• It shall have multi bend construction and inter locking design and full length over lapping doors and side. It shall have two - way bolting mechanism. It shall be done with Spot welding & Co2 welding. It should have anticorrosion treated components oven baked powder coating process. The outer dimension should be 1980 x 916 x 486mm. Top, bottom LH& RH sides material thickness = 0.9mm CRCA. Its back material thickness = 0.8mm CRCA & its removable RH&LH main door with stiffeners material thickness = 0.9 mm CRCA . It should have brass handle with two way locking mechanism with shooting bolt. It should have four no. of full adjustable shelf of thickness 0.7 mm CRCA. It should have M 10 screw type leveller with hex plastic base. The finish should be epoxy powder coated to the thickness of 50 microns(+/- 10)	25



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

11	<p><u>Commode Chair</u></p> <ul style="list-style-type: none"> • Wheels 4", 2 wheels with brakes and 2 without brake wheels. • Seat - commode seat with collecting pan. • 50 mm 50 PU density foam cushioned top and back covered with leathered Rexene of 2 mm thickness • Seat height - adjustable type (suitable adjustment for western commode). • Can be used for use bathing / commode. • All the Stainless Steel should be 304 grade/ 16 gauge . 	2
12	<p><u>Bedpan</u></p> <ul style="list-style-type: none"> • Standard SS bedpan. 	40
13	<p><u>Urinal</u></p> <ul style="list-style-type: none"> • Made of SS • Lightweight, durable and easy to clean • Hold 1000 cc with graduation marks of at least 100 marks each to measure output. • Should have a sturdy grip for handling. • Can be used in several positions by bedridden patient. 	40
14	<p><u>Anti-vibration table</u></p> <ul style="list-style-type: none"> • Required for table top centrifuge & weighing machine. • The table should be heavily weighted pre-fabricated & rubber mounted to dampen the minute's vibration. • A small drawer should be provided for the convenience for storing weights, accessories etc. • The height from ground of one table should be between 34 to 37" and another should be 37-40" • Size : ~ 2 feet X 3 feet 	2
15	<p><u>Examination Couch</u></p> <ul style="list-style-type: none"> • Assisted adjustable backrest of 450 mm. • Three sliding door & three drawers for storage. • It should have overall size: 1890 mm L x 560 mm W x 840 mm H approximately. • It should have fixed rexine upholstered top 64 mm thick in two sections. Body frame work made from 20G CRCA sheet & 20 mm x 40 mm x 16G.MS. Rectangular Tubes. • It should have fitted stainless steel Legs & powder coated on all metal surfaces. • It should have Headrest adjustable on gas spring. • It should have upper section of box approx. size 1220 mm L 	15



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

	<p>x 460 mm W x 630 mm H with three sliding drawers of approx. size 320 mm x L x 430 mm W x 75 mm H.</p> <ul style="list-style-type: none"> • It should have lower section comprising of three cabinets of approx. inside size 350 mm L x 440 mm W x 430 mm H with Separate doors. • It should have sliding foot step under the front side of lower middle cabinet. • It should have BP apparatus tray made of 18 G MS sheet of approx. size 350 mm L x 120 mm W x 20 mm H provided on swinging rod rotating through a bush fixed on the body of the couch. • All MS parts should be pre-treated & powder coated & SS Parts finished with Matt polish. 	
16	<p><u>Gynaecology table</u></p> <ul style="list-style-type: none"> • Overall Approx. Size: 1830mmL x 600mmW x 800mmH with U notch. • Frame work made of MS tube. • Two section cushioned top • Complete with the pair of S.S. lithotomy rod made of 12 mm S.S. Rod • Back rest adjustable by hand lever • Epoxy powder coated finish. • To be provided with 18 G MS steel double foot step .Size 510mm LX 300mm WX 450 H 	1
17	<p><u>Instrument trolley</u></p> <ul style="list-style-type: none"> • Overall approx. size: 680mm L x 450mm W x 900mm H. • Stainless steel tubular frame work made of 25.4mm OD x 18 G verticals mounted on 125 mm dia non –rusting castor two with brakes. Castor housing and wheels made from high grade non floor-staining synthetic materials with integrated thread guards. Wheel centre having precision ball bearing to run smoothly. • Two stainless steel shelves of 20G thickness, with protective railings on three sides. The railing shall be made from dia 10 mm S.S.rod. Only 304 grade stainless steel should be used for trolley frame work and SS shelves. SS parts finished with Matt Polish. 	36



All India Institute of Medical Sciences Jodhpur

Admin/General/98/2013-AIIMS.JDH

18	<u>Mayo's Trolley</u> <ul style="list-style-type: none">Adjustable height of SS tray from approx. 760 mm to 1270 mm. Tray size approx. : 560 mm L x 400 mm W. SS tubular frame mainly made of 31.7mm dia x 18G round tube. Tray made from 0.9 mm thick S.S. 304 grade sheet. Tray frame made from 25 x 5mm, a cantilever of 10 mm dia SS 304 grade rod and SQ. bar of 16 mm S.S. 304 Q mounted on four swivels, twin wheel non rusting castors, 50 mm wheel dia, all without brake. SS parts finished with Matt Polish.	15
19	<u>Cardiac table/Over bed table</u> <ul style="list-style-type: none">Overall size: 112 mm LX 450 mmWX 965mm HTop Size: 760 mm LX 450 mm WTop made from 12 mm MDF with 0.6 mm thick laminate pasted on it.Bidding on all four side of top. Tubular frame work with 50 mmDia swivelling caster .All without brakeAll MS parts are 8 tank Pre-treated & powder coated.	175