

Tender

For

Equipments required for Department of Pediatric Surgery

At

All India Institute of Medical Sciences, Jodhpur

NIT Issue Date : 15th January, 2016
NIT No. : Admn/Tender/Pediatric Surgery/2015-AIIMS.JDH
Pre-Bid Meeting : 04th February, 2016 at 03:00 PM.
Last Date of Submission : 22nd February, 2016 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur, Rajasthan-342005.

Telephone: 0291- 2012978, email: aoadmin@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department

All India Institute of Medical Sciences (AIIMS), Jodhpur, Rajasthan, an apex healthcare institute being established by an Act of Parliament of India under aegis of Ministry of Health & Family Welfare, Government of India, invites sealed tenders for supply & installation of the following items for the Department of Pediatric Surgery at the institute. You are requested to quote your best offer along with the complete details of specifications, terms & conditions.

Chapter-I

S.No	NIT No.	Item Description	Qty	EMD (Rs.)
01.	Admn/Tender/Pediatric Surgery/01/2015-AIIMS.JDH	Transport Incubator	1	10,000
02.	Admn/Tender/Pediatric Surgery/02/2015-AIIMS.JDH	LED Head Light with Cold Illumination	1	6,000
03.	Admn/Tender/Pediatric Surgery/03/2015-AIIMS.JDH	Endoscopic Rectoscopy System (Pediatric Set)	1	26,000
04.	Admn/Tender/Pediatric Surgery/04/2015-AIIMS.JDH	Composite Offset camera Pediatric Cystoscope 10 Fr	1	30,000
05.	Admn/Tender/Pediatric Surgery/05/2015-AIIMS.JDH	Pediatric Open Surgical Instruments Sets		
		a) Pediatric Retractor System	1	16,000
		b) Pediatric Open Surgical Instruments	1	8,000
		c) Instruments Case & Trays	1	4,000
06.	Admn/Tender/Pediatric Surgery/06/2015-AIIMS.JDH	Operating Loupes	2	2,400
07.	Admn/Tender/Pediatric Surgery/07/2015-AIIMS.JDH	Patient Warming System	3	78,000

Quotation should be sealed and super-scribed with tender number and address to:

“Administrative Officer
All India Institute of Medical Sciences, Jodhpur
Basni, Phase-II
Jodhpur-342005, Rajasthan”.

The sealed quotations should reach the Institute, latest by 22nd February, 2016 at 03:00 PM and it will be opened on same day at 04:30 PM in the Conference Hall, Academic Block, AIIMS Jodhpur in the presence of the bidder(s) or their authorized representative(s), who will present at the scheduled date and time.

Terms & Conditions:

- 1. Preparation and Submission of Tender:** The tender should be submitted in two parts i.e. Technical Bid and Financial Bid. The Technical Bid and the Financial Bid should be sealed by the bidder in two separate covers “**Technical Bid for Tender for Supply of (Item Name)**” and “**Financial Bid for Tender for Supply of (Item Name)**”. Both Sealed Envelopes (separate for each NIT) should be kept in a main/ bigger envelope super-scribed as “**Tender for Supply of (Item Name)**”.
- 2. Earnest Money Deposit: Earnest Money Deposit:** The bidder shall be required to submit the Earnest Money Deposit (EMD) for an amount as per Chapter-I by way of demand drafts and Bank Guarantee only. The demand drafts shall be drawn in favour of “**All India Institute of Medical Sciences, Jodhpur**”. The demand drafts and Bank Guarantee for earnest money deposit must be

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department

enclosed in the envelope containing the technical bid. The EMD of the successful bidder shall be returned after the successful submission of Bank Guarantee/ Security Deposit and for unsuccessful bidder(s) it would be returned after award of the contract. Bid(s) received without demand drafts of EMD will be rejected.

The firms who are registered with National Small Industries Corporation (NSIC) / OR Small Scale Industrial (SSI) are exempted to submit the EMD (copy of registration must be provide along with).

3. **Tender Fee:** Tender fee will be Non-refundable amount of Rs. 1,000/- (Rupees One Thousand only) **for Each Item** by way of demand drafts only. The demand drafts shall be drawn in favour of “**All India Institute of Medical Sciences, Jodhpur**”. Bids received without demand draft of tender will be rejected.
4. **Validity:** The quoted rates must be valid for a period for 180 days from the date of closing of the tender. The overall offer for the assignment and bidder(s) quoted price shall remain unchanged during the period of validity. If the bidder quoted the validity shorter than the required period, the same will be treated as unresponsive and it may be rejected.
5. **"PRE –BID Meeting" with the intending bidders shall be held on 04th February, 2016 from 03:00 P.M. onwards at AIIMS, Jodhpur.** All the prospective bidders are requested to send comments/ representations on or before pre-bid meeting. Intending bidder will be allowed to seek clarification on specification, Conditions of Contract, etc. in writing to AIIMS, Jodhpur, within 48 hours after the pre-bid meeting.
6. In case the tenderer withdraws, modifies or change his offer during the validity period, bid is liable to be rejected and the earnest money deposit shall be forfeited without assigning any reason thereof. The tenderer should also be ready to extend the validity, if required, without changing any terms, conditions etc. of their original tender.
7. **Delivery and Installation:**
 - i) For goods supplied from India:**
All the goods ordered shall be delivered and Installed at AIIMS, Jodhpur within **30 days** from the date of issue of supply order.
 - ii) For goods imported directly from abroad:**
All the goods ordered shall be delivered and Installed at AIIMS, Jodhpur within **60 days** from the date of opening of Letter of Credit for shipment.

All the aspects of safe delivery, installation and commissioning shall be the exclusive responsibility of the supplier. If the supplier fails to deliver, install and commission the goods on or before the stipulated date, then a penalty at the rate of 0.5% per week of the total order value shall be levied subject to maximum of 10% of the total order value. The successful tenderer will also provide required training for supplied items at AIIMS Jodhpur. The goods should be manufactured after adoption of latest technology.

If at any time during the currency of the contract, the supplier encounters conditions hindering timely of the goods and performance of services, the supplier shall promptly inform the AIIMS, Jodhpur for extension of the delivery schedule accordingly. On receiving the supplier's communication, the AIIMS, Jodhpur shall examine the situation as soon as possible and, at its discretion, may agree to extend the delivery schedule, with or without liquidated damages for completion of supplier's contractual obligations by issuing an amendment to the contract.

In the case of package supply where the delayed portion of supply materially hampers installation and commissioning of the systems, liquidated damages charges shall be levied as above on the total value of the concerned package of the purchase order. Quantum of liquidated damages assessed and levied by the purchaser shall be final and not challengeable by the supplier.

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department

- 8. Signing the Contract:** - The successful bidder shall be required to execute the Contract Agreement accepting all terms and conditions stipulated herein on a non-judicial stamp paper of Rs. 100/- (Rs. One Hundred only) along with performance security within fifteen days of the issue of the Letter of notification of award. In the event of failure on the part of the successful bidder to sign the Contract within the period stipulated above, the EMD shall be forfeited and the acceptance of BID shall be considered as cancelled.
- 9. Performance Security:** As a guarantee towards due performance and compliance of the contract work, the successful bidder (contractor) will deposit an amount equal to 10% of order value and should be kept valid for a period of 60 day beyond completion of all the contractual obligation, including CMC period towards security deposit by way of demand draft/ bank Guarantee in favour of “All India Institute of Medical Sciences, Jodhpur “drawn on any Nationalized Bank/Scheduled Bank and payable at Jodhpur within fifteen days of the issue of the Letter of notification of award along with non-judicial stamp paper of Rs. 100/- (Contract agreement).
- 10. Incidental Services:** The supplier shall be required to perform the following services:-
- Installation & Commissioning, Supervision and Demonstration of the goods.
 - Providing required jigs and tools for assembly, minor civil works required for the completion of the installation.
 - On Site Training to Doctors/ Technicians/ Staff is to be provided by Supplier for operation and maintenance of the equipment for a period of 30 working days after successful installation of the machine, as per direction of user department.
 - Supplying required number of operation & maintenance manual for the goods.
 - To provide non-locked open software and standard interface inter-operability conditions for networked equipment’s in hospital management information system, wherever applicable.
- 11. Accessories & Consumables:** The separate price list of all accessories and consumables, if any, must be attached/ enclosed along with the Financial Bid.
- 12. After Sales Service:** After sales service centre should be available on 24 (hrs.) X 7 (days) X 365 (days) basis. Complaints should be attended properly, maximum within 24 hrs to ensure an uptime of minimum 95%, wherever applicable, failing which the necessary penalty measures shall be enforced.
- 13. Inspection:**
- AIIMS, Jodhpur shall have the right to inspect and/or to test the goods to confirm their conformity to the NIT Specifications at no extra cost to the Purchaser.
 - AIIMS, Jodhpur right to inspect, test and, where necessary, reject the Goods after the goods arrival at the final destination shall in no way be limited or waived by reason of the Goods having previously been inspected, tested and passed by AIIMS, Jodhpur prior to the goods shipment.
 - The Director, AIIMS Jodhpur shall be the final authority to reject full or any part of the supply which is not confirming to the specification and other terms and conditions.
 - No payment shall be made for rejected Stores. Rejected items must be removed by the Bidders within two weeks of the date of rejection at their own cost and replaced immediately. In case these are not removed, these will be auctioned at the risk and responsibility of the suppliers without any further notice.
- 14. Documents:**
- All pages of the Tender should be numbered and indexed.**
 - The bidder shall provide in its tender the required as well as the relevant documents like technical data, literature, drawings etc. to establish that the goods and services offered in the tender fully confirm to the goods and services specified by the purchaser in the tender documents. For this purpose the bidder shall also provide a clause-by-clause commentary on the technical specifications and other technical details incorporated by the purchaser in the tender documents to establish technical responsiveness of the goods and services offered in its tender duly indicating relevant page numbers in the product literature.

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department

- c. The bidder shall provide a list of major Government and Private Institutions where its relevant bid item has been supplied during last one year.
- 15. Manufacturer Authorisation:** The bidder (if not original equipment manufacturer must submit Original Equipment Manufacturer authorization certificate that the tenderer is authorized for selling and maintain the equipment quoted for. Performa attached at **Annexure- III**.
- 16.** The bidders are required to submit user certificate for the relevant equipment on the letter head of the institution (Government/ Private).
- 17.** The successful bidder will be required to submit order copies of the supply of the equipment in Government institutions in last 12 month for rate reasonability purpose.
- 18. Insurance:** - The supplier shall make arrangements for insuring the goods against loss or damage incidental to manufacture or acquisition, transportation, storage and delivery. If the equipment's is not commissioned and handed over to AIIMS, Jodhpur within specified period, the insurance will have to be extended by the supplier at their cost till the successful installation, testing, commissioning and handing over of the goods to the AIIMS, Jodhpur.
- 19. Tender Currencies:**
- The bidder supplying indigenous goods or already imported goods shall quote only in Indian Rupees. Further, imported goods to be imported and supplied by the bidder are also required to be quoted in Indian Rupees.
 - For imported goods if supplied directly from abroad, prices shall be quoted in any freely convertible currency say US Dollar, Euro, GBP or Yen. As regards price(s) for allied services, if any, required with the goods, the same shall be quoted in Indian Rupees only, if such services are to be performed /undertaken in India.
 - Tenders, where prices are quoted in any other way shall be treated as non -responsive and rejected.
- 20. Tender Prices:** While filling up the columns of the Financial Bid, the following aspects should be noted for compliance:
- For domestic goods or goods of foreign origin located within India, the prices in the corresponding Financial Bid shall be entered separately in the following manner:**
- The price of the goods, quoted ex-factory/ ex-showroom/ ex-warehouse/ off-the-shelf, as applicable, including all taxes and duties like sales tax, CST/ VAT, CENVAT, Custom Duty, Excise Duty etc. already paid or payable on the components and raw material used in the manufacture or assembly of the goods quoted ex-factory etc. or on the previously imported goods of foreign origin quoted ex-showroom etc.;
 - Any sales tax or other taxes and any duties including excise duty, which will be payable on the goods in India if the contract is awarded;
 - Charges towards Packing & Forwarding, Inland Transportation, Insurance, Loading/Unloading and other local costs incidental to delivery of the goods to their final destination as specified in the List of Requirements and Financial Bid;
 - The price of Incidental Services, as mentioned in List of Requirements and Financial Bid;
 - The prices of Turnkey (if any), as mentioned in List of Requirements, Technical Specification and Financial Bid; and
 - The price of annual CMC, as mentioned in List of Requirements, Technical Specification and Financial Bid.

For goods offered from abroad, the prices in the corresponding Financial Bid shall be entered separately in the following manner:

- The price of goods quoted FOB port of shipment, as indicated in the List of Requirements and Financial Bid;
- The price of goods quoted CIF port of entry in India as indicated in the List of Requirements and Financial Bid;

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department

- c. The price of goods quoted for delivery at AIIMS, Jodhpur as indicated in the List of Requirements, Financial Bid and Consignee List;
- d. Wherever applicable, the amount of custom duty with CDEC applicable on CIF value on the goods to be imported;
- e. The charges for Loading/Unloading, Inland transportation, Insurance and other local costs, Incidental cost to delivery of the goods from the port of entry in India to AIIMS, Jodhpur, as specified in the List of Requirements and Financial Bid;
- f. The charges for Incidental Services, as in the List of Requirements and Financial Bid;
- g. The prices of Turnkey (if any), as mentioned in List of Requirements, Technical Specification and Financial Bid; and
- h. The price of annual CMC, as mentioned in List of Requirements, Technical Specification and Financial Bid.

Additional information and instruction on Duties and Taxes: If the Bidder desires to ask for excise duty, sales tax/CST / VAT/ CENVAT, Custom Duty, Service Tax, Works Contract Tax etc. to be paid extra, the same must be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such duties and taxes and no claim for the same will be entertained later.

Excise Duty:

- a. If reimbursement of excise duty is intended as extra over the quoted prices, the supplier must specifically say so also indicating the rate, quantum and nature of the duty applicable. In the absence of any such stipulation it will be presumed that the prices quoted are firm and final and no claim on account of excise duty will be entertained after the opening of tenders.
- b. If a Bidder chooses to quote a price inclusive of excise duty and also desires to be reimbursed for variation, if any, in the excise duty during the time of supply, the Bidder must clearly mention the same and also indicate the rate and quantum of excise duty included in its price. Failure to indicate all such details in clear terms may result in rejection of that tender.
- c. Subject to sub clauses (i) & (ii) above, any change in excise duty upward/downward as a result of any statutory variation in excise duty taking place within contract terms shall be allowed to the extent of actual quantum of excise duty paid by the supplier. In case of downward revision in excise duty, the actual quantum of reduction of excise duty shall be reimbursed to the purchaser by the supplier. All such adjustments shall include all reliefs, exemptions, rebates, concession etc. if any obtained by the supplier.

Sales Tax: - If a bidder asks for sales tax/CST / VAT/CENVAT, Service Tax and Works Contract Tax to be paid extra, the rate and nature of sales tax applicable should be shown separately. The CST / VAT/CENVAT, Service Tax and Works Contract Tax will be paid as per the rate at which it is liable to be assessed or has actually been assessed provided the transaction of sale is legally liable to sales tax/ CST / VAT/CENVAT, Service Tax and Works Contract Tax and is payable as per the terms of the contract.

Octroi Duty and Local Duties & Taxes:- Normally, goods to be supplied to Government departments against Government contracts are exempted from levy of town duty, Octroi duty, terminal tax and other levies of local bodies. However, on some occasions, the local bodies (like town body, municipal body etc.) as per their regulations allow such exemptions only on production of certificate to this effect from the concerned Government department. Keeping this in view, the supplier shall ensure that the goods to be supplied by the supplier against the contract placed by the AIIMS, Jodhpur are exempted from levy of any such duty or tax and, wherever necessary, obtain the exemption certificate from the AIIMS, Jodhpur. However, if a local body still insists upon payment of such local duties and taxes, the same should be paid by the supplier to the local body to avoid delay in supplies and possible demurrage charges and obtain a receipt for the same. The supplier should forward the receipt obtained for such payment to the AIIMS, Jodhpur to enable the AIIMS, Jodhpur reimburse the supplier and take other necessary action in the matter.

Customs Duty: In respect of imported goods offered from abroad, the bidder shall specify the rate as well as the total amount of customs duty payable with Custom Duty Exemption Certificate, if

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department

applicable, on the quoted goods in the Financial Bid. The bidder shall also indicate the corresponding Indian Customs Tariff Number applicable for the goods.

- a. For transportation of imported goods offered from abroad, relevant instructions as incorporated shall be followed.
- b. For insurance of goods to be supplied, relevant instructions as provided shall be followed.
- c. Unless otherwise specifically indicated in this NIT document, the terms FCA, FOB, FAS, CIF, CIP etc. for imported goods offered from abroad, shall be governed by the rules & regulations prescribed in the current edition of INCOTERMS, published by the International Chamber of Commerce, Paris.
- d. The need for indication of all such price components by the bidders, as required in this clause is for the purpose of comparison of the tenders by the purchaser and will no way restrict the AIIMS, Jodhpur right to award the contract on the selected bidder on any of the terms offered.

21. Indian Agent:- If a foreign bidder has engaged an agent in India in connection with its bid, the foreign bidder, in addition to indicating Indian agent's commission, if any, shall also furnish the following information:

- a. The complete name and address of the Indian Agent and its Permanent Account Number as allotted by the Indian Income Tax authority.
- b. The details of the services to be rendered by the agent for the subject requirement.
- c. Details of Service outlets in India, nearest to the AIIMS, Jodhpur to render services during Warranty and CMC period.

22. Firm Price

- a. Unless otherwise specified in the NIT, prices quoted by the bidder shall remain firm and fixed during the currency of the contract and not subject to variation on any account.
- b. However, as regards taxes and duties, if any, chargeable on the goods and payable, the conditions stipulated will apply.

23. Conversion of tender currencies to Indian Rupees: - In case the bid document permits the bidders to quote their prices in different currencies, all such quoted prices of the responsive bidders will be converted to a single currency viz., Indian Rupees for the purpose of equitable comparison and evaluation, as per the closing exchange rates established by the Reserve Bank of India for similar transactions, as on the date of '**Last Date of Submission of Tender**'.

24. Payment Terms:

i) Payment for goods supplied from India:

100% payment of the total order value shall be released after the successful installation/ commissioning of the ordered goods against the submission of the inspection report.

ii) Payment for Imported goods:

For imported goods payment shall be made in the following manner:

- a) **On shipment:** 75 % payment of the contract price shall be paid 60 days after presentation of shipping documents {goods shipped shall be paid through irrevocable, non-transferable Letter of Credit (LC) opened in favour of the supplier in a bank in his country} and upon the submission of the following documents:
 - i. Four copies of Supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
 - ii. Original and four copies of the clean, on-board Bill of Lading/ Airway bill, marked freight prepaid and four copies of non-negotiable Bill of Lading/Airway bill.
 - iii. Insurance Certificate;
 - iv. Certificate of origin by the chamber of commerce of the concerned country;
 - v. Certificate of country of origin;
 - vi. Manufacturer's / Supplier's warranty certificate;
 - vii. Manufacturer's own factory inspection report.
- b) **On Acceptance:** 25 % payment would be made after satisfactory installation & commissioning on issuance of Inspection certificate by the AIIMS, Jodhpur.

Note:-The supplier shall not claim any interest or any other payment under the contract.

25. Custom Clearance: For the Goods to be imported and supplied, the Institute will provide Custom Duty Exemption Certificate (CDEC) to successful bidder for availing concessional rate of duty as per prevailing Custom Tariff. In case, the bidder requires CDEC certificate, then the same should be specifically mentioned in the bid. The supplier is solely responsible for getting the material clearance from customs. Institute will provide all custom documents for custom clearance on the demand of supplier. Transportation of goods up to AIIMS, Jodhpur and its successful installation and commissioning is also the responsibility of the supplier. All charges/ expenses incurred in this process will be borne by the supplier. **NO DEMURRAGE / WHARFAGE CHARGES WILL BE PAYABLE BY THE INSTITUTE UNDER ANY CIRCUMSTANCES. NO ADVANCE PAYMENT WILL BE PAYABLE FOR CUSTOM CLEARANCE/ FREIGHT/INSURANCE ETC.**

26. Guarantee / Warrantee Period:

26.1 For the equipment value upto Rs. 5 Lakh: The Tenderers must quote for **02 years** comprehensive warranty (Including all Spares, Accessories and Labour) from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. Also the bidders are requested to submit their quote (Rates) for subsequent **03 years** Comprehensive Maintenance Contract (CMC) (Including All Spares, Accessories and Labour). Failure to comply this condition will entail the rejection of the bids. The price comparison shall be taking into account on basic price and post warranty CMC.

26.2 For the equipment value above Rs. 5 Lakh: The Tenderers must quote for **05 years** comprehensive warranty (Including all Spares, Accessories and Labour) from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. Also the bidders are requested to submit their quote (Rates) for subsequent **05 years** Comprehensive Maintenance Contract (CMC) (Including All Spares, Accessories and Labour). Failure to comply this condition will entail the rejection of the bids. The price comparison shall be taking into account on basic price and post warranty CMC.

27. Uptime guarantee: The firm should provide uptime guarantee of 95%

28. Downtime penalty Clause

- a. During the comprehensive warranty period, the guarantee uptime of 95% of 365 days will be ensured. In case the down time exceeds the 5% limit penalty of extension of guaranty period by two days for each additional day of down time will be enforced. The vendor must undertake to supply all spares for optimal upkeep of the equipment for at least FIVE YEARS after handling over the unit to the Institute. If accessories / other attachment of the system are procured from the third party, then the vendor must produce cost of accessory / other attachment and the CMC from the third party separately along with the main offer and the third party will have to sign the CMC with the Institute if required.
- b. The principals or their authorized service providers are required to submit a certificate that they have satisfactory service arrangements and fully trained staff available to support the uptime guarantee.

29. Arbitration: If any difference arises concerning this agreement, its interpretation on payment to the made there-under, the same shall be settled out by mutual consultation and negotiation. If attempts for conciliation do not yield any result within a period of 30 days, either of the parties may make a request to the other party for submission of the dispute for decision by an Arbitral Tribunal containing Sole Arbitrator to be appointed by the Secretary, Department of Legal Affairs. Such requests shall be accompanied with a panel of names of three persons to act as the sole arbitrator. In case of such arbitrator refusing, unwilling or becoming incapable to act or his mandate having been terminated under law, another arbitrator shall be appointed in the same manner from among the panel of three persons to be submitted by the claimant. The provision of Arbitration and Conciliation Act, 1990 and the rule framed there under and in force shall be applicable to such proceedings.

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department

- 30. Subletting of Work:** The firm shall not assign or sublet the work/job or any part of it to any other person or party without having first obtained permission in writing of AIIMS, Jodhpur, which will be at liberty to refuse if it thinks fit. The tender is not transferable. Only one tender shall be submitted by one tenderer.
- 31. Breach of Terms and Conditions:** In case of breach of any terms and conditions as mentioned above, the Competent Authority, will have the right to cancel the work order/ job without assigning any reason thereof and nothing will be payable by AIIMS, Jodhpur in that event the security deposit shall also stand forfeited.
- 32. Insolvency etc:** In the event of the firm being adjudged insolvent or having a receiver appointed for it by a court or any other order under the Insolvency Act made against them or in the case of a company the passing any resolution or making of any order for winding up, whether voluntary or otherwise, or in the event of the firm failing to comply with any of the conditions herein specified AIIMS, Jodhpur shall have the power to terminate the contract without any prior notice.
- 33. Force Majeure:** If, at any time during the subsistence of this contract, the performance in whole or in part by either party of any obligation under this contract is prevented or delayed by reasons of any war or hostility, act of public enemy, civil commotion, sabotage, fire, floods, explosion, epidemics, quarantine restriction, strikers lockout or act of God (hereinafter referred to as events) provided notice of happening of any such eventuality is given by party to other within 21 days from the date of occurrence thereof, neither party shall by reason of such event be entitled to terminate this contract nor shall either party have any claim for damages against other in respect of such non-performance or delay in performance, and deliveries have been so resumed or not shall be final and conclusive.
- Further, that if the performance in whole or in part of any obligation under this contract is prevented or delayed by reason of any such event for a period exceeding 60 days, either party may, at least option to terminate the contract.
- 34.** Bidder shall submit a copy of the tender document and addenda thereto, if any, with each page of this document should be signed and stamped to confirm the acceptance of the entire terms & conditions as mentioned in the tender enquiry document.
- 35.** The quantity of item given in the tender is tentative, which may be increased or decreased as per the institute's requirement.
- 36.** Signed & stamped compliance sheet of the technical specification of the goods with technical printed literature must be enclosed with the bid.
- 37.** After due evaluation of the bid(s) Institute will award the contract to the lowest evaluated responsive tenderer
- 38.** Conditional bid will be treated as unresponsive and it may be rejected.
- 39. Demonstration:** - AIIMS Jodhpur reserves the right to ask the tenderers for arranging demonstration of their equipment for which rates have been quoted, to the concerned committee, if required.
- 40.** The Institute reserves the right to accept in part or in full or reject any or more tender(s) without assigning any reasons or cancel the tendering process and reject all tender(s) at any time prior to award of contract, without incurring any liability, whatsoever to the affected bidder or bidder(s).
- 41. Applicable Law:**
- The contract shall be governed by the laws and procedures established by Govt. of India, within the framework of applicable legislation and enactment made from time to time concerning such Commercial dealings / processing.

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department

- Any disputes are subject to exclusive jurisdiction of Competent Court and Forum in Jodhpur, Rajasthan, India only.
- The Arbitration shall be held in accordance with the provisions of the Arbitration and Conciliation Act, 1996 and the venue of arbitration shall be at Jodhpur. The decision of the Arbitrator shall be final and binding on both the parties.
- Force Majeure: Any delay due to Force Majeure will not be attributable to the supplier.

**Administrative Officer
AIIMS, Jodhpur**

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department
Annexure-I

S.No	Item	Technical Specification	Qty.
01.	Transport Incubator	<ol style="list-style-type: none"> 1. Required for transportation of premature babies and neonates and it can be used for long distance transportation. 2. It should be mobile intensive care station including transport ventilator, incubator, and power supply unit and infusion stand 3. It should be mounted on collapsible trolley having lockable rust free castors of the size 4 inches or more and with two A type Aluminum oxygen cylinders on rack under the Incubator . 4. Single walled incubator with at least two large port holes for access. Iris ports for ventilator & other tubings. Bed level at least 80 cms. above ground level. Two shelves cabinet with door. 3 Width: app 80 cm+ 5 cms. , Depth 30 cm + 5 cm, height 115 + 5 cms, Mattress to hood distance at least 30 cms. 5. Air Mode: adjustable set temperatures between 20 – 39 C. Display of set temperatures with resolution of 0.1 C. Skin mode adjustable set temperatures between 34 – 38 C. Display of set temperatures with resolution of 0.1 C. 6. Alarms of High , Low and Probe failure for the set air mode up to +2.5 C and skin mode of + 0.5 C of temperatures 7. Oxygen monitor in incubator hood with display of 21 – 100% Oxygen alarms for high, low and probe failure. 8. Heart and Oxygen saturation monitor: Fixed, built monitors, dual wavelength probe for Oxygen saturation with Digital LED display for Heart rate and Oxygen saturation. Alarms for high and low for Heart Rate, Oxygen saturation and probe failure 9. The system should have an internal rechargeable maintenance free battery to ensure continued functioning of the unit for at east 4 hours during transport. It should have automatic switch circuit for change over from battery to AC and vice versa. 10. One suction apparatus with negative suction pressure of 5-120 mm Hg should be provided. IV fluid stand should support two infusion bottles 11. One Syringe infusion pump with stand compatible with 10, 20, and 50 ml syringes compatible with locally available brand of syringes. Range of infusion rate 1 – 99 ml / hr.in steps of 0.1ml. Display infusion rates, Alarms for occlusions, end of infusion with internal rechargeable battery should be provided along with the quoted price 12. Height less than 60”, depth less than 30”, width 33”-36”. Weight 90-100 kg. With wheel mounted. (All dimensions in approximation of +/-10%) 	01

		<p>13. Environmental Factors</p> <ol style="list-style-type: none"> a. Shall meet IEC-60601-1-2 :2001(Or Equivalent BIS) General Requirements of Safety for Electromagnetic Compatibility.or should comply with 89/366/EEC; EMC-directive. b. The unit shall be capable of being stored continuously in ambient temperature of 0 -50deg C and relative humidity of 15-90% c. The unit shall be capable of operating continuously in ambient temperature of 10 -40 deg C and relative humidity of 15-90% <p>14. Power Supply</p> <ol style="list-style-type: none"> a. Power input to be 220-240VAC, 50Hz b. Voltage corrector/stabilizer of appropriate ratings meeting ISI Specifications.(Input 160-260 V and output 220-240 V and 50 Hz) <p>15. Standards, Safety and Training</p> <ol style="list-style-type: none"> a. Should be US FDA or European CE approved b. Manufactures/Supplier should have ISO certificate to Quality Standard <p>16. List of important spare parts and accessories with their part number and costing should be provided</p> <p>17. User/Technical/Maintenance manuals to be supplied in English.</p>	
<p>02.</p>	<p>LED Headlight</p>	<ol style="list-style-type: none"> 1. Should be comfortable and mobile LED Head Light and easy to use without any cables 2. Should have a maximum luminous Intensity of around 2,00,000 lux or more 3. Should have color temperature of 5000K 4. Should have minimum battery service life of 6 hours after full charge 5. Batteries should be capable of being charged easily without powering off 6. Should be able to have rapid USG charging from wall adaptor or personal computer 7. Should have integrated battery status indicators 8. Should have continous digital dimming facility 9. Should be light weight and less than 600 gms 10. Should be able to focus down to about 65 mm 11. Should be comfortable through a removable, washable headband 12. Should have an extra spare battery 13. Should be provided with easily attachable 2 Dioptre and 3 dioptre Magnifying Lenses 14. Shoud be European CE or US FDA approved and of International safety standards 15. Prior Demo is required 	<p>01</p>

<p>03</p>	<p>Endoscopic Rectoscopy System for Children</p>	<ol style="list-style-type: none"> 1. Rectoscope Tube – 1 <ol style="list-style-type: none"> 1. Outer diameter – 12mm 2. Working Length – 20 cms 3. With Obturator 2. Rectoscope Tube – 1 <ol style="list-style-type: none"> 1. Outer Diameter – 20 mm 2. Working Length – 25 cms 3. With Obturator 3. Compatible Illumination head for Rectoscope Tube - 1 <ol style="list-style-type: none"> 1. With fiber optic light transmission with Luer Lock jub for rubber insufflation bulb 4. Glass Window Plug for Illumination head - 1 5. Glass Window for Sigmoidoscope - 1 6. Adjustable Magnifier for Illumination head with sealing cap -1 <ol style="list-style-type: none"> 1. For use with Instruments size 10mm 2. Reducer for use with instruments size 5 mm - 3 7. Rubber Insufflation bulb -2 <ol style="list-style-type: none"> 1. Detachable Rubber tube -2 2. Rubber bulb -2 8. Airtight Sliding Cap <ol style="list-style-type: none"> 1. for Centering a 10mm telescope 9. Compatible Forward oblique telescope - 1 <ol style="list-style-type: none"> 1. 30 degree 2. 45 degree angled eyepiece 3. Diameter 4 mm 4. Length – 28.5 cms 10. Telescope Stopper – 100 <ol style="list-style-type: none"> 1. Size 11 mm 2. Sterile, Single use 11. Working Attachment for rectoscope tube - 1 <ol style="list-style-type: none"> 1. Channel for Telescope 2. Two working channels for instruments with diameter upto 5 mm 12. Biopsy Forceps – 30 cms - 1 <ol style="list-style-type: none"> 1. Through Cutting 2. Single action jaws 3. Small jaws 4. Compatible with Working Channel of Working attachment for Rectoscope tube 13. Biopsy Forceps – 30 cms - 1 <ol style="list-style-type: none"> 1. Through Cutting 2. Single action jaws 3. Medium Size jaws 4. Compatible with Working Channel of Working attachment for Rectoscope tube 14. Sponge Forceps – 1 <ol style="list-style-type: none"> 1. 30 cms length 2. Compatible with Working Channel of Working 	<p>01</p>
------------------	---	--	------------------

		<p>attachement for Rectoscope tube</p> <p>15. Suction and coagulation tube – insulated with unipolar coagulation – 1</p> <ol style="list-style-type: none"> 1. 30 cms Length 2. Compatible with working channel of working attachment for rectoscope tube <p>16. Snare Guide for Rectal Polyp – 2</p> <ol style="list-style-type: none"> 1. 40 cms length 2. Insulated 3. Connector for Unipolar coagulation 4. Self Forming wire snare 5. Should be provided with 10 extra Wire Snares <p>17. Polyp Grasping Forceps for use with snare guided – 1</p> <ol style="list-style-type: none"> 1. Length – 40 cms <p>18. Compatible Plastic Container for Sterilizing and Storage, perforated, with transparent lid, with inserts for telescope and silicone mat</p> <p>19. Plastic Container for Sterilizing and Storage, perforated, with lid, for two-level storage, for use with 30 cm and 36 cm instruments, external dimensions (w x d x h): 600 mm x 145 mm x 260 mm</p>	
<p>04</p>	<p>Composite Offset Pediatric Operating Cysto-urethroscope</p>	<ol style="list-style-type: none"> 1. Composite 9.5 – 10 Fr Cysto-urethroscope with Angled Eyepiece and working channels - 1 <ol style="list-style-type: none"> 1. 9.5 Fr-10 Fr Tip 2. Autoclavable 3. With 45 degree Angled eyepiece/ocular 4. Length – 150-170mm 5. Working channel of 6 Fr or more for Operating instruments of 5 Fr 6. Incorporated Fiber optic light transmission 7. With 2 lateral channels irrigation ports 2. Composite 7.5-8 Fr Cysto-urethroscope with Angled Eyepiece and Working Channels – 1 <ol style="list-style-type: none"> 1. 7.5-8 Fr Tip 2. Autoclavable 3. With 45 degree Angled eyepiece/ocular 4. Length – 210-230 mm 5. Working Channel of more than 4 Fr for Operating Instruments of 3 Fr 6. Incorporated Fiber optic light transmission 7. With 2 lateral channels irrigation ports 3. Should be supplied with following accessories <ol style="list-style-type: none"> 1. Instrument port with sealing system with One channel – 1 2. Seal for working channel - 10 3. Luer Lock tube connector 4. Cleaning adaptor 5. Insterion Aid for guide wires 6. Instrument post with sealing sysem – 2 channels (One Straight and 1 Lateral Channel) 7. Injection Needle – Rigid 3 Fr – 10 	<p>01</p>

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department

		<p>8. Grasping Forceps for Stone fragments – 5Fr , Length 30 cms</p> <p>9. Grasping forceps for Large Stone fragments – 5 Fr, Length 30 cms</p> <p>4. Metal Tray, for sterilization and storage of operating cystourethroscope, perforated, lid with silicone bridges, external dimensions (w x d x h): 373 x 178 x 35 mm.</p> <p>5. Plastic Container for storage of accessories, suitable for gas and hydrogen peroxide sterilization as well as storage, external dimensions (w x d x h): 550 x 260 x 90 mm.</p> <p>6. US FDA/European CE Approved</p>	
5	Pediatric Open Surgical Instruments		
a)	Pediatric Retractor System - 1	<p>a. Flexible Table mounted retraction system for Pediatric Patients of all sizes</p> <p>b. Adjustable for smaller operative fields</p> <p>c. Should have small, independently-adjustable and removable frame arms to follow the contours of any pediatric operative field</p> <p>d. Should allow for multi-plane, multi-position hands-free retraction</p> <p>e. Should allow for fast and accurate set up</p> <p>f. Should be US FDA Approved</p> <p>g. Should be supplied with following components -</p> <p>i. Sterile Field Post – attachable to the operating table – 1</p> <p>ii. Support Arm – 1</p> <p>iii. Small Curved (wishbone) frame arm – 2</p> <p>iv. Snap on Clamps for attaching retractors – 6</p> <p>v. Rake Retractor (2.2x1.3 cms) – 1</p> <p>vi. Rake Retractor – 2.2x 2.5 cms) – 1</p> <p>vii. Mayo Swivel Retractors (5cms x 5cms)- 2</p> <p>viii. Mayo Swivel Retractors (7 cms x 5 cms) – 2</p> <p>ix. Mayo Swivel Retractor (2.5 cms x 2.5 cms) – 1</p> <p>x. Mayo swivel retractor (5 cms x 3.8 cms)- 2</p> <p>xi. Malleable Swivel Retractor – (2.5 cms x 2.5 cms) – 1</p> <p>xii. Malleable Swivel Retractor – (1.3cms x 2.5 cms) – 1</p> <p>xiii. Malleable Swivel Retractor (1.9x5 cms) – 1</p> <p>xiv. Malleable Swivel Retractor (1.9x7.6cms) - 1</p> <p>xv. Malleable Swivel Retractor (2.5x10.2cms) – 1</p> <p>xvi. Malleable Swivel Retractor (2.5x12.7 cms) – 1</p> <p>xvii. Malleable Swivel Retractor (2.5 x 15.2 cms) – 1</p> <p>xviii. Malleable Swivel Retractor (3.8x7.6cms) – 1</p> <p>xix. Malleable Swivel Retractor (3.8x12.7cms) – 1</p> <p>xx. Malleable Swivel Retractor (3.8x15.2 cms) – 1</p> <p>xxi. Splanchnic Swivel Retractor (3.8x8.9cms) – 2</p>	01
b)	Pediatric Open Surgical Instruments	<p>a. Finochietto Chest Retractor – Neonatal - 1</p> <p>1. Blades Size – 10mm x 30 mm</p> <p>2. Retractor Spread – 75 mm</p>	01

3. Arm Length – 50 mm
4. Rust Proof Stainless Steel
5. US FDA/European CE certification

- 2. Finochietto Chest Retractor – Neonatal - 2**
 1. Blades Size – 10mm x 15 mm
 2. Retractor Spread – 50 mm
 3. Arm Length – 50 mm
 4. Rust Proof Stainless Steel
 5. US FDA/European CE certification

- 3. Finochietto Chest Retractor – Infant – 2**
 1. Blades Size – 12mm x 40 mm
 2. Retractor Spread – 90 mm
 3. Arm Length – 75 mm
 4. Rust Proof Stainless Steel
 5. US FDA/European CE certification

- 4. Finochietto Chest Retractor – Small Infant – 2**
 1. Blades Size – 12mm x 34 mm
 2. Retractor Spread – 90 mm
 3. Arm Length – 75 mm
 4. Rust Proof Stainless Steel
 5. US FDA/European CE certification

- 5. Hoffman Shunt Passer - 2**
 1. Stainless Steel
 2. Resuable
 3. Suitable for subcutaneous tunneling for VP Shunt
 4. Tube – 3.2 mm Internal diameter, 4.2 mm Outer diameter
 5. Size - 35 – 42cms
 6. US FDA/European CE certification

- 6. Hoffman Shunt Passer - 2**
 1. Stainless Steel
 2. Resuable
 3. Suitable for subcutaneous tunneling for VP Shunt
 4. Tube – 3.2 mm Internal diameter, 4.2 mm Outer diameter
 5. Size - 60 - 65 cms
 6. US FDA/European CE certification

- 7. Ventricular Cannula – 2**
 1. For Hydrocephalus
 2. Reusable
 3. Stainless steel
 4. Closed end with three side holes
 5. Graduated
 6. Size – 5 Fr
 7. Length – 10 cms

8. US FDA/European CE certification

8. Ventricular Cannula – 2

1. For Hydrocephalus
2. Reusable
3. Stainless steel
4. Closed end with three side holes
5. Graduated
6. Size – 7 Fr
7. Length – 10 cms
8. US FDA/European CE certification

9. Ventricular Cannula – 2

1. For Hydrocephalus
2. Reusable
3. Stainless steel
4. Closed end with three side holes
5. Graduated
6. Size – 9 Fr
7. Length – 10 cms
8. US FDA/European CE certification

10. Mayo Dissecting Scissors – 15

1. Stainless steel
2. Length – 4 Inch
3. Curved
4. Blunt Tip
5. US FDA/European CE certification

11. Mayo Dissecting Scissors – 15

1. Stainless steel
2. Length – 6 Inch
3. Curved
4. Blunt Tip
5. Ring Handle
6. US FDA/European CE certification

12. Castroviejo blade breaker and holder for Ophthalmogy - 5

1. short model, squared jaws, round handle with slide lock
2. Length – 5 Inch
3. Tip – 12 mm, Blunt Tip
4. Straight with Round Knurnel Handle
5. US FDA/European CE certification

13. Swiss Model Blade Breaker and holder for Ophthalmology– 5

1. Pencil Type
2. Length – 90 mm
3. Tip – 12mm, Blunt tip
4. US FDA/European CE certification

14. Ribbon Type Malleable Retractor – 5

1. Size 1 1/2 inch width, Length – 13 inches
2. Malleable
3. Ribbon type
4. US FDA/European CE certification

15. Ribbon Type Malleable Retractor – 5

1. Size 1 inch width, Length – 13"
2. Malleable
3. Ribbon type
4. US FDA/European CE certification

16. Ribbon Type Malleable Retractor – 5

1. Size 1/2 inch width, Length – 7 inches
2. Malleable
3. Ribbon type
4. US FDA/European CE certification

17. Ribbon Type Malleable Retractor – 5

1. Size 10mm width, Length – 5 inches
2. Malleable
3. Ribbon type
4. US FDA/European CE certification

18. Denis Browne Abdominal Retractor – Child Size – 1

1. Ring/Frame Only
2. Size – 18x14 cms
3. Stainless Steel
4. Oval Sproket Frame
5. US FDA/European CE Certification

19. Denis Browne Abdominal Retractor – Adult Size – 1

1. Ring /Frame Only
2. Size – 25x18 cms
3. Stainless Steel
4. Oval Sproket Frame
5. US FDA/European CE

20. Valve Allien Retractor Blades for Denis Browne Abdominal Retractors – 2

1. 40x40 mm bades
2. US FDA/European CE

21. Valve Allien Retractor Blades for Denis Browne Abdominal Retractors – 2

1. 30x40 mm bades
2. US FDA/European CE

		<p>22. Valve Allien Retractor Blades for Denis Browne Abdominal Retractors – 2</p> <ol style="list-style-type: none"> 1. 50x40 mm bades 2. US FDA/European CE <p>* It should be supplied with two sterilization case from the same manufacturer.</p> <p>* The company must quote all items.</p>	
<p>c)</p>	<p>Instrument cases and trays</p>	<ol style="list-style-type: none"> 1. Instrument Sterlization Case/tray – 10 <ol style="list-style-type: none"> 1. Anodized Aluminum Case 2. Hinged top 3. Size 8"X14"X1" Inches with Cover 4. Full Silicone Finger Mat Below and slilicone cushion above 5. Should be good quality and durable. 6. Prior Demo required 2. Full Size Double Decker Laproscopic Instrument Tray – 5 <ol style="list-style-type: none"> 1. Should be suitable for holding full sized Laparoscopy Instruments 2. Should have holders for 3mm, 5mm and 10 mm Instruments 3. Should have silicone mat to protect the instruments 4. Size – 23inch x 11 inch x 8 inch 5. Should accomodate minimum 12 instruments 6. Should be made from High grade anodized aluminium 7. Should be good quality and durable. 8. Prior Demo required 3. Clear Top Telescope Trays – 10 <ol style="list-style-type: none"> 1. Should be suitable for securely holding Laparoscopic/cystoscopic telescopes 2. Length – 15 inches, Width – 2.5 inches, Height – 1.5 inch 3. Should be suitable for telescopes from 1mm to 10 mm 4. Should be able to accomodate two telescopes 5. Should have soft silicone base to prevent damage to instruments 6. Should have a robust locking mechanism to prevent inadvertant opening 7. Should be good quality and durable. 8. Prior Demo required 4. Instrument trays – 10 <ol style="list-style-type: none"> 1. Stainless steel 2. Size 20inch x 12 inch x 2.5 inches 3. With Cover 4. Should be good quality and durable. 5. Prior Demo required 5. Wire Baskets for Storage and Sterlization – 10 <ol style="list-style-type: none"> 1. Stainless steel 2. Size – 19 inch x 10 inch x 2 inches 	<p>01</p>

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department

		<ol style="list-style-type: none"> 3. Should be provided with compatible wire mesh cover 4. Should be good quality and durable. 5. Prior Demo required <p>6. Wire Baskets for Storage and Sterlization – 10</p> <ol style="list-style-type: none"> 1. Stainless steel 2. Size – 10 inch x 10 inch x 2 inches 3. Should be provided with compatible wire mesh cover 4. Should be good quality and durable. 5. Prior Demo required 										
6	Surgical Operating Loupes	<ul style="list-style-type: none"> • Magnification – 2.3 – 2.5X Magnification • Working Distance – 420 mm • Should have individually adjustable telescopes with interpupillary distance adjustment • Should include multiple adjustment options including height • Should have Suitable carrying case included with the loupe • Should be sufficiently light • Should have US FDA or European CE Certification 	02									
7	Patient Warming System	<ol style="list-style-type: none"> 1. Should be suitable for intra-operative applications for Pediatric Patients. 2. Should consist of active warming arm-cum-shoulder section, pair of leg segments and abdominal segment to cover the entire body. 3. Should be based on semiconductor polymer foil for precise warming of entire patient body during & after surgery. 4. Size <table style="margin-left: 20px; border: none;"> <tr> <td style="padding-right: 10px;">Abdominal Segment</td> <td style="padding-right: 10px;">:</td> <td>(55-60) cm x (25-30) cm</td> </tr> <tr> <td>Arm & Shoulder Section</td> <td>:</td> <td>(110-115) cm x (30-35) cm</td> </tr> <tr> <td>Leg Segment</td> <td>:</td> <td>(55-60) cm X (25-30) cm</td> </tr> </table> 5. Control unit should be capable of warming minimum four segments at a time. 6. Control unit should have Color TFT touch screen for easy operation. 7. Control unit should have touch screen display to select & display temperature of all segments at a time. 8. Control unit should automatically detect the number of segments which are connected to the unit and display the same on the screen. 9. Should offer precise digital temperature control with selectable temperature range of 30 to 40° C in steps of 0.1°C 10. Arm cum shoulder segment should be divided in two sections capable of being switched ON or OFF independently depending upon the nature of surgery and condition of patient. 11. Should have facility to measure & display the real time core body temperature of the patient continuously on the screen. 12. Should also have on screen graphical display of patient body temperature for the entire duration of surgery. 13. Should have facility to independently adjust the temperature of individual segment. 14. Should have a provision to connect whole body blanket, adult size blanket, jelly based warming mattress / pad to the same control unit for future requirement. 15. Should have safety features such as Automatic check, Precise 	Abdominal Segment	:	(55-60) cm x (25-30) cm	Arm & Shoulder Section	:	(110-115) cm x (30-35) cm	Leg Segment	:	(55-60) cm X (25-30) cm	03
Abdominal Segment	:	(55-60) cm x (25-30) cm										
Arm & Shoulder Section	:	(110-115) cm x (30-35) cm										
Leg Segment	:	(55-60) cm X (25-30) cm										

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department

		<p>temperature control between warming system and patient, Autostop on detecting any problem</p> <ol style="list-style-type: none">16. Should have non latex anti-bacterially coated, blood and fluid resistant, washable and replaceable covers17. Three replacable covers for each Segment (Abdominal/arm shoulder/leg) should be provided with each patient warming system18. Storage unit to store each segment should be provided.19. The control unit should be light weight, small in size and easily attachable to IV rod/OT table with fixing claw.20. Should have low energy consumption and noiseless operation21. Should confirm to international Quality and safety standards including European CE or US FDA ,European Directive 93/42/EEC	
--	--	--	--

Annexure-II

TECHNICAL BID

Name of Firm/Contractor/Supplier	
Complete Address & Telephone No.	
Name of Proprietor/Partner/Managing Director/Director.	
Phone No:- Mobile No:- Email Id:-	
Name and address of service centre nearby Jodhpur.	
Whether the firm is a registered firm Yes/No (attached copy of certificate).	
PAN No. (enclose the attested copy of PAN Card).	
Service Tax No. (enclose the attested copy of Service Tax Certificate).	
VAT No. (enclose the attested copy of VAT Certificate).	
Whether the firm has enclosed the Bank Draft/Pay Order/Banker's cheque of Earnest Money Deposit.	
Whether the Firm/Agency has signed each and every page of Tender/NIT.	
Please provide full list of consumables.	
Any other information, if necessary.	

Authorized signatory of the bidder with seal.

Annexure-III

MANUFACTURER's / PRINCIPAL's AUTHORIZATION FORM

To

The Administrative Officer,
All India Institute of Medical Sciences, Jodhpur

Sir,

TENDER: _____.

we, _____, who are established and reputable manufacturers of _____, having factories at _____ and _____, hereby authorize Messrs. _____ (name and address of agents) to bid, negotiate and conclude the contract with you against Tender

No. _____ for the above goods manufactured by us. No company or firm or individual other than Messrs. _____ are authorized to bid, negotiate and conclude the contract in regard to this business against this specific tender.

We hereby extend our full guarantee and warranty as per the conditions of tender for the goods offered for supply against this tender by the above firm.

The authorization is valid up to _____

Yours faithfully,

(Name)

For and on behalf of Messrs. _____
(Name of manufacturers)/Principal.

Annexure-IV

Financial Bid

A) FINANCIAL BID FOR DOMESTIC GOODS OR GOODS OF FOREIGN ORIGIN LOCATED WITHIN INDIA OR GOODS TO BE IMPORTED AND SUPPLIED AGAINST PAYMENT IN INDIAN RUPEES

1	2	3	4	5							6
Schedule	Brief Description of Goods	Country of Origin	Quantity (Nos.)	Price per unit (Rs.)							Total Price (at AIIMS-Jodhpur) basis (Rs.) = {4 x 5(g)}
				Ex - factory/ Ex-warehouse /Ex-showroom /Off-the shelf (a)	Excise Duty(if any) [%age & value] (b)	Sales Tax/ VAT(if any) [%age & value] (c)	Packing and Forwarding charges (d)	Inland Transportation, Insurance, loading/unloading and Incidental costs at AIIMS-Jodhpur (e)	Incidental Services (including Installation & Commissioning, Supervision, Demonstration and Training) at AIIMS-Jodhpur (f)	Unit Price (at AIIMS-Jodhpur) basis (g)= (a+b+c+d+e+f)	

Total Tender price in Rupees: _____

In words: _____

Note: -

1. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
2. The charges for Annual CMC after warranty shall be quoted separately.
3. The Bidder must quote price for "GOODS TO BE IMPORTED AND SUPPLIED AGAINST PAYMENT IN INDIAN RUPEES" after having taken in to account, the provision of Custom Duty Exemption Certificate (CDEC) by the Purchaser, as per Customs Tariff Act.

Place:

Date:

Name:

Business Address:

Signature of Bidder:

Seal of the Bidder:

Supply, Installation and Commissioning of Equipment required in Pediatric Surgery Department

Financial Bid

B) FINANCIAL BID FOR GOODS TO BE IMPORTED FROM ABROAD

1	2	3	4	5				6
Schedule	Brief Description of Goods	Country of Origin	Quantity (Nos.)	Price per unit (Rs.)				Total price on Destination + Insurance (local transportation and storage) = {4X 5 (d)}
				FOB price at port/ airport of Lading (a)	Carriage & Insurance (port of loading to port of entry) and other Incidental costs** (b)	Incidental Services (including Installation & Commissioning, Supervision, Demonstration and Training) at AIIMS-Jodhpur ** (c)	Unit Price on DDP AIIMS-Jodhpur + Extended Insurance (local transportation and storage) (d) = a+b+c	

** To be paid in Indian Currency (Rs.)

Total Tender price in foreign currency: _____

In words: _____

Note: -

1. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
2. The charges for Annual CMC after warranty shall be quoted.
3. The Bidder will be fully responsible for the safe arrival of the goods AIIMS-Jodhpur in good condition as per terms of DDP as per INCOTERMS, if applicable.

Indian Agent:

Indian Agency Commission - ___% of FOB

Place:

Date:

Name:

Business Address:

Signature of Bidder:

Seal of the Bidder:

Financial Bid

C) FINANCIAL BID FOR ANNUAL COMPREHENSIVE MAINTENANCE CONTRACT AFTER WARRANTY PERIOD:

1. S.No.	2. DESCRIPTION OF GOODS	3. QUANTITY. (Nos.)	4. Annual Comprehensive Maintenance Contract Cost for Each Unit year wise*.					5. Total Annual Comprehensive Maintenance Contract Cost for 5 Years [3 x (4a+4b+4c+4d+4e)]
			1 st	2 nd	3 rd	4 th	5 th	
			a	b	c	d	e	

* After completion of Warranty period.

Service Tax: Whether extra or inclusive, if extra, indicates the rate _____.

NOTE:-

1. In case of discrepancy between unit price and total prices, THE UNIT PRICE shall prevail.
2. The cost of Comprehensive Maintenance Contract (CMC) which includes preventive maintenance including testing & calibration as per technical/ service /operational manual, labour and spares, after satisfactory completion of Warranty period may be quoted as per NIT conditions on yearly basis for complete equipment and Turnkey (if any).
3. The cost of CMC may be quoted along with taxes applicable. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
4. Cost of CMC will be added for Ranking/Evaluation purpose.
5. All software updates should be provided free of cost during CMC period.
6. The stipulations in Technical Specification will supersede above provisions
7. The supplier shall keep sufficient stock of spares required during Annual Comprehensive Maintenance Contract period. In case the spares are required to be imported, it would be the responsibility of the supplier to import and get them custom cleared and pay all necessary duties.

Date:

Place:

Name:

Business Address:

Signature of Bidder:

Seal of the Bidder: