Tender

For

Equipments required for Department of Pulmonary Medicine

At

All India Institute of Medical Sciences, Jodhpur

NIT Issue Date : 15th January, 2016

NIT No. : Admn/Tender/Pulmonary Medicine/2015-AIIMS.JDH

Pre-Bid Meeting : 01st February , 2016 at 11:00 AM.

Last Date of Submission : 17th February , 2016 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur, Rajasthan-342005.

Telephone: 0291- 2012978, email: <u>aoadmin@aiimsjodhpur.edu.in</u> www.aiimsjodhpur.edu.in

All India Institute of Medical Sciences (AIMS), Jodhpur, Rajasthan, an apex healthcare institute being established by an Act of Parliament of India under aegis of Ministry of Health & Family Welfare, Government of India, invites sealed tenders for supply & installation of the following items for the Department of Pulmonary Medicine at the institute. You are requested to quote your best offer along with the complete details of specifications, terms & conditions.

Chapter-I

S.No	NIT No.	Item Description	Qty	EMD (Rs.)
01.	Admn/Tender/Pulmonary Medicine/1/2015 -AIIMS.JDH	Polysomnography lab equipments: a) Level 1 sleep Lab b) Universal titration device c) ETCO2 probe d) Esophageal manometry e) Actigraphy	1	50,000
02.	Admn/Tender/Pulmonary Medicine/2/2015 -AIIMS.JDH	Forced Oscillometry	1	24,000
03.	Admn/Tender/Pulmonary Medicine/3/2015 -AIIMS.JDH	Exhaled Nitric oxide/FeNO analyser	1	34,000
04.	Admn/Tender/Pulmonary Medicine/4/2015 -AIIMS.JDH	Portable USG with Echocardiography Probe	1	40,000
05.	Admn/Tender/Pulmonary Medicine/5/2015 -AIIMS.JDH	Fiberoptic bronchoscope (Adult)	1	12,000
06.	Admn/Tender/Pulmonary Medicine/6/2015 -AIIMS.JDH	Thin adult Videobronchoscope	1	32,000
07.	Admn/Tender/Pulmonary Medicine/7/2015 -AIIMS.JDH	Medical grade ultrasonic cleaner for biopsy forceps.	1	8,000

Quotation should be sealed and super-scribed with tender number and address to:

"Administrative Officer
All India Institute of Medical Sciences, Jodhpur
Basni, Phase-II
Jodhpur-342005, Rajasthan".

The sealed quotations should reach the Institute, latest by 17th February, 2016 at 03:00 PM and it will be opened on same day at 03:30 PM in the Conference Hall, Academic Block, AIIMS Jodhpur in the presence of the bidder(s) or their authorized representative(s), who will present at the scheduled date and time.

Terms & Conditions:

- 1. Preparation and Submission of Tender: The tender should be submitted in two parts i.e. Technical Bid and Financial Bid. The Technical Bid and the Financial Bid should be sealed by the bidder in two separate covers "Technical Bid for Tender for Supply of (Item Name)" and "Financial Bid for Tender for Supply of (Item Name)". Both Sealed Envelopes (separate for each NIT) should be kept in a main/ bigger envelope super-scribed as "Tender for Supply of (Item Name)".
- 2. Earnest Money Deposit: Earnest Money Deposit: The bidder shall be required to submit the Earnest Money Deposit (EMD) for an amount as per Chapter-I by way of demand drafts and Bank Guarantee only. The demand drafts shall be drawn in favour of "All India Institute of Medical Sciences, Jodhpur". The demand drafts and Bank Guarantee for earnest money deposit must be enclosed in the envelope containing the technical bid. The EMD of the successful bidder shall be returned after the successful submission of Bank Guarantee/ Security Deposit and for unsuccessful

Supply, Installation and Commissioning of Equipment required in Pulmonary Medicine Department bidder(s) it would be returned after award of the contract. Bid(s) received without demand drafts of EMD will be rejected.

The firms who are registered with National Small Industries Corporation (NSIC) / OR Small Scale Industrial (SSI) are exempted to submit the EMD (copy of registration must be provide along with).

- 3. Tender Fee: Tender fee will be Non-refundable amount of Rs. 1,000/- (Rupees One Thousand only) for Each Item by way of demand drafts only. The demand drafts shall be drawn in favour of "All India Institute of Medical Sciences, Jodhpur". Bids received without demand draft of tender will be rejected.
- **4. Validity:** The quoted rates must be valid for a period for 180 days from the date of closing of the tender. The overall offer for the assignment and bidder(s) quoted price shall remain unchanged during the period of validity. If the bidder quoted the validity shorter than the required period, the same will be treated as unresponsive and it may be rejected.
- 5. "PRE –BID Meeting" with the intending bidders shall be held on 01st February, 2016 from 11:00 A.M. onwards at AIIMS, Jodhpur. All the prospective bidders are requested to send comments/ representations on or before pre-bid meeting. Intending bidder will be allowed to seek clarification on specification, Conditions of Contract, etc. in writing to AIIMS, Jodhpur, within 48 hours after the pre-bid meeting.
- **6.** In case the tenderer withdraws, modifies or change his offer during the validity period, bid is liable to be rejected and the earnest money deposit shall be forfeited without assigning any reason thereof. The tenderer should also be ready to extend the validity, if required, without changing any terms, conditions etc. of their original tender.

7. Delivery and Installation:

i) For goods supplied from India:

All the goods ordered shall be delivered and Installed at AIIMS, Jodhpur within **30 days** from the date of issue of supply order.

ii) For goods imported directly from abroad:

All the goods ordered shall be delivered and Installed at AIIMS, Jodhpur within **60 days** from the date of opening of Letter of Credit for shipment.

All the aspects of safe delivery, installation and commissioning shall be the exclusive responsibility of the supplier. If the supplier fails to deliver, install and commission the goods on or before the stipulated date, then a penalty at the rate of 0.5% per week of the total order value shall be levied subject to maximum of 10% of the total order value. The successful tenderer will also provide required training for supplied items at AIIMS Jodhpur. The goods should be manufactured after adoption of latest technology.

If at any time during the currency of the contract, the supplier encounters conditions hindering timely of the goods and performance of services, the supplier shall promptly inform the AIIMS, Jodhpur for extension of the delivery schedule accordingly. On receiving the supplier's communication, the AIIMS, Jodhpur shall examine the situation as soon as possible and, at its discretion, may agree to extend the delivery schedule, with or without liquidated damages for completion of supplier's contractual obligations by issuing an amendment to the contract.

In the case of package supply where the delayed portion of supply materially hampers installation and commissioning of the systems, liquidated damages charges shall be levied as above on the total value of the concerned package of the purchase order. Quantum of liquidated damages assessed and levied by the purchaser shall be final and not challengeable by the supplier.

8. Signing the Contract: - The successful bidder shall be required to execute the Contract Agreement accepting all terms and conditions stipulated herein on a non-judicial stamp paper of Rs. 100/- (Rs. One Hundred only) along with performance security within fifteen days of the issue of the Letter of

Supply, Installation and Commissioning of Equipment required in Pulmonary Medicine Department notification of award. In the event of failure on the part of the successful bidder to sign the Contract within the period stipulated above, the EMD shall be forfeited and the acceptance of BID shall be considered as cancelled.

- 9. **Performance Security:** As a guarantee towards due performance and compliance of the contract work, the successful bidder (contractor) will deposit an amount equal to 10% of order value and should be kept valid for a period of 60 day beyond completion of all the contractual obligation, including CMC period towards security deposit by way of demand draft/ bank Guarantee in favour of "All India Institute of Medical Sciences, Jodhpur "drawn on any Nationalized Bank/Scheduled Bank and payable at Jodhpur within fifteen days of the issue of the Letter of notification of award along with non-judicial stamp paper of Rs. 100/- (Contract agreement).
- 10. Incidental Services: The supplier shall be required to perform the following services:
 - a. Installation & Commissioning, Supervision and Demonstration of the goods.
 - b. Providing required jigs and tools for assembly, minor civil works required for the completion of the installation.
 - c. On Site Training to Doctors/ Technicians/ Staff is to be provided by Supplier for operation and maintenance of the equipment for a period of 30 working days after successful installation of the machine, as per direction of user department.
 - d. Supplying required number of operation & maintenance manual for the goods.
 - e. To provide non-locked open software and standard interface inter-operability conditions for networked equipment's in hospital management information system, wherever applicable.
- 11. Accessories & Consumables: The separate price list of all accessories and consumables, if any, must be attached/ enclosed along with the Financial Bid.
- 12. After Sales Service: After sales service centre should be available on 24 (hrs.) X 7 (days) X 365 (days) basis. Complaints should be attended properly, maximum within 24 hrs to ensure an uptime of minimum 95%, wherever applicable, failing which the necessary penalty measures shall be enforced.

13. Inspection:

- a. AIIMS, Jodhpur shall have the right to inspect and/or to test the goods to confirm their conformity to the NIT Specifications at no extra cost to the Purchaser.
- b. AIIMS, Jodhpur right to inspect, test and, where necessary, reject the Goods after the goods arrival at the final destination shall in no way be limited or waived by reason of the Goods having previously been inspected, tested and passed by AIIMS, Jodhpur prior to the goods shipment.
- c. The Director, AIIMS Jodhpur shall be the final authority to reject full or any part of the supply which is not confirming to the specification and other terms and conditions.
- d. No payment shall be made for rejected Stores. Rejected items must be removed by the Bidders within two weeks of the date of rejection at their own cost and replaced immediately. In case these are not removed, these will be auctioned at the risk and responsibility of the suppliers without any further notice.

14. Documents:

- a. All pages of the Tender should be numbered and indexed.
- b. The bidder shall provide in its tender the required as well as the relevant documents like technical data, literature, drawings etc. to establish that the goods and services offered in the tender fully confirm to the goods and services specified by the purchaser in the tender documents. For this purpose the bidder shall also provide a clause-by-clause commentary on the technical specifications and other technical details incorporated by the purchaser in the tender documents to establish technical responsiveness of the goods and services offered in its tender duly indicating relevant page numbers in the product literature.
- c. The bidder shall provide a list of major Government and Private Institutions where its relevant bid item has been supplied during last one year.

- **15. Manufacturer Authorisation:** The bidder (if not original equipment manufacturer must submit Original Equipment Manufacturer authorization certificate that the tenderer is authorized for selling and maintain the equipment quoted for. Performa attached at **Annexure-III.**
- **16.** The bidders are required to submit user certificate for the relevant equipment on the letter head of the institution (Government/ Private).
- 17. The successful bidder will be required to submit order copies of the supply of the equipment in Government institutions in last 12 month for rate reasonability purpose.
- **18. Insurance:** The supplier shall make arrangements for insuring the goods against loss or damage incidental to manufacture or acquisition, transportation, storage and delivery. If the equipment's is not commissioned and handed over to AIIMS, Jodhpur within specified period, the insurance will have to be extended by the supplier at their cost till the successful installation, testing, commissioning and handing over of the goods to the AIIMS, Jodhpur.

19. Tender Currencies:

- a. The bidder supplying indigenous goods or already imported goods shall quote only in Indian Rupees. Further, imported goods to be imported and supplied by the bidder are also required to be quoted in Indian Rupees.
- b. For imported goods if supplied directly from abroad, prices shall be quoted in any freely convertible currency say US Dollar, Euro, GBP or Yen. As regards price(s) for allied services, if any, required with the goods, the same shall be quoted in Indian Rupees only, if such services are to be performed /undertaken in India.
- c. Tenders, where prices are quoted in any other way shall be treated as non -responsive and rejected.
- **20. Tender Prices:** While filling up the columns of the Financial Bid, the following aspects should be noted for compliance:

For domestic goods or goods of foreign origin located within India, the prices in the corresponding Financial Bid shall be entered separately in the following manner:

- a. The price of the goods, quoted ex-factory/ ex-showroom/ ex-warehouse/ off-the-shelf, as applicable, including all taxes and duties like sales tax, CST/ VAT, CENVAT, Custom Duty, Excise Duty etc. already paid or payable on the components and raw material used in the manufacture or assembly of the goods quoted ex-factory etc. or on the previously imported goods of foreign origin quoted ex-showroom etc.;
- b. Any sales tax or other taxes and any duties including excise duty, which will be payable on the goods in India if the contract is awarded;
- c. Charges towards Packing & Forwarding, Inland Transportation, Insurance, Loading/Unloading and other local costs incidental to delivery of the goods to their final destination as specified in the List of Requirements and Financial Bid;
- d. The price of Incidental Services, as mentioned in List of Requirements and Financial Bid;
- e. The prices of Turnkey (if any), as mentioned in List of Requirements, Technical Specification and Financial Bid; and
- f. The price of annual CMC, as mentioned in List of Requirements, Technical Specification and Financial Bid.

<u>For goods offered from abroad, the prices in the corresponding Financial Bid shall be entered</u> separately in the following manner:

- a. The price of goods quoted FOB port of shipment, as indicated in the List of Requirements and Financial Bid;
- b. The price of goods quoted CIF port of entry in India as indicated in the List of Requirements and Financial Bid:
- c. The price of goods quoted for delivery at AIIMS, Jodhpur as indicated in the List of Requirements, Financial Bid and Consignee List;

- d. Wherever applicable, the amount of custom duty with CDEC applicable on CIF value on the goods to be imported;
- e. The charges for Loading/Unloading, Inland transportation, Insurance and other local costs, Incidental cost to delivery of the goods from the port of entry in India to AIIMS, Jodhpur, as specified in the List of Requirements and Financial Bid;
- f. The charges for Incidental Services, as in the List of Requirements and Financial Bid;
- g. The prices of Turnkey (if any), as mentioned in List of Requirements, Technical Specification and Financial Bid; and
- h. The price of annual CMC, as mentioned in List of Requirements, Technical Specification and Financial Bid.

Additional information and instruction on Duties and Taxes: If the Bidder desires to ask for excise duty, sales tax/CST / VAT/ CENVAT, Custom Duty, Service Tax, Works Contract Tax etc. to be paid extra, the same must be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such duties and taxes and no claim for the same will be entertained later.

Excise Duty:

- a. If reimbursement of excise duty is intended as extra over the quoted prices, the supplier must specifically say so also indicating the rate, quantum and nature of the duty applicable. In the absence of any such stipulation it will be presumed that the prices quoted are firm and final and no claim on account of excise duty will be entertained after the opening of tenders.
- b. If a Bidder chooses to quote a price inclusive of excise duty and also desires to be reimbursed for variation, if any, in the excise duty during the time of supply, the Bidder must clearly mention the same and also indicate the rate and quantum of excise duty included in its price. Failure to indicate all such details in clear terms may result in rejection of that tender.
- c. Subject to sub clauses (i) & (ii) above, any change in excise duty upward/downward as a result of any statutory variation in excise duty taking place within contract terms shall be allowed to the extent of actual quantum of excise duty paid by the supplier. In case of downward revision in excise duty, the actual quantum of reduction of excise duty shall be reimbursed to the purchaser by the supplier. All such adjustments shall include all reliefs, exemptions, rebates, concession etc. if any obtained by the supplier.

<u>Sales Tax:</u> - If a bidder asks for sales tax/CST / VAT/CENVAT, Service Tax and Works Contract Tax to be paid extra, the rate and nature of sales tax applicable should be shown separately. The CST / VAT/CENVAT, Service Tax and Works Contract Tax will be paid as per the rate at which it is liable to be assessed or has actually been assessed provided the transaction of sale is legally liable to sales tax/ CST / VAT/CENVAT, Service Tax and Works Contract Tax and is payable as per the terms of the contract.

Octroi Duty and Local Duties & Taxes:- Normally, goods to be supplied to Government departments against Government contracts are exempted from levy of town duty, Octroi duty, terminal tax and other levies of local bodies. However, on some occasions, the local bodies (like town body, municipal body etc.) as per their regulations allow such exemptions only on production of certificate to this effect from the concerned Government department. Keeping this in view, the supplier shall ensure that the goods to be supplied by the supplier against the contract placed by the AIIMS, Jodhpur are exempted from levy of any such duty or tax and, wherever necessary, obtain the exemption certificate from the AIIMS, Jodhpur. However, if a local body still insists upon payment of such local duties and taxes, the same should be paid by the supplier to the local body to avoid delay in supplies and possible demurrage charges and obtain a receipt for the same. The supplier should forward the receipt obtained for such payment to the AIIMS, Jodhpur to enable the AIIMS, Jodhpur reimburse the supplier and take other necessary action in the matter.

<u>Customs Duty:</u> In respect of imported goods offered from abroad, the bidder shall specify the rate as well as the total amount of customs duty payable with Custom Duty Exemption Certificate, if applicable, on the quoted goods in the Financial Bid. The bidder shall also indicate the corresponding Indian Customs Tariff Number applicable for the goods.

- a. For transportation of imported goods offered from abroad, relevant instructions as incorporated shall be followed.
- b. For insurance of goods to be supplied, relevant instructions as provided shall be followed.
- c. Unless otherwise specifically indicated in this NIT document, the terms FCA, FOB, FAS, CIF, CIP etc. for imported goods offered from abroad, shall be governed by the rules & regulations prescribed in the current edition of INCOTERMS, published by the International Chamber of Commerce, Paris.
- d. The need for indication of all such price components by the bidders, as required in this clause is for the purpose of comparison of the tenders by the purchaser and will no way restrict the AIIMS, Jodhpur right to award the contract on the selected bidder on any of the terms offered.
- 21. Indian Agent:- If a foreign bidder has engaged an agent in India in connection with its bid, the foreign bidder, in addition to indicating Indian agent's commission, if any, shall also furnish the following information:
 - a. The complete name and address of the Indian Agent and its Permanent Account Number as allotted by the Indian Income Tax authority.
 - b. The details of the services to be rendered by the agent for the subject requirement.
 - c. Details of Service outlets in India, nearest to the AIIMS, Jodhpur to render services during Warranty and CMC period.

22. Firm Price

- a. Unless otherwise specified in the NIT, prices quoted by the bidder shall remain firm and fixed during the currency of the contract and not subject to variation on any account.
- b. However, as regards taxes and duties, if any, chargeable on the goods and payable, the conditions stipulated will apply.
- 23. Conversion of tender currencies to Indian Rupees: In case the bid document permits the bidders to quote their prices in different currencies, all such quoted prices of the responsive bidders will be converted to a single currency viz., Indian Rupees for the purpose of equitable comparison and evaluation, as per the closing exchange rates established by the Reserve Bank of India for similar transactions, as on the date of 'Last Date of Submission of Tender'.

24. Payment Terms:

i) Payment for goods supplied from India:

100% payment of the total order value shall be released after the successful installation/commissioning of the ordered goods against the submission of the inspection report.

ii) Payment for Imported goods:

For imported goods payment shall be made in the following manner:

- a) On shipment: 75 % payment of the contract price shall be paid 60 days after presentation of shipping documents {goods shipped shall be paid through irrevocable, non-transferable Letter of Credit (LC) opened in favour of the supplier in a bank in his country} and upon the submission of the following documents:
 - i. Four copies of Supplier's invoice showing contract number, goods description, quantity, unit price and total amount;
 - ii. Original and four copies of the clean, on-board Bill of Lading/Airway bill, marked freight prepaid and four copies of non-negotiable Bill of Lading/Airway bill.
 - iii. Insurance Certificate;
 - iv. Certificate of origin by the chamber of commerce of the concerned country;
 - v. Certificate of country of origin;
 - vi. Manufacture's / Supplier's warranty certificate;
 - vii. Manufacturer's own factory inspection report.
- **b)** On Acceptance: 25 % payment would be made after satisfactory installation & commissioning on issuance of Inspection certificate by the AIIMS, Jodhpur.

Note:- The supplier shall not claim any interest or any other payment under the contract.

25. Custom Clearance: For the Goods to be imported and supplied, the Institute will provide Custom Duty Exemption Certificate (CDEC) to successful bidder for availing concessional rate of duty as per prevailing Custom Tariff. In case, the bidder requires CDEC certificate, then the same should be specifically mentioned in the bid. The supplier is solely responsible for getting the material clearance from customs. Institute will provide all custom documents for custom clearance on the demand of supplier. Transportation of goods up to AIIMS, Jodhpur and its successful installation and commissioning is also the responsibility of the supplier. All charges/ expenses incurred in this process will be borne by the supplier. NO DEMURRAGE / WHARFAGE CHARGES WILL BE PAYBALE BY THE INSTITUTE UNDER ANY CIRCUMSTANCES. NO ADVANCE PAYMENT WILL BE PAYABLE FOR CUSTOM CLEARANCE/ FREIGHT/INSURANCE ETC.

26. Guarantee / Warrantee Period:

- **26.1** For the equipment value upto Rs. 5 Lakh: The Tenderers must quote for **02** years comprehensive warranty (Including all Spares, Accessories and Labour) from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. Also the bidders are requested to submit their quote (Rates) for subsequent **03** years Comprehensive Maintenance Contract (CMC) (Including All Spares, Accessories and Labour). Failure to comply this condition will entail the rejection of the bids. The price comparison shall be taking into account on basic price and post warranty CMC.
- **26.2** For the equipment value above Rs. 5 Lakh: The Tenderers must quote for **05** years comprehensive warranty (Including all Spares, Accessories and Labour) from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. Also the bidders are requested to submit their quote (Rates) for subsequent **05** years Comprehensive Maintenance Contract (CMC) (Including All Spares, Accessories and Labour). Failure to comply this condition will entail the rejection of the bids. The price comparison shall be taking into account on basic price and post warranty CMC.
- 27. Uptime guarantee: The firm should provide uptime guarantee of 95%

28. Downtime penalty Clause

- a. During the comprehensive warranty period, the guarantee uptime of 95% of 365 days will be ensured. In case the down time exceeds the 5% limit penalty of extension of guaranty period by two days for each additional day of down time will be enforced. The vendor must undertake to supply all spares for optimal upkeep of the equipment for at least FIVE YEARS after handling over the unit to the Institute. If accessories / other attachment of the system are procured from the third party, then the vendor must produce cost of accessory / other attachment and the CMC from the third party separately along with the main offer and the third party will have to sign the CMC with the Institute if required.
- b. The principals or their authorized service providers are required to submit a certificate that they have satisfactory service arrangements and fully trained staff available to support the uptime guarantee.
- 29. Arbitration: If any difference arises concerning this agreement, its interpretation on payment to the made there-under, the same shall be settled out by mutual consultation and negotiation. If attempts for conciliation do not yield any result within a period of 30 days, either of the parties may make a request to the other party for submission of the dispute for decision by an Arbitral Tribunal containing Sole Arbitrator to be appointed by the Secretary, Department of Legal Affairs. Such requests shall be accompanied with a panel of names of three persons to act as the sole arbitrator. In case of such arbitrator refusing, unwilling or becoming incapable to act or his mandate having been terminated under law, another arbitrator shall be appointed in the same manner from among the panel of three persons to be submitted by the claimant. The provision of Arbitration and Conciliation Act, 1990 and the rule framed there under and in force shall be applicable to such proceedings.
- **30. Subletting of Work:** The firm shall not assign or sublet the work/job or any part of it to any other person or party without having first obtained permission in writing of AIIMS, Jodhpur, which will be

Supply, Installation and Commissioning of Equipment required in Pulmonary Medicine Department at liberty to refuse if thinks fit. The tender is not transferable. Only one tender shall be submitted by one tenderer.

- **31. Breach of Terms and Conditions:** In case of breach of any terms and conditions as mentioned above, the Competent Authority, will have the right to cancel the work order/ job without assigning any reason thereof and nothing will be payable by AIIMS, Jodhpur in that event the security deposit shall also stands forfeited.
- **32. Insolvency etc:** In the event of the firm being adjudged insolvent or having a receiver appointed for it by a court or any other order under the Insolvency Act made against them or in the case of a company the passing any resolution or making of any order for winding up, whether voluntary or otherwise, or in the event of the firm failing to comply with any of the conditions herein specified AIIMS, Jodhpur shall have the power to terminate the contract without any prior notice.
- 33. Force Majeure: If, at any time during the subsistence of this contract, the performance in whole or in part by either party of any obligation under this contract is prevented or delayed by reasons of any war or hostility, act of public enemy, civil commotion, sabotage, fire, floods, explosion, epidemics, quarantine restriction, strikers lockout or act of God (hereinafter referred to as events) provided notice of happening of any such eventuality is given by party to other within 21 days from the date of occurrence thereof, neither party hall by reason of such event be entitled to terminate this contract nor shall either party have any claim for damages against other in respect of such non-performance or delay in performance, and deliveries have been so resumed or not shall be final and conclusive.

Further, that if the performance in whole or in part of any obligation under this contract is prevented or delayed by reason of any such event for a period exceeding 60 days, either party may, at least option to terminate the contract.

- **34.** Bidder shall submit a copy of the tender document and addenda thereto, if any, with each page of this document should be signed and stamped to confirm the acceptance of the entire terms & conditions as mentioned in the tender enquiry document.
- **35.** The quantity of item given in the tender is tentative, which may be increased or decreased as per the institute's requirement.
- **36.** Signed & stamped compliance sheet of the technical specification of the goods with technical printed literature must be enclosed with the bid.
- **37.** After due evaluation of the bid(s) Institute will award the contract to the lowest evaluated responsive tenderer
- **38.** Conditional bid will be treated as unresponsive and it may be rejected.
- **39. Demonstration:** AIIMS Jodhpur reserves the right to ask the tenderers for arranging demonstration of their equipment for which rates have been quoted, to the concerned committee, if required.
- **40.** The Institute reserves the right to accept in part or in full or reject any or more tender(s) without assigning any reasons or cancel the tendering process and reject all tender(s) at any time prior to award of contract, without incurring any liability, whatsoever to the affected bidder or bidder(s).

41. Applicable Law:

• The contract shall be governed by the laws and procedures established by Govt. of India, within the framework of applicable legislation and enactment made from time to time concerning such Commercial dealings / processing.

- Any disputes are subject to exclusive jurisdiction of Competent Court and Forum in Jodhpur, Rajasthan, India only.
- The Arbitration shall be held in accordance with the provisions of the Arbitration and Conciliation Act, 1996 and the venue of arbitration shall be at Jodhpur. The decision of the Arbitrator shall be final and binding on both the partied.
- Force Majeure: Any delay due to Force Majeure will not be attributable to the supplier.

Administrative Officer AIIMS, Jodhpur

Item No.	Name of Items	Details Description and Technical Specification	Qty
No. 01	Polysomnography	1. It should be US FDA approved	
U1	Machine	2. It should have following channels and parameters.	
	(Portable)	Pressure-based airflow with snoring detection	
	(1 of table)	Pulse rate	-
		• SpO2	-
		Chest or abdominal effort: Z rip technology	_
		Body position changes	
		Pressure from Sleep Therapy device	
		Patient event button	
		Software should have facility to auto scoring and manual scoring	
		3. Maximum recording time 20 Hours	
		4. Sample rates	1 Nos.
		Airflow: 200 Hz (respiration)	1 1103.
		• Effort : 200 Hz	
		• Oximetry: 80 Hz	
		• Pressure therapy parameters: * 0.2 to 10 Hz	
		5. Minimum battery life 17 Hours	
		6. Parameter ranges	
		• Flow 4 – 40 BPM	
		• Effort 4 – 40 BPM	
		• Position Supine and non-supine SpO2 50 – 100%	
		• Pulse rate 20 – 200 BPM Pressure* 0 – 30 cm H2O	
		7. Manufacture should preferable having office in India	1
		8. Auto CPAP macine- USFDA approved	
	Sleep Lab	1 Description of Function	1 Set
	Polysomnography	1.1 Polysomnography and sleep studies are performed to	1 361
	System along	diagnose illness or syndromes related to sleep	
	with the standard	2. Operational Requirements	1
	items/ accessories	2.1 System completes with all hardware and software is required.	1
	mentioned in	3. Technical Specification	1
	Technical	3.1 Total Number of Channels should be minimum 64	
	Specification		
	column	a) EEG input should be: Minimum 32 b) The machine should have EMG input of:minimum 3	1
		c) The machine should have EMG input of .iminimum 3	1
		d) The machine should have ECG: 2 d) The machine should have ECG: 4 (2 Physical and 2 derived)	-
			_
		e) Nasal /Oral/Airflow: 1	_
		f) Flow (Thermal): 1	
		l D '	
		g) Respiratory Effort , zRip : 2	
		h) Snoring, Microphone: 1	
		h) Snoring, Microphone : 1 i) Body Position : 1	
		h) Snoring, Microphone : 1 i) Body Position : 1 j) CPAP Pressure : 1	- - - -
		h) Snoring, Microphone : 1 i) Body Position : 1 j) CPAP Pressure : 1 k) Limb Movement : 2	-
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1	
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1 m) Pulse rate: 1	-
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1 m) Pulse rate: 1 n) Light sensor: 1	
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1 m) Pulse rate: 1 n) Light sensor: 1 o) Sensors 5 sets (each set includes flow)	
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1 m) Pulse rate: 1 n) Light sensor: 1 o) Sensors 5 sets (each set includes flow) p) etCO2 sensor compatible with sleep lab: 1	
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1 m) Pulse rate: 1 n) Light sensor: 1 o) Sensors 5 sets (each set includes flow) p) etCO2 sensor compatible with sleep lab: 1 q) Sensor(Two sets)- pediatric and adult, microphone, pulse	
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1 m) Pulse rate: 1 n) Light sensor: 1 o) Sensors 5 sets (each set includes flow) p) etCO2 sensor compatible with sleep lab: 1	
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1 m) Pulse rate: 1 n) Light sensor: 1 o) Sensors 5 sets (each set includes flow) p) etCO2 sensor compatible with sleep lab: 1 q) Sensor(Two sets)- pediatric and adult, microphone, pulse	
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1 m) Pulse rate: 1 n) Light sensor: 1 o) Sensors 5 sets (each set includes flow) p) etCO2 sensor compatible with sleep lab: 1 q) Sensor(Two sets)- pediatric and adult, microphone, pulse Oximeter sensor and leads of ECG, EEG, Leg & chin EMG, EOG	
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1 m) Pulse rate: 1 n) Light sensor: 1 o) Sensors 5 sets (each set includes flow) p) etCO2 sensor compatible with sleep lab: 1 q) Sensor(Two sets)- pediatric and adult, microphone, pulse Oximeter sensor and leads of ECG, EEG, Leg & chin EMG, EOG 3.2 Additional DC channels for external peripherals like	
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1 m) Pulse rate: 1 n) Light sensor: 1 o) Sensors 5 sets (each set includes flow) p) etCO2 sensor compatible with sleep lab: 1 q) Sensor(Two sets)- pediatric and adult, microphone, pulse Oximeter sensor and leads of ECG, EEG, Leg & chin EMG, EOG 3.2 Additional DC channels for external peripherals like capnography, PH, blood pressure Monitor etc.	
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1 m) Pulse rate: 1 n) Light sensor: 1 o) Sensors 5 sets (each set includes flow) p) etCO2 sensor compatible with sleep lab: 1 q) Sensor(Two sets)- pediatric and adult, microphone, pulse Oximeter sensor and leads of ECG, EEG, Leg & chin EMG, EOG 3.2 Additional DC channels for external peripherals like capnography, PH, blood pressure Monitor etc. 3.3 Other Technical Specifications	
		h) Snoring, Microphone: 1 i) Body Position: 1 j) CPAP Pressure: 1 k) Limb Movement: 2 l) SaO2: 1 m) Pulse rate: 1 n) Light sensor: 1 o) Sensors 5 sets (each set includes flow) p) etCO2 sensor compatible with sleep lab: 1 q) Sensor(Two sets)- pediatric and adult, microphone, pulse Oximeter sensor and leads of ECG, EEG, Leg & chin EMG, EOG 3.2 Additional DC channels for external peripherals like capnography, PH, blood pressure Monitor etc. 3.3 Other Technical Specifications a) Pressure transducer for nasal air flow	

AIIMS-Jodhpur Page 12

protection for 60 minutes back up.

6.1 Power input to be 220-240VAC, 50Hz fitted with Indian plug 6.2 UPS of suitable rating with voltages regulation and spike

6 Power Supply

	7 Standards, Safety and Training	
	7.1 Should be US FDA approved	
	7.2Manufacturer should have ISO certification for quality standards.	
	7.3Comprehensive training for lab staff and support services	
	till familiarity with the system should be provided.	
	7.4 Electrical safety should conform to standards for electrical safety IEC60601-1(OR EQUIVALENT international/national standard) General requirement for Electrical safety of	
	Medical Equipment.	
	8 Documentation	
	8.1User/Technical/Maintenance manuals to be supplied in English.	
	8.2 List of important spare parts and accessories with their part	
	number and costing.	
	8.3 Compliance Report to be submitted in a tabulated and point	
	wise manner clearly mentioning the page/para number of original	
	catalogue/data sheet. Any point, if not substantiated with	
	authenticated catalogue/manual, will not be considered.	
	8.4 Certificate of calibration and inspection.	
	8.5 List of Equipment available for providing calibration and	
	routine Preventive Maintenance Support. as per manufacturer	
	documentation in service/technical manual	
Forced/Impulse Oscillometry	1. It should be PC based using Forced Oscillations Technique (FOT).	01
•	2. The Forced Oscillation should be based on Signal Optimized	
	Pseudorandom Noise (PRN) between 4Hz and 48Hz.	
	3. It should measure Total Respiratory Impedance: Resistance and	
	Reactance	
	4. Auto BTPS correction	
	5. System should be suitable for use in both children and adults patients.	
	6. Processing time should be preferably less than 10 seconds.	
	7. Should meet all international safety standards, US FDA and	
	European CE for human applications	
	8. Complete disassembly for disinfection and sterilization	
	9. All necessary software should be supplied with the equipment.	
	10. The software should be latest and original.	
	Essential Accessories:	
	1. The system should be supplied with all essential accessories required.	
	2. Antibacterial Filters 500	
	3. Nose Clips 5	
	4. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 4GB RAM, 1TB	
	Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB	
	(3.0) port)	
	5. Laser Colour Printer: 01	
	6. UPS (1 KVA): 01	
	7. Disposable Mouthpiece to perform the test on 500 patients	
	should be provided. 8. All essential accessories to make unit fully functional should	

Schaled Nitric Schalation parameters: Schalation Flow: 50 + 5 m/Sec Schalation pressure: 10 - 20 cm/D2 Concentration range: 5-300 ppb mire oxide Accuracy: ± 5ppb of measured value 50ppb, ± 10% of measured value >50ppb Sensor sensitivity: 5ppb Breath test time: Adult 10 - 12 seconds / Child 10 seconds Response time: < 2 min Temp stabilisation time: < 30 min Ambient NO measurement time: < 4 min Operating temp range: 16 - 30 or Operating relative hemidity: 20 - 60 % Rh Sensor operating life: 1 year Maximum ambient NO concentration: < 300 ppb Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be latest and original. Essential Accessories: 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 78 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3:0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to signain Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3:0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to support speckle reduction imaging for hetter citissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D. Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System should support transducers. 6. System must display at a maximum depth of 35 cm. 9. The system must have a dedicat	Suppiy, 1	1		Λ1
Exhalation pressure: 10 − 20 cmH2O Concentration range: 5-300 pph nitric oxide Accuracy: ± 5pph of measured value ≤50ppb, ± 10% of measured value >50ppb Sensor sensitivity: 5pph Breath test time: Adult 10 − 12 seconds / Child 10 seconds Response time: ≥ 2 min Temp stabilisation time: < 30 min Ambient NO measurement time: < 4 min Operating temp range: 16 − 30 oC Operating relative humidity: 20 − 60 % Rh Sensor operating life: 1 year Maximum ambient NO concentration: < 300 ppb Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches I.CD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 78 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 4 Portable USG with Echocardiography Probe 4 Portable current of the composition of the proper	3		<u>-</u>	01
Concentration range: 5-300 ppb nitric oxide Accuracy: 4 ppb of measured value ≤50ppb, ± 10% of measured value ≤50ppb Sensor sensitivity: 5ppb Breath test time: Adult 10 - 12 seconds / Child 10 seconds Response time: 2 min Temp stabilisation time: <30 min Ambient NO measurement time: <4 min Operating emp range: 16 - 30 oC Operating relative humidity: 20 - 60 % Rh Sensor operating life: 1 year Maximum ambient NO concentration: <300 ppb Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should he hatest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM. ITB Hard disc, original Windows 78.8 with original Microsoft Office and DVD RW and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 7.				
Accuracy: + 5ppb of measured value ≤50ppb, + 10% of measured value >50ppb Sensor sensitivity: 5ppb Breath test time: Adult 10 - 12 seconds / Child 10 seconds Response time: ≥ 2 min Temp stabilisation time: < 30 min Ambient NO measurement time: < 4 min Operating temp range: 16 - 30 oC Operating temp range: 16 - 30 oC Operating temp range: 18 oC Operating temp range: 18 oC Operating temp range:		analyser		
scansor sensitivity: 5ppb Breath test time: Adult 10 - 12 seconds / Child 10 seconds Response time: < 2 min Temp stabilisation time: < 30 min Ambient NO measurement time: < 4 min Operating temp range: 16 - 30 oC Operating relative humidity: 20 - 60 % Rh Sensor operating life: 1 year Maximum ambient NO concentration: < 300 ppb Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, ITB Hard disc, original Windows 78 with original Microsoft Office and DVD R'W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 7				
Sensor sensitivity: 5ppb Breath test time: Adult 10 - 12 seconds / Child 10 seconds Response time: < 2 min Temp stabilisation time: < 30 min Ambient NO measurement time: < 4 min Operating temp range: 16 - 30 oC Operating relative humidity: 20 - 60 % Rh Sensor operating life: 1 year Maximum ambient NO concentration: < 300 ppb Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, TTB Hard disc, original Windows 7/8 with original Microsoft Office and DVD RW and atleast 4 USB port: 01 (Winimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 4 Portable USG with Echocardiography Probe 4 Portable compound imaging with at least 5 sights of lines for better cardiac contrast resolution, itssue differentiation and edge detection, equivalent to high end cart based systems. 5. System should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, itssue differentiation and edge detection, equivalent to high end cart based systems. 5. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 5. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 6. System should support transducer technologies like TEE (Transesophageal echo, phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must display at a maximum depth of 35 cm.				
Breath test time: Adult 10 - 12 seconds / Child 10 seconds Response time: < 2 min Temp stabilisation time: < 30 min Ambient NO measurement time: < 4 min Operating temp range: 16 - 30 o.C Operating relative humidity: 20 - 60 % Rh Sensor operating life: 1 year Maximum ambient NO concentration: < 300 ppb Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be lates and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM. TIB Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 4 Portable USG with Echocardiogrphy Probe 1. Unit should be able to give high quadity images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end card based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must have fast start up to scanning in less than 30 sec			* *	
Response time: < 2 min Temp stabilisation time: < 30 min Ambient NO measurement time: < 4 min Operating temp range: 16 - 30 oC Operating relative humidity: 20 - 60 % Rh Sensor operating life: 1 year Maximum ambient NO concentration: < 300 ppb Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 7% with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA); 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge eduction, equivalent to high end card based systems. 2. System should be able to support speckle reduction imaging for better itsue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System should support transducer technologies like TEE (Transesphageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 9. The system shall process a dynamic range that is at least 165db. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and			Sensor sensitivity: 5ppb	
Temp stabilisation time: < 30 min Ambient NO measurement time: < 4 min Operating temp range: 16 - 30 oC Operating relative humidity: 20 - 60 % Rh Sensor operating life: 1 yever Maximum ambient NO concentroin: < 300 ppb Should meat all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 78 with original Microsoft Office and DVD RW and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA):			Breath test time: Adult 10 - 12 seconds / Child 10 seconds	
Temp stabilisation time: < 30 min Ambient NO measurement time: < 4 min Operating temp range: 16 - 30 oC Operating relative humidity: 20 - 60 % Rh Sensor operating life: 1 yever Maximum ambient NO concentroin: < 300 ppb Should meat all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 78 with original Microsoft Office and DVD RW and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA):				
Ambient NO measurement time: «4 min Operating temp range: 16 - 30 oC Operating relative humidity: 20 - 60 % Rh Sensor operating life: 1 year Maximum ambient NO concentration: < 300 ppb Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 78 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA); 01 4 Portable USG with Echocardiogrphy Probe 1 Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system shall process a dynamic range that is at least 165db. The system must flave dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to br				
Operating temp range: 16 - 30 oC Operating relative humidity: 20 - 60 % Rh Sensor operating life: 1 year Maximum ambient NO concentration: < 300 ppb Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 17B Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 8. System should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 5. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 5. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 5. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 6. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 7. Crine memory on all modes. 8. The system should support transducer technologies			<u> </u>	
Operating relative humidity: 20 - 60 % Rh Sensor operating life: 1 year Maximum ambient NO concentration: < 300 ppb Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA); 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, itssue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support specckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System shall process a dynamic range that is at least 165db. The system must have fast start up to scanning in less than 50 seconds from off condition, for use in critical and emergency situations. 9. The system must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 9. The system must have a addicated cardiac accludation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact				
Sensor operating life: 1 year Maximum ambient NO concentration: < 300 ppb Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 78 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 8. System should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cline memory on all modes. 8. The system must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be compact, portable and lightweight, weighing less than 5 kg. 12. Flat LCI/ TFT monitor of				
Maximum ambient NO concentration: < 300 ppb Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end card based systems. 2. System should be able to support speckle reduction imaging for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end card based systems. 2. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system shuld be hard surface. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TTF monitor o				
Should meet all international safety standards, US FDA and European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 78 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, itsuse differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system sust have a dedicated cardiac calculation packages with PISA, TDI calculation package and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to santitiv				
European CE for human applications All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 78 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/TFI monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid				
All necessary software should be supplied with the equipment. The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM. ITB Hard disc, original Windows 78 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 7. U			· · · · · · · · · · · · · · · · · · ·	
The software should be latest and original. Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system shall process a dynamic range that is at least 105db. The system must display at a maximum depth of 35 cm. 1. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be compact, portable and lightweight, weighing less than 5 kg. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including ke				
Essential Accessories: 1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 17B Hard disc, original Windows 7/8 with original Microsoft Office and DVD RW and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system must display at a maximum depth of 35 cm. 9. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid			All necessary software should be supplied with the equipment.	
1. The system should be supplied with all essential accessories required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system hall process a dynamic range that is at least 165db. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid			The software should be latest and original.	
required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid			Essential Accessories:	
required. 2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid				
2. Disposable Antibacterial Filters 500 3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid				
3. Sensor for 500 tests 4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall for hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid			•	
4. Nose Clips 5 Nos 5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid			*	
5. Compatible Desktop PC all in one i.e. inbuilt CPU with atleast 19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 78 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid				
19 inches LCD monitor, with atleast 8GB RAM, 1TB Hard disc, original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 4 Portable USG with Echocardiogrphy Probe 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid			*	
original Windows 7/8 with original Microsoft Office and DVD R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid				
R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port) 6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid				
6. Laser Colour Printer: 01 7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid				
7. UPS (1 KVA): 01 1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid			R/W and atleast 4 USB port: 01 (Minimum 2 USB (3.0) port)	
1. Unit should be able to give high quality images with advance technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must display at a maximum depth of 35 cm. 9. The system must display at a maximum depth of 35 cm. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid			6. Laser Colour Printer: 01	
technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid			7. UPS (1 KVA): 01	
technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid				
Fechocardiogrphy Probe for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	Portable USG	1. Unit should be able to give high quality images with advance	01
detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4			01
 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid 	4	with	technologies like compound imaging with at least 5 sights of lines	01
better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge	01
3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems.	01
zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for	01
 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid 	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement.	01
Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write)	01
available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility.	01
5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave	01
seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be	01
situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers.	01
 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid 	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30	01
esophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency	01
esophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency	01
7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations.	01
8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Trans-	01
The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc.	01
9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes.	01
with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db.	01
package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm.	01
10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages	01
less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations	01
11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package.	01
against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing	01
12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg.	01
image. 13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit	01
13. Alphanumeric soft keys keyboard with easy access scans controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface.	01
controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free	01
controls, facility to sanitize the system including keyboard to avoid	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free	01
	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image.	01
	4	with Echocardiogrphy	technologies like compound imaging with at least 5 sights of lines for better cardiac contrast resolution, tissue differentiation and edge detection, equivalent to high end cart based systems. 2. System should be able to support speckle reduction imaging for better tissue differentiation and edge enhancement. 3. System should have both online (Read) as well as offline (Write) zoom facility. 4. Imaging modes of Real time 2D, Colour Doppler, Pulsed wave Doppler, Continuous wave Doppler, Power Doppler must be available on all cardiac transducers. 5. System must have fast start up to scanning in less than 30 seconds from off condition, for use in critical and emergency situations. 6. System should support transducer technologies like TEE (Transesophageal echo), phased array, convex, linear etc. 7. Cine memory on all modes. 8. The system shall process a dynamic range that is at least 165db. The system must display at a maximum depth of 35 cm. 9. The system must have a dedicated cardiac calculation packages with PISA, TDI calculation packages and vascular calculations package. 10. The unit must be compact, portable and lightweight, weighing less than 5 kg. 11. Unit must be sturdy, resistant to breakage & damage on hit against the wall or hard surface. 12. Flat LCD/ TFT monitor of at least 15 inches with flicker free image. 13. Alphanumeric soft keys keyboard with easy access scans	01

Supply,		nmissioning of Equipment required in Pulmonary Medicine Depar	
FFIJ		14. The system must have the ability to function by AC/DC or	
		battery power with the same degree of functionality, the battery life	
		(run time) shall be al least 2 (Two) hours, this need to be	
		demonstrated.	
		15. The system must have archive capability for storage and	
		retrieval of images and clips data.	
		16. The system shall support the all DICOM functionality, Storage,	
		Print, and Work List, also ready to connect to PACS.	
		17. System should possess software for Enhanced Needle	
		Visualization to track the needle clearly at steep angles during the	
		procedures while maintaining striking image quality of the target	
		structures and the surrounding anatomy with simple On/Off	
		functionality. This Facility should be available on both High	
		frequency Linear and Curvilinear probes for superficial as well as	
		deeper blocks.	
		18. The system and transducers should be both USFDA and	
		European CE certified.	
		19. Following transducers to be supplied as standard	
		a. Curvilinear 2-5 MHz +/-1 with foot print size 50-60 mm for	
		Nerve, abdominal & FAST applications.	
		b. High Frequency Linear transducer 6-12 MHz +/-1 with foot print	
		in between 35-40 mm for vascular access, small parts, Nerve	
		Blocks.	
		c. 1-5 MHz +/-1 broad band phased array transducer18-22 mm for	
		cardiac applications.	
		d.5-12 MHz +/-1 with foot print size less than 30 mm broadband	
		linear array for vascular, nerve, superficial, and venous ultrasound	
		e. Mobile Cart with transducer holder.	
		20. The firm must have minimum of 25 installations of the same	
		model in India, attach list of installations, and also provide	
		performance certificates.	
5	Eibanantia	1 E' 11 CX' 1 111 120 1	
J	Fiberoptic	1. Field of View should be 120 degree or more	01
5	Bronchoscope	2. Depth of field should be $3 - 50$ mm or better	01
5	_	 2. Depth of field should be 3 – 50 mm or better 3. Distal end diameter should be 4.9 mm or less 	01
J	Bronchoscope	 2. Depth of field should be 3 – 50 mm or better 3. Distal end diameter should be 4.9 mm or less 4. Insertion tube diameter should be 5.0 mm or less 	01
J	Bronchoscope	 2. Depth of field should be 3 – 50 mm or better 3. Distal end diameter should be 4.9 mm or less 4. Insertion tube diameter should be 5.0 mm or less 5. Channel diameter should be 2.2 mm or more 	01
J	Bronchoscope	 2. Depth of field should be 3 – 50 mm or better 3. Distal end diameter should be 4.9 mm or less 4. Insertion tube diameter should be 5.0 mm or less 5. Channel diameter should be 2.2 mm or more 6. Should be light weight and easy to use 	01
J	Bronchoscope	 2. Depth of field should be 3 – 50 mm or better 3. Distal end diameter should be 4.9 mm or less 4. Insertion tube diameter should be 5.0 mm or less 5. Channel diameter should be 2.2 mm or more 6. Should be light weight and easy to use 7. Working length should be 600 mm or more 	01
J	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree 	01
J	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece 	01
J	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water 	01
J	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. 	01
J	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure 	01
J	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. 	01
J	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. 	01
J	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester Unit should be compact and light weight. 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester Unit should be compact and light weight. Light source minimum 150 watt halogen/LED. 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester Unit should be compact and light weight. Light source minimum 150 watt halogen/LED. Should have Inbuilt air pump. 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester Unit should be compact and light weight. Light source minimum 150 watt halogen/LED. Should have Inbuilt air pump. Essential Accessories 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester Unit should be compact and light weight. Light source minimum 150 watt halogen/LED. Should have Inbuilt air pump. Essential Accessories Ten reusable cups with needle type biopsy forceps compatible 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester Unit should be compact and light weight. Light source minimum 150 watt halogen/LED. Should have Inbuilt air pump. Essential Accessories Ten reusable cups with needle type biopsy forceps compatible with the scope channel diameter. 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester Unit should be compact and light weight. Light source minimum 150 watt halogen/LED. Should have Inbuilt air pump. Essential Accessories Ten reusable cups with needle type biopsy forceps compatible with the scope channel diameter. Five reusable cup biopsy forceps compatible with the scope 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester Unit should be compact and light weight. Light source minimum 150 watt halogen/LED. Should have Inbuilt air pump. Essential Accessories Ten reusable cups with needle type biopsy forceps compatible with the scope channel diameter. Five reusable cup biopsy forceps compatible with the scope channel diameter. 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester Unit should be compact and light weight. Light source minimum 150 watt halogen/LED. Should have Inbuilt air pump. Essential Accessories Ten reusable cups with needle type biopsy forceps compatible with the scope channel diameter. Five reusable alligator type biopsy forceps compatible with the 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester Unit should be compact and light weight. Light source minimum 150 watt halogen/LED. Should have Inbuilt air pump. Essential Accessories Ten reusable cups with needle type biopsy forceps compatible with the scope channel diameter. Five reusable alligator type biopsy forceps compatible with the scope channel diameter. Five reusable alligator type biopsy forceps compatible with the scope channel diameter. 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester Unit should be compact and light weight. Light source minimum 150 watt halogen/LED. Should have Inbuilt air pump. Essential Accessories Ten reusable cups with needle type biopsy forceps compatible with the scope channel diameter. Five reusable alligator type biopsy forceps compatible with the scope channel diameter. Five reusable alligator type biopsy forceps compatible with the scope channel diameter. Ten transbronchial aspiration needles. 	01
	Bronchoscope	 Depth of field should be 3 – 50 mm or better Distal end diameter should be 4.9 mm or less Insertion tube diameter should be 5.0 mm or less Channel diameter should be 2.2 mm or more Should be light weight and easy to use Working length should be 600 mm or more Angulation should be Up 180 degree and Down 130 degree Should have telescopic eye piece Can be fully immersed in disinfectant solution and water Should have autoclavable suction valve. Should have facility to check leakage by automatic pressure regulated leak tester. Should be compatible halogen light source. Must have USFDA/ European CE certification. Should have 5 years warranty and 5 years CMC subsequently. Specification for Light Source Should be compatible with semi-automatic leakage tester Unit should be compact and light weight. Light source minimum 150 watt halogen/LED. Should have Inbuilt air pump. Essential Accessories Ten reusable cups with needle type biopsy forceps compatible with the scope channel diameter. Five reusable alligator type biopsy forceps compatible with the scope channel diameter. Five reusable alligator type biopsy forceps compatible with the scope channel diameter. 	01

6	Thin adult	1. Should have latest colour CCD chip technology built into the	01
	Videobronchosco	distal end	
	pe	2. Should enable the bronchoscopist to access peripheral areas	
		including upper lobes	
		3. Should be light weight and easy to use	
		4. Easy to pass endotherapy accessories including endoscopic	
		ultrasound and various electrosurgical accessories like electro	
		cautery, biopsy forceps, needle for aspiration, brush for cytology	
		5. Having good suction quality	
		6. Other characteristics	
		§ Field of view (degree): 100 degree or more	
		§ Direction of view : Forward	
		§ Depth of field (mm): 2-50mm or more	
		§ Distal end outer diameter (mm) : 4.2 mm	
		§ Insertion tube diameter (mm) : 4.1 mm or less	
		§ Insertion tube rotation : Yes	
		§ Working channel inner diameter (mm): 2.00 mm	
		§ Flexible tip angulation: At least 210 upwards and 130	
		downwards	
		§ Working length (mm): Minimum 600	
		o Can be fully immersed in disinfectant solution and water	
		o Should have facility to check leakage by automatic pressure	
		regulated leak tester (automatic leak tester)	
		o Good ergometric design with simple and easy to remember	
		configuration of switches in control section facilitating single	
		handed operation	
		o Minimum number of remote switches on control section (four)	
		o Joint less channel configuration	
		o Exterior design to minimise surface protrusions and interior	
		design should be simplified	
		o Compatible with all standard accessories like biopsy forceps,	
		brush, needle aspiration	
		o Should be fully insulated to minimize any potential risk during	
		electro cautery	
		o Should be supplied with all standard accessories including	
		different type biopsy forceps, cleaning brushes and storage boxes	
		Video imaging system	
		o Delivering high resolution images and true to life colour	
		reproduction for accuracy	
		o Should be a single chip camera technology	
		o Full height display mode filling the screen vertically to provide	
		much larger images without distortion	
		o Capable of storing at least 40 patient data	
		o Image storing and retrieving settings	
		o Should have image freezing capability on the scope	
		o Should have automatic white balance and memory function	
		o Should have fully automatic exposure control	
		o Should have horizontal resolution of more than 450 lines	
		o Should be supplied with flat TV monitor and a trolley	
		o Should work with AC on 200-240V and 50 Hz supply Light	
		source	
		o Should be a xenon /LED light source.	
		o Should have dual fan cooling system	
		z i	
		o Should work with AC on 200-240V and 50 Hz supply	
		o Should have minimum running cost	
		o Should be supplied with suitable trolley	
	İ	o Trolley should have at least 5 power socket to connect camera,	i .

Supply,	Installation and Con	nmissioning of Equipment required in Pulmonary Medicine Depar	rtment			
7	Medical grade	Cleaning tank should be made of stainless steel	01			
	ultrasonic cleaner	User-friendly and clear operating panel.				
	for biopsy forceps	• Easily mobile				
		• LED-Display showing set and remaining time of cleaning period,				
		pre-set and actual temperature				
		• Turning knob for setting continued and short-period operation				
		from 1 to 30 min				
		Temperature-controlled ultrasonic operation				
		• Ultrasonic frequency should be more than 35 kHz				
		• Should work on AC supply with 200-250 v and 50 amp				
		• The tank internal dimension should be at least 500/400/250 mm ³				
		• The basket internal dimension should be at least 450/250/190				
		mm3				
		• Size of processing chamber should be at least 7.5 litre				
		• Sweep function for an optimised sound field distribution in the				
		cleaning liquid by frequency modulation				
		• The drain valve and tank design should ensure efficient drainage.				
		Tray design should facilitate even distribution of ultrasonic				
		waves.				
		The whole system should be compliant with Indian electrical safety				
		standards				

TECHNICAL BID

37 10 10 11	
Name of Firm/Contractor/Supplier	
Complete Address &	
Telephone No.	
Name of Proprietor/Partner/Managing	
Director/Director.	
Phone No:-	
Mobile No:-	
Email Id:-	
Name and address of service centre nearby Jodhpur.	
Whether the firm is a registered firm Yes/No (attached copy of certificate).	
PAN No. (enclose the attested copy of PAN Card).	
Service Tax No. (enclose the attested copy of Service Tax Certificate).	
VAT No. (enclose the attested copy of VAT Certificate).	
Whether the firm has enclosed the Bank Draft/Pay Order/Banker's cheque of Earnest Money Deposit.	
Whether the Firm/Agency has signed each and every page of Tender/NIT.	
Please provide full list of consumables.	
Any other information, if necessary.	

Authorized signatory of the bidder with seal.

MANUFACTURER'S / PRINCIPAL'S AUTHORIZATION FORM

То	
The Administrative Officer, All India Institute of Medical Sciences, Jodhpur	
Sir,	
TENDER:	
we,	, who are
established and reputable manufacturers of	, having factories at
and	, hereby authorize Messrs.
(name and addre	ess of agents)
to bid, negotiate and conclude the contract with you against Tend	der
No for	the above goods manufactured by
us. No company or firm or individual other than Messrs	
bid, negotiate and conclude the contract in regard to this business	s against this specific tender.
We hereby extend our full guarantee and warranty as per the co	nditions of tender for the goods offered for
supply against this tender by the above firm.	
The authorization is valid up to	
	Yours faithfully,
	(Name)
For	and on behalf of Messrs
	(Name of manufacturers)/Principal.

Annexure-IV

Financial Bid

A) FINANCIAL BID FOR DOMESTIC GOODS OR GOODS OF FOREIGN ORIGIN LOCATED WITHIN INDIA OR GOODS TO BE IMPORTED AND SUPPLIED AGAINST PAYMENT IN INDIAN RUPEES

1	2	3	4		5					6	
Schedule	Brief Description of Goods	Country of Origin	Quantity (Nos.)	Ex - factory/ Ex- warehous e /Ex- showroo m /Off- the shelf (a)	Excise Duty(if any) [%age & & value] (b)	Sales Tax/ VAT(if any) [%age & value] (c)	Packing and Forwarding charges (d)	Inland Transportation, Insurance, loading/ unloading and Incidental costs at AIIMS- Jodhpur (e)	Incidental Services (including Installation & Commissioning, Supervision, Demonstration and Training) at AIIMS- Jodhpur (f)	Unit Price (at AIIMS- Jodhpur) basis (g)= (a+b+c+d+e+ f)	Total Price (at AIIMS- Jodhpur) basis (Rs.) = {4 x 5(g)}

Total Tender price in Rupees:
In words:

Note: -

- 1. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
- 2. The charges for Annual CMC after warranty shall be quoted separately.
- 3. The Bidder must quote price for "GOODS TO BE IMPORTED AND SUPPLIED AGAINST PAYMENT IN INDIAN RUPEES" after having taken in to account, the provision of Custom Duty Exemption Certificate (CDEC) by the Purchaser, as per Customs Tariff Act.

Place:	Name:
Date:	Business Address:
	Signature of Bidder:
	Seal of the Bidder

Financial Bid B) FINANCIAL BID FOR GOODS TO BE IMPORTED FROM ABROAD

1	2	3	4	5				
					Total price on			
Schedule	Brief Description of Goods	Country of Origin	Quantity (Nos.)	FOB price at port/ airport of Lading (a)	Carriage & Insurance (port of loading to port of entry) and other Incidental costs** (b)	Incidental Services (including Installation & Commissioning, Supervision, Demonstration and Training) at AIIMS-Jodhpur **	Unit Price on DDP AIIMS- Jodhpur + Extended Insurance (local transportation and storage) (d) = a+b+c	Destination + Insurance (local transportation and storage) $= \{4X \ 5 \ (d)\}$

** To be paid in Indian Currency (Rs.)	
Total Tender price in foreign currency:	
In words:	
Note: -	

- 1. If there is a discrepancy between the unit price and total price THE UNIT PRICE shall prevail.
- 2. The charges for Annual CMC after warranty shall be quoted.
- 3. The Bidder will be fully responsible for the safe arrival of the goods AIIMS-Jodhpur in good condition as per terms of DDP as per INCOTERMS, if applicable.

Indian Agent: Indian Agency Commission% of FOB	
mulan rightly commission =	
Place:	Name:
Date:	Business Address:
	Signature of Bidder
	Seal of the Bidder:

Financial Bid

C) FINANCIAL BID FOR ANNUAL COMPREHENSIVE MAINTENANCE CONTRACT AFTER WARRANTY PERIOD:

1. S.No.	2. DESCRIPTION OF GOODS	3. QUANTITY. (Nos.)	4. Annual Comprehensive Maintenance Contract Cost for Each Unit year wise*. 1 st 2 nd 3 rd 4 th 5 th a b c d e				5. Total Annual Comprehensive Maintenance Contract Cost for 5 Years [3 x (4a+4b+4c+4d+4e)]	

^{*} After completion of Warranty period.

Service Tax: Whether extra or inclusive, if extra, indicates the rate_____.

NOTE:-.

- 1. In case of discrepancy between unit price and total prices, THE UNIT PRICE shall prevail.
- 2. The cost of Comprehensive Maintenance Contract (CMC) which includes preventive maintenance including testing & calibration as per technical/ service /operational manual, labour and spares, after satisfactory completion of Warranty period may be quoted as per NIT conditions on yearly basis for complete equipment and Turnkey (if any).
- 3. The cost of CMC may be quoted along with taxes applicable. The taxes to be paid extra, to be specifically stated. In the absence of any such stipulation the price will be taken inclusive of such taxes and no claim for the same will be entertained later.
- 4. Cost of CMC will be added for Ranking/Evaluation purpose.
- 5. All software updates should be provided free of cost during CMC period.
- 6. The stipulations in Technical Specification will supersede above provisions
- 7. The supplier shall keep sufficient stock of spares required during Annual Comprehensive Maintenance Contract period. In case the spares are required to be imported, it would be the responsibility of the supplier to import and get them custom cleared and pay all necessary duties.

Date:	
Place:	Name:
	Business Address:
	Signature of Bidder:
	Seal of the Bidder: