

Tender
For
Dental Chair (Advanced & Basic)
At
All India Institute of Medical Sciences, Jodhpur

NIT No.	÷	Admin/General/42/2013-AIIMS.JDH.
NIT Issue Date	÷	September 02, 2013
Pre-Bid Meeting	÷	September 18, 2013 at 03:00 PM.
Last Date of Submission	÷	October 01, 2013 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur
Basni Phase - II, Jodhpur, Rajasthan—342005.
Telephone: 0291- 2740532, email: aoadmin@aiimsjodhpur.edu.in
www.aiimsjodhpur.edu.in

All India Institute of Medical Sciences (AIIMS), Jodhpur, Rajasthan, an apex healthcare institute being established by Parliament of India under aegis of Ministry of Health & Family Welfare, Government of India, invites sealed tenders for supply & installation of the following items at the institute. You are requested to quote your best offer along with the complete details of specifications, terms & conditions.

ANNEXURE 'A'

S.No	NIT No.	Qty	EMD (in Rs.)	Item Description
1.	Admin/General/42/2013-AIIMS.JDH	1. Nos	8,000	Dental Chair (Advanced)
2.	Admin/General/42-A/2013-AIIMS.JDH	1. Nos	4,000	Dental Chair (Basic)

(Refer Specifications Details as per Annexure-'B')

Quotation should be sealed and superscribed with tender number and address to:

“Administrative Officer
All India Institute of Medical Sciences, Jodhpur
Basni, Phase-II
Jodhpur-342005, Rajasthan”.

The sealed quotations should reach the Institute, latest by October 01, 2013 at 03:00 PM and it will be opened on same day at 04:00 PM in the Project Cell, Residential Complex, AIIMS, Jodhpur of the Institute in the presence of the bidder(s) or their authorized representative(s), who will present at the scheduled date and time.

Terms & Conditions:

- 1. Earnest Money Deposit:** The bidder shall be required to submit refundable amount as Earnest Money Deposit (EMD) and a non-refundable tender fee for an amount of **1,000/- (Rupees One Thousand only)** for each NIT by way of demand drafts only as mentioned in Annexure 'A'. The demand drafts shall be drawn in favour of **“All India Institute of Medical Sciences, Jodhpur”**. The demand drafts for earnest money deposit must be enclosed in the envelope containing the technical bid.

The EMD of the successful bidder shall be returned after the successful completion of contract / order and for unsuccessful bidder(s) it would be returned after award of the contract. Bid(s) received without demand drafts of EMD shall be liable for rejection.

The firms who are registered with National Small Industries Corporation (NSIC) / OR Small Scale Industrial (SSI) are exempted to submit the EMD (Copy of registration must be provided alongwith).

- 2. Rate:** Rates should be quoted in Indian Rupees (INR) on DOOR Delivery Basis at AIIMS, Jodhpur, Rajasthan, Inclusive of all the Charges, with break-ups as:
 - Basic Cost.
 - VAT /CST as applicable.
 - Total Cost (F.O.R at AIIMS Jodhpur).

- 3. Validity:** The quoted rates must be valid for a period for 120 days from the date of closing of the tender. The overall offer for the assignment and bidder(s) quoted price shall remain unchanged during the period of validity. If the bidder quoted the validity shorter than the required period, the same will be treated as unresponsive and it may be rejected.

In case the tenderer withdraws, modifies or change his offer during the validity period, bid is liable to be rejected and the earnest money deposit shall be forfeited without assigning any reason thereof. The tenderer should also be ready to extend the validity, if required, without changing any terms, conditions etc. of their original tender.

- 4. Delivery & Installation:** The items shall be delivered within 30 days of issue of supply order at AIIMS, Jodhpur. Satisfactory installation / commissioning and handover of the items will be completed within two weeks from the date of receipt of the goods at the AIIMS, Jodhpur premises. The successful tenderer will also provide required training for supplied items at AIIMS-Jodhpur.

The goods should be manufactured after adoption of latest technology.

- 5. Sample:** AIIMS Jodhpur reserves the right to ask the tenderers for submitting the sample of each item for which rates have been quoted, Technically Qualified Bidders may be asked to submit samples along with their quoted items nos. and their firm name without indicating any prices before opening of Financial Bid to AIIMS, Jodhpur for Inspection.

- 6. Guarantee / Warrantee Period: For the equipment value upto Rs. 5 Lakh:** The Tenderers must quote for 2 years comprehensive warranty (Including all Spares, Accessories and Labour) from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. Also the bidders are requested to submit their quote (Rates) for subsequent 3 years Comprehensive Maintenance Contract (CMC) (Including All Spares, Accessories and Labour). Failure to comply this condition will entail the rejection of the bids. The price comparison shall be taking into account on basic price and post warranty CMC.
CMC Charges will be payable separately at the time of start of CMC.

- 7. Guarantee / Warrantee Period: For the equipment value above Rs. 5 Lakh:** The Tenderers must quote for 5 years comprehensive warranty (Including all Spares, Accessories and Labour) from the date of completion of the satisfactory installation. The warranty charges shall not be quoted separately otherwise the offer shall be summarily rejected. Also the bidders are requested to submit their quote (Rates) for subsequent 5 years Comprehensive Maintenance Contract (CMC) (Including All Spares, Accessories and Labour). Failure to comply this condition will entail the rejection of the bids. The price comparison shall be taking into account on basic price and post warranty CMC.

- 8. Performance Security:** The supplier shall require to submit the performance security in the form of irrevocable Bank Guarantee (BG) / or Fixed Deposit Receipt (FDR) issued by any Nationalised Bank for an amount of which is equal to the 10% of the order value and should be kept valid for a period of 60 day beyond completion of all the contractual obligation including CMC period.

9. Payment Term:

- 90% payment of the total order value shall be released after the successful installation/ commissioning of the ordered goods against the submission of the test report.
- Balance 10% of the order value shall be released after the submission of the performance security.

10. Bidder shall submit a copy of the tender document and addenda thereto, if any, with each page of this document should be signed and stamped to confirm the acceptance of the entire terms & conditions as mentioned in the tender enquiry document.

11. Signed & stamped compliance sheet of the technical specification of the goods with technical printed literature must be enclosed with the bid.

12. After due evaluation of the bid(s) Institute will award the contract to the lowest evaluated responsive tenderer.

13. Equipment should be USFDA or European CE approved / certified.

14. Conditional bid will be treated as unresponsive and it may be rejected.

15. The Institute reserves the right to accept in part or in full or reject any or more quotation(s) without assigning any reasons or cancel the tendering process and reject all quotations at any time prior to award of contract, without incurring any liability, whatsoever to the affected bidder or bidder(s).

16. Quantity: The quantity of item given in the tender is tentative, which may be increased or decreased as per the institute's requirement.

17. Applicable Law:

- The contract shall be governed by the laws and procedures established by Govt. of India, within the framework of applicable legislation and enactment made from time to time concerning such Commercial dealings / processing.
- Any disputes are subject to exclusive jurisdiction of Competent Court and Forum in Jodhpur, Rajasthan, India only.
- The Arbitration shall be held in accordance with the provisions of the Arbitration and Conciliation Act, 1996 and the venue of arbitration shall be at Jodhpur. The decision of the Arbitrator shall be final and binding on both the parties.
- Force Majeure: Any delay due to Force Majeure will not be attributable to the supplier.

"PRE –BID Meeting" with the intending bidders shall be held on 18th Sep 2013 from 03:00 P.M. onwards at AIIMS, Jodhpur.

Specifications

Annexure B

S.No	Tender Name	Specification	QTY
1.	Dental Chair (Advanced)	<p>1 Description of Function</p> <p>1.1 Dental Chair Medium is the Dental Chair required for dental and surgical procedures.</p> <p>2 Operational Requirements</p> <p>2.1 Physiological Dental Chair operated by electricity</p> <p>3 Technical Specifications</p> <p>3.1.Dental unit should have latest overhead delivery system</p> <p>3.2.It should have two 3 way syringes (tip autoclavable, with 4 spare tips) one on unit side and other on the assistant side.</p> <p>3.3.It should have one high speed Air Rotor terminal with water control on coupling supplied with original CE 380000 rpm handpiece.</p> <p>3.4.It should have one high speed fiber-optic air-rotor terminal with handpiece</p> <p>3.5.One brushless micromotor (40000 rpm) terminal having straight and contra angle handpieces</p> <p>3.6.It should have LED light cure unit on unit sides (Min. Intensity 800 mW/cm² and wave length range - 370 - 500 nm output)</p> <p>3.7.It should have one in-built Piezon Ultrasonic Scaler (frequency 28-36 KHz) with 3scaler tips and one set of perio-curette tips</p> <p>3.8. All handpieces/terminals should be kept on Autoclavable pads. 6 spare autoclavable pads should be supplied</p> <p>3.9Removable auxillary tray (stainless steel)</p> <p>3.10.It should have latest LED/halogen Light with two intensity and sensor control(min 35,000 LUX)</p> <p>3.11.It should have high and low vacuum motorized with continuous nonstop function suction</p> <p>3.12.It should have two erasable programmable and return to Zero position</p> <p>3.13It should have LED based X-ray viewer</p> <p>3.14. It should be provided with rotatable right arm</p> <p>3.15. It should have multifunctional foot control base (fixed or mobile)</p> <p>3.16. It should be provided with one doctor's stool and one assistant's stool with adjustable backrest tilt including an adjustable ring for foot rest.</p> <p>3.17. Oil Free Air Compressor, 1.5 HP with silicon filter and dryer (Medical Grade)</p> <p>3.18 Should have attached monitor mounting Arm</p> <p>4 System Configuration Accessories, spares and consumables</p> <p>4.1 System as specified</p>	1. Nos

		<p>4.2 All consumables required for installation and standardization of system to be given free of cost.</p> <p>4.3 All fixtures and furniture for 10×12 feet surgeries. Fixtures should include modular work station with acid and fire resistant table top, high quality wash basin, modular drawer system of approximate size of 6 feet×2.5 feet×3 feet and one instrument trolley.</p> <p>4.4 All the outlet and inlet for the services to chair should be concealed in a box at foot area or within the unit for infection control purpose.</p> <p>5 Environmental factors</p> <p>5.1 The unit shall be capable of operating continuously in ambient temperature of 10-40deg</p> <p>5.2 Complete installation of the system including water input and drainage system has to be installed</p> <p>6 Power Supply</p> <p>6.1 Power input to be 220-240VAC, 50Hz</p> <p>6.2 Five KV Servo Voltage stabilizer of appropriate ratings meeting ISI Specifications.(Input 160-260 V and output 220-240 V and 50 Hz)</p> <p>7 Standards, Safety and Training</p> <p>7.1 Should be USFDA/ European CE approved product</p> <p>7.2 Manufacturer/Supplier should have ISO certification for quality standards.</p> <p>7.3 Electrical safety conforms to standards for electrical safety IEC-60601 / IS-13450.</p> <p>8 Documentation</p> <p>8.1 User/Technical/Maintenance manuals to be supplied in English.</p> <p>8.2 Certificate of calibration and inspection.</p> <p>8.3 List of important spare parts, handpieces, and accessories with their part number and costing</p> <p>8.4 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.</p>	
2.	Dental Chair (Basic)	<p>1 Description of Function</p> <p>1.1 Dental Chair is the Dental Chair required for dental examination and dental procedures.</p> <p>2 Operational Requirements</p> <p>2.1 Physiological Dental Chair (Basic) operated by electricity</p> <p>3 Technical Specifications</p> <p>3.1.Dental unit with modular delivery system</p> <p>3.2.It should have two 3 way syringes (tip autoclavable)</p> <p>3.3.It should high speed Air Rotor terminal supplied with original CE mark 380000 rpm handpiece.</p> <p>3.4.Should have one brushless micromotor(40000rpm) terminal</p>	1. Nos

		<p>with straight and contra angle handpieces</p> <p>3.5.It should have one in-built Piezon Ultrasonic Scaler (frequency 28-36 KHz) with 8 scaler tips.</p> <p>3.6. It should have LED/halogen Light with sensor (min 25,000 LUX)</p> <p>3.7. .It should High and low motorised suction with auto drain and flush system.</p> <p>3.10.Removable auxillary tray (stainless steel)</p> <p>3.11. Auto connection for tumbler and spittoon</p> <p>3.12 It should have LED based X-ray viewer</p> <p>3.13 It should be provided with one doctor’s stool with adjustable backrest tilt including an adjustable ring for foot rest.</p> <p>4 System Configuration Accessories, spares and consumables</p> <p>4.1 System as specified</p> <p>4.2 All consumables required for installation and standardization of system to be given free of cost.</p> <p>5 Environmental factors</p> <p>5.1 The unit shall be capable of operating continuously in ambient temperature of 10-40deg C and relative humidity of 15-90%</p> <p>5.2 Complete installation of the system including water input and drainage system has to be installed</p> <p>6 Power Supply</p> <p>6.1 Power input to be 220-240VAC, 50Hz</p> <p>6.2 Five KV Servo Voltage stabilizer of appropriate ratings meeting ISI Specifications.(Input 160-260 V and output 220-240 V and 50 Hz)</p> <p>7 Standards, Safety and Training</p> <p>7.1 Should be USFDA/ European CE approved product</p> <p>7.2 Manufacturer/Supplier should have ISO certification for quality standards.</p> <p>7.3 Electrical safety conforms to standards for electrical safety IEC-60601 / IS-13450.</p> <p>8 Documentation</p> <p>8.1 User/Technical/Maintenance manuals to be supplied in English.</p> <p>8.2 Certificate of calibration and inspection.</p> <p>8.3 List of important spare parts, handpieces, and accessories with their part number and costing</p> <p>8.4 Log book with instructions for daily, weekly, monthly and quarterly maintenance checklist. The job description of the hospital technician and company service engineer should be clearly spelt out.</p>	
--	--	--	--

Annexure-C

Inviting of sealed quotations for supply and installation of Dental Chair
(Advance & Basic) for Department of Dentistry at AIIMS, Jodhpur

TECHNICAL BID

Name of Firm/Contractor/Supplier	
Complete Address & Telephone No.	
Name of Proprietor/Partner/Managing Director/Director.	
Phone & Mobile No.	
Name and address of service centre near by Jodhpur.	
Whether the firm is a registered firm Yes/No (attached copy of certificate)	
PAN No. (enclose the attested copy of PAN Card)	
Service Tax No. (enclose the attested copy of Service Tax Certificate)	
VAT No. (enclose the attested copy of VAT Certificate)	
Whether the firm has enclosed the Tender Fees as per Annexure 'A'	
Whether the firm has enclosed the Bank Draft/Pay Order/Banker's cheque as Earnest Money Deposit as per Annexure 'A'	
Whether the Firm/Agency has signed each and every page of Tender/NIT	
Please provide full list of consumables with rates	
Any other information, if necessary	

Authorized signatory of the bidder with seal.

Annexure-D

Financial Bid

(To be submitted on the letterhead of the company / firm separately for each item)

S.No.	Item Description	QTY	Rate	Vat/ Tax	Amount

CMC Charges as applicable (excluding Service Tax)	
1 st Year	
2 st Year	
3 st Year	
4 st Year	
5 st Year	

1. I/We have gone through the terms & conditions as stipulated in the tender enquiry document and confirm to accept and abide the same.
2. No other charges would be payable by the Institute.

(Authorized signatory of the bidder with seal)