



**All India Institute of Medical Sciences  
Jodhpur**

Date: - 25/09/2014

**Corrigendum for Computerised Pulmonary function test Machine with  
Body Plethysmography.**

NIT Issue Date : September 02<sup>nd</sup>, 2014.  
Pre Bid Meeting held on : September 15<sup>th</sup>, 2014 at 03:00 PM  
Last Date of Submission : October 01<sup>st</sup>, 2014 at 03:00 PM  
**Revised Last date of Submission** : October 09<sup>th</sup>, 2014 at 03:00 PM

**1. The following revised and additional terms & conditions will be added:-**

**1. Page No. 6, Point No. 22, Payment Terms:**

**For**

**Payment for goods supplied from India:**

a. 100% payment released after the submission of the performance security.

**Read**

Payment for goods supplied from India:

a. 100% payment of the total order value shall be released after the successful installation/ commissioning of the ordered goods against the submission of the inspection report.

**2. The following revised and additional specification will be added:-**

**1. Page No. 10, Annexure-I, Specifications, Point No. f:**

**For**

Intra Breath Diffusion capacity.

**Read**

Intra Breath Diffusion capacity or equivalent technique of customising breath hold time.

**2. Page No. 10, Annexure-I, Specification, Box No.09:**

**For**

Should have a multiuse heated screen/Lilly/Fleisch Pneumotachograph with range, linearity, resistance and accuracy meeting or exceeding American Thoracic Society – European.

**Read**

Should have a multiuse heated screen/Lilly/Fleisch Pneumotachograph/ Differential Pressure Pneumotach with range, linearity, resistance and accuracy meeting or exceeding American Thoracic Society -European Respiratory Society standards

**Administrative Officer  
AIIMS Jodhpur**