



**All India Institute of Medical Sciences  
Jodhpur**

Date: - 07/11/2014

**Corrigendum**

**for**

**Deep Freezers- 20° C & Deep Freezers-80° C.**

NIT Issue Date	:	September 14 <sup>th</sup> , 2014.
Pre Bid Meeting held on	:	October 21 <sup>st</sup> , 2014 at 04:00 PM
Nit No.	:	Admn/Tender/14,15/2014-AIIMS.JDH
Extended Last Date of Submission	:	November 10 <sup>th</sup> , 2014 at 03:00 PM
Revise Last Date of Submission	:	November 20 <sup>th</sup> , 2014 at 03:00 PM

**1. The following revised and additional terms & conditions will be added:-**

**1. Page No. 03, Point No. 09, Para 01:**

**For**

Delivery and Installation: The items shall be delivered within **30 days** of issue of supply order at AIIMS, Jodhpur. Satisfactory installation / commissioning and handover of the items will be completed within two weeks from the date of receipt of the goods at the AIIMS, Jodhpur premises. The successful tenderer will also provide required training for supplied items at AIIMS-Jodhpur.

**Read**

Delivery and Installation: The items shall be delivered within **60 days** of issue of supply order at AIIMS, Jodhpur. Satisfactory installation / commissioning and handover of the items will be completed within two weeks from the date of receipt of the goods at the AIIMS, Jodhpur premises. The successful tenderer will also provide required training for supplied items at AIIMS-Jodhpur.

**2. The following revised and additional specification will be added:-**

**1. Page No. 10, Annexure-I, Technical Specifications, Item 01:**

**1. Point No. 1:**

**For**

Upright, 9.53 to 13 cu ft.

**Read**

Upright, with capacity of 270 liters or more.

**2. Point No. 2:**

**For**

Capacity to hold atleast 20,000 to 25,000 vials.

**Read**

Deleted and merged with point no. 1.

**3. Point No. 3:**

**For**

1 to 2 horsepower compressors.

**Read**

½ to ¾ horsepower compressors.

**4. Point No. 4:**

**For**

Freezer must have one lockable insulated solid door, Durable Steel interior with Inner doors.

**Read**

Freezer must have one lockable insulated solid door.

**5. Point No. 5:**

**For**

Freezer must come standard with four baskets and two pull cover drawers.

**Read**

Freezer must come standard with four drawer/baskets and two pull cover drawers.

**6. Point No. 7:**

**For**

Freezer must have interior lighting with external on/off switch. Freezer must have hasp standard for attachment of secondary lock.

**Read**

Freezer must have interior lighting with external on/off switch. Freezer with hasp standard for attachment of secondary lock preferred.

**7. Point No. 8:**

**For**

Triple -sealing silicone door gasket.

**Read**

Deleted

**8. Point No. 18:**

**For**

Front to back airflow with Removable, cleanable air filter.

**Read**

Deleted.

**9. Point No. 19:**

**For**

Hinged grill swings out for easy access to filter and battery.

**Read**

Deleted.

**10. Point No. 21.2:**

**For**

Freezer must have compressor on/off indicator on control panel.

**Read**

Deleted.

**11. Point No. 26:**

**For**

Accessories - cryogloves, replaceable air filters, 3 hours battery backup for cooling.

**Read**

Accessories- cryogloves.

Optional accessories- 3 hours battery backup for cooling.

**2. Page No. 11, Annexure-I, Technical Specifications, Item 02:**

**1. Point No. 01:**

**For**

Upright, 10 to 13 cu ft.

**Read**

Upright, 17 cu ft or more.

**2. Point No. 02:**

**For**

Capacity to hold atleast 20,000 to 25,000 vials.

**Read**

30,000 vials or more.

**3. Point No. 24:**

**For**

Accessories - cryogloves, replaceable air filters, 3 hours battery backup for cooling.

**Read**

Accessories- cryogloves, replaceable air filters.

Optional accessories- Online UPS backup for 3 hours battery backup for cooling.

**Administrative Officer  
AIIMS Jodhpur**