Invitation of quotation

for

Supply of Transport Pulse Oximeter

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.:	:	Admin/Gen/09-04/2022-AIIMS.JDH
Inquiry Issue Date	:	06 th May, 2022
Last Date of Submission	:	12 th May, 2022 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan Telefax: 0291- 2740741, email: **procurement@aiimsjodhpur.edu.in** www.aiimsjodhpur.edu.in

Invitation of quotation for Supply of Transport Pulse Oximeter at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Supply of Transport Pulse Oximeter for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 12.05.2022 03:00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

<u>"QUOTATION FOR SUPPLY OF TRANSPORT PULSE OXIMETER</u> <u>AGAINST INQUIRY NO. ADMN/GEN/09-04/2022-AIIMS.JDH" DUE ON</u> <u>12.05.2022 03:00 PM"</u>

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible.
 The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in "Quotation Box" located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
 - Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST No.
 - The firm should not be black listed by any Govt. Agency/Dept.
- J) Quotations qualified by such vague and indefinite expressions such as "subject to prior

confirmation", "subject to immediate acceptance" etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

- K) Delivery Period within 30 days from Purchase order.
- L) **Liquidated Damage:** If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of "Terms & Conditions" of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.
- 2. Special Terms & Conditions:
- A) Bidder must quote the product as per specification provided in Annexure 1.
- **B**) Catalog must be attached with quotation for technical evaluation.
- C) The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

Deputy Director (Admin)

Encl.: Annexure 1 (Specification) Annexure 2 (Format of price bid)

Annexure 1

S. No.	Particular	Qty.	
1.	Transport Pulse Oximeter	01 Nos.	

Specifications:-

- 1. Should have heart rate, saturation (SpO2), ECG, RR, temperature, NIBP and IBP facility
- 2. Monitor should measure neonatal and pediatric patient's parameters.
- 3. Display should be maximum 6-6.3" TFT color antimicrobial, corning Gorilla Glass screen with auto rotation feature (It can change from portrait to landscape) with respect to gravity. Should have transport mountings so that it can be mounted on a transport trolley and strong handle making it comfortable to carry.
- 4. Can withstand bumps & bounces for transporting babies in a transport trolley. Can survive repeated fall from height of one meter i.e. it can absorb the shock due to a fall.
- 5. Can withstand any cleaning agents or aggressive disinfectants to make the display clean and sterile without damaging the monitor.
- 6. Weight should not be more than 3 kgs with battery.
- 7. Monitor should be charged when connected to Main monitor and can be used as independent transport Monitor when detached from Main monitor. Should be Upgradable to work on both AC mains and battery (Li-Ion battery).
- 8. It should be a high-resolution touch display with configurable screen layouts.
- 9. Rechargeable Li-Ion battery should be available with back up of minimum 4 hrs.
- 10. Monitor should have minimum 24 hrs. of trend memory
- 11. Audio and visual alarms should be available
- 12. Monitor should have capability to be used as module/ server for high-end ICU monitors /modular monitors with facility to transfer patient data seamlessly to the main ICU monitor when patient is back on bedside after in-hospital/ out of hospital transportation.
- 13. Monitor should be able to display 12 leads of ECG simultaneously on screen through 5 lead able. It should have minimum 5 waveforms channel display capability simultaneously.
- 14. ECG a) 12 lead ECG through 5/6 Leads ECG cable. b) Heart rate range 30-300 bpm, this should be the minimum range a range if wider than this is also acceptable. c) Accuracy should be min. (+/-) 1 % and resolutions should be min. 1 bpm
- 15. NIBP
 - a) Oscillometric method, should be operate on manual, automatic and STAT modes
 - b) Can be used for neonatal and pediatric patients
 - c) Range should be 10-200 mmHg
- 16. Respirations
 - a) Range 0-100 rpm
 - b) Resolution should be minimum (+/-) 1 rpm
- 17. Temperatures
 - a) C & F selectable
 - b) Temperature range 25-45 degree
- 18. SPO2
 - a) Masimo Signal Extraction Technology (SET) 0-100%
- 19. Monitor should have the facility for invasive BP monitoring
- 20. System should be European CE or US FDA approved
- 21. System should follow standard electrical norms and should be supplied with three pin plug (not two pin) with 240 VAC and 10-15 Amps
- 22. Should have local service facility and should have the necessary equipments to carry out preventive maintenance test
- 23. Onsite physical demonstration and training of the equipment to all the end users with all the requested facilities will be mandatory

- 24. List of essential spares, accessories, expendables and consumables expected to be used in should be quoted separately. Prices so quoted to be frozen for entire duration of warranty and CMC
- 25. Warranty for at least 2 years and CAMC for 5 years after warranty is over
- 26. Availability of spares for at least 10 years after date of installation
- 27. Monitor should able to withstand the temperature range from 0-50 degree C
- 28. Accessories should be supplied along with monitor a ECG 3 lead Trunk cable 01 no. along with pre-attached lead wire set (set of 3) 20 nos. b SPO2 Neonatal wrap around sensor- 01 no. c NIBP disposable cuff along with size 1, 2 & 3 -- 10 nos. of each d Temperature Esophaegeal probe for infants/ neonates 01 no. e NIBP Interconnect cable for disposable cuff 01 no.

Note: -

- **♦** Catalog/ Brochure must be attached for Technical Evaluation.
- The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

[On the letterhead of firm] ANNEXURE "2" PRICE BIDFORM

To,

Deputy Director (Admin), AIIMS, Jodhpur.

Dear Sir,

- 2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
- 3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty.	Moko	Exclusive of	Other	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1.	Transport Pulse Oximeter <u>Specification:-</u> As per annexure – 1	01 Pcs.						

Note:-

- 1. The Bidder must quote only single Make & Model.
- 2. The bidder must quoted their quotation only in above said format on the letter of firm otherwise quotation will be <u>REJECTED</u>.
- 3. Catalog must be attached with quotation for technical evaluation.
- 4. The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

Date	(Name)
Place	Name of Firm/Company/Agency
	GSTIN No.:
	Bank Name:
	Bank Account No.:
	IFSC Code:-
	Branch Name:
	Phone No
	Email:
	(Signature of Authorized Person)
	Seal: