Invitation of quotation

for

Supply of Under Water Seal Bottle with Valve (Double Chamber)

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/28-01(ii)/2018-AIIMS.JDH

Inquiry Issue Date : 17th May, 2018

Last Date of Submission : 23rd May, 2018 at 05:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan Telefax: 0291- 2012978, email: <u>procurement@aiimsjodhpur.edu.in</u> www.aiimsjodhpur.edu.in

<u>Invitation of quotation for Supply of Under Water Seal Bottle</u> <u>with Valve (Double Chamber) at AIIMS Jodhpur</u>

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for Supply of Under Water Seal Bottle with Valve (Double Chamber) for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 23.05.2018 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

"QUOTATION FOR SUPPLY OF UNDER WATER SEAL BOTTLE WITH VALVE (DOUBLE CHAMBER) AGAINST INQUIRY NO. ADMN/GEN/28-01(ii)/2018-AIIMS.JDH" DUE ON 23.05.2018 05.00 PM"

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in "Quotation Box" located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
 - Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST/Other taxes and IT PAN.
 - The firm should not be black listed by any Govt. Agency/Dept.

INQUIRY NO. Admin/Gen/28-01(ii)/2018-AIIMS.JDH

- J) Quotations qualified by such vague and indefinite expressions such as "subject to prior confirmation", "subject to immediate acceptance" etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** within 30 days from Purchase order.
- L) **Liquidated Damage:** If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of "Terms & Conditions" of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.
- 2. **Special Terms & Conditions:**
- A) Bidder must quote the product as per specification provided in Annexure 1.
- B) Catalog must be attached with quotation for technical evaluation.
- C) The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AHMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.
- D) Inspection committee will check the products thoroughly, if somehow inspection committee found any discrepancy and is not satisfied with the final product (final product should be same as reviewed ones at the institution) then AIIMS, Jodhpur has the right to reject the supply and cancel the order, and no claim for payment in this regard will be entertained

Administrative Officer

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

INQUIRY NO. Admin/Gen/28-01(ii)/2018-AIIMS.JDH

Annexure 1

S. No	Particular	Size				
1.	Under Water Seal Bottle with Valve (Double Chamber)	 Design patented, excess negative pressure control mechanism and positive pressure control valve, under water seal drainage system. Safeguards against damage to patient's vital organs. Equipped with a "pressure control valve" to ensure controlled, safe drainage. The valve limits the maximum negative pressure to -50cm H2O and releases the excess pressure if applied erroneously. The mechanism also controls the positive pressure if generated into the system anytime. Easy to read graduations on the bottle help note the drainage volume. Clearly marked initial level ensures effective under water seal. Write on facility on the side helps to monitor the drainage volume precisely. Suction port is provided for connection with suction unit. Metallic hanger for easy attachment to patient's bedside. Sterile, individually packed. Maxve pressure:50 cm H2O Capacity: - 2000 ml Allows optimal drainage With excess negative/positive pressure control valve. 	50 Nos.			

Note: The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

INQUIRY NO. Admin/Gen/28-01(ii)/2018-AIIMS.JDH

[On the letterhead of firm] ANNEXURE "2" PRICE BIDFORM

			<u>DIDIDI</u>	<u> </u>			
To,							
	istrative Officer, , Jodhpur.						
Dear Si	r,						
quotation WITH Admn/	I/We	TATION CHAMBE S.JDH" D	FOR SUER) AT	PPLY OF UN AIIMS AGA 3.05.2018 05.0	DER W AINST	ATER SEAL THE INQU	IRY N
docume	I/We thoroughly examined ent, failing which my quotat	tion will b	e rejected	out rightly.	& conditi	ons given in t	he enqui
S. No	Particular	Qty	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/ Other Taxes	Price/ Unit Inclusive of GST (INR)	MRP
1.	Under Water Seal Bottle with Valve (Double Chamber)	50 Nos.		0.0 1 (0.1.03)			
 The other Cat The have den 	e Bidder must quote only see bidder must quoted the erwise quotation will be Realog must be attached with esupplier may be asked to been quoted, to the nonstrating the items will	eir quota EJECTE h quotatio to arrang AIIMS	ition only <u>D</u> . on for tecl ging demo Jodhpur,	in above said nnical evaluation onstration of the	on. heir equ	ipment for w	hich rat
_ mcc			(Signatu	re of Authorize	ed Perso	n)	
			(Name)_				
			Name of	Firm/Compan	y/Agenc	ey	
			Phone N	0			

Email:_____