

Invitation of quotation
for
CO2 Cable Connection between Central Pendant
to CO2 Insufflator

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/32-16/2020-AIIMS.JDH

Inquiry Issue Date : 15th March, 2021

Last Date of Submission : 18th March, 2021 at 03:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan

Telefax: 0291- 2740741, email: procurement@aiimsjodhpur.edu.in

www.aiimsjodhpur.edu.in

**Invitation of quotation for CO2 Cable Connection between
Central Pendent to CO2 Insufflator at AIIMS Jodhpur**

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for CO2 Cable Connection between Central Pendent to CO2 Insufflator for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 18.03.2021 03.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

**“QUOTATION FOR CO2 CABLE CONNECTION BETWEEN
CENTRAL PENDENT TO CO2 INSUFFLATOR AGAINST INQUIRY
NO. ADMN/GEN/32-16/2020-AIIMS.JDH” DUE ON 18.03.2021 03.00 PM”**

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. **The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.**
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in “**Quotation Box**” located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
- Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST/Other taxes and IT PAN.
 - **The firm should not be black listed by any Govt. Agency/Dept.**
- J) Quotations qualified by such vague and indefinite expressions such as “subject to prior

INQUIRY NO. Admin/Gen/32-16/2020-AIIMS.JDH

confirmation”, “subject to immediate acceptance” etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.

- K) **Delivery Period** – within 30 days from Purchase order.
- L) **Liquidated Damage:** - If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of “Terms & Conditions” of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. Special Terms & Conditions:

- A) **Bidder must quote the product as per specification provided in Annexure 1.**
- B) **Catalog must be attached with quotation for technical evaluation.**
- C) **The Bidder must submit the sample of quoted make for technical evaluation at Central Store, IPD Ground Floor, AIIMS Jodhpur at the time of submission of their quotation. Failure to submit the sample at the time of submission of their quotation/offer will be summarily rejected.**

Administrative Officer

Encl.: Annexure 1 (Specification)
Annexure 2 (Format of price bid)

Annexure 1

Sr No.	Item Name	Qty
1.	CO2 Cable Connection between Central Pendant to CO2 Insufflator (Karl Storz)	6
2.	CO2 Cable Connection between Central Pendant to CO2 Insufflator (Richard Wolf)	5
3.	CO2 Cable Connection between Central Pendant to CO2 Insufflator (AirSeal-Robotic)	3

Specifications:

- CO2 HOSE PIPE should be 1.5 meter SS 316 steel braided hose with endurance of 75 bar pressure.
- One end of Hose should have adapter with O-ring 2mm thickness diameter and of thread M22x1.5 pitch.
- Adapter O-ring should be made by NBR/BUNA having hardness 75 +/-5 on measurement with SHORE A Meter.
- Another end adapter compatible with CO2 insufflator of thread 14bsp.
- CO2 Hose Pipe with O-ring Adapter should be ISO and CE Certified.
- Warranty : 6 Months

Note:- The Bidder must submit the sample of quoted make for technical evaluation at **Central Store, IPD Ground Floor, AIIMS Jodhpur** at the time of submission of their quotation. Failure to submit the sample at the time of submission of their quotation/offer will be summarily rejected.

[On the letterhead of firm]

ANNEXURE "2"
PRICE BIDFORM

To,
Administrative Officer,
AIIMS, Jodhpur.

Dear Sir,

1. I/We Submitted the quotation for Enquiry No. **"QUOTATION FOR CO2 CABLE CONNECTION BETWEEN CENTRAL PENDENT TO CO2 INSUFFLATOR AT AIIMS AGAINST THE INQUIRY NO. Admn/Gen/32-16/2020-AIIMS.JDH" DUE ON 18.03.2021 03.00 PM** for CO2 Cable Connection between Central Pendent to CO2 Insufflator at AIIMS Jodhpur".
2. I/We thoroughly examined, understood and accepted terms & conditions given in the enquiry document, failing which my quotation will be rejected out rightly.
3. I/We hereby offer to supply at the following rates.

S. No	Particular	Qty.	Quoted Make	Price/Unit Exclusive of GST (INR)	GST/ Other Taxes	Price/ Unit Inclusive of GST (INR)	Total Cost Inclusive of GST (INR)	MRP
1.	CO2 Cable Connection between Central Pendent to CO2 Insufflator (Karl Storz)	6						
2	CO2 Cable Connection between Central Pendent to CO2 Insufflator (Richard Wolf)	5						
3	CO2 Cable Connection between Central Pendent to CO2 Insufflator (AirSeal-Robotic)	3						

INQUIRY NO. Admin/Gen/32-16/2020-AIIMS.JDH

Note:-

- 1. The Bidder must quote single Make.**
2. The bidder must quoted their quotation only in above said format on the letter of firm otherwise quotation will be REJECTED.
3. Catalog must be attached with quotation for technical evaluation.
4. The supplier may be asked to arranging demonstration of their equipment for which rates have been quoted, to the AIIMS Jodhpur, if required. The expenditure incurred for demonstrating the items will be borne by the supplier.

Date _____

Place _____

(Name) _____

Name of Firm/Company/Agency _____

GSTIN No.: _____

Bank Name:- _____

Bank Account No.: _____

IFSC Code:- _____

Branch Name: _____

Phone No. _____

Email: _____

(Signature of Authorized Person) _____

Seal: _____