



**ALL INDIA INSTITUTE OF MEDICAL SCIENCES,
JODHPUR**

Dated: 03th December, 2018

CORRIGENDUM

Invitation of Quotation

For

Biological Indicator for Plasma & Incubator for

Plasma

At

All India Institute of Medical Sciences, Jodhpur

Issue Date : 27th November, 2018
Inquiry No. : Admin/Gen/34-10/2018-AIIMS.JDH
Last Date of Submission : 04th December, 2018 at 05:00 PM.



ALL INDIA INSTITUTE OF MEDICAL SCIENCES, JODHPUR

The following revisions / modifications will be made:

1. Annexure 1 > Specification:

S.No.	Items Name	Specifications	Quantity
1	Incubator for Plasma	<p>1. Plasma Incubator which should provide result within 30 Minutes.</p> <p>2. Positive and Negative results should be within 5 to 7 Minutes indicate on the Auto-reader LCD display. (To be read as deleted)</p> <p>3. The ANSI/AAMI/ISO 11138 series of standards.</p> <p>4. Should have USFDA Approved.</p> <p>5. The Connectivity of the Incubator should be done Online by the Internet so that the online result on the system are possible.</p>	01 Nos