



All India Institute of Medical Sciences,
Jodhpur

Dated: 31st October, 2022

CORRIGENDUM

Invitation of quotation

for

Incise Drape

At

All India Institute of Medical Sciences, Jodhpur

Inquiry No.: : Admin/Gen/46-10(ii)/2022-AIIMS.JDH

Inquiry Issue Date : 22nd October, 2022

Last Date of Submission : 31st October, 2022 at 03:00 PM.

Subject- Extension of date and Revision in Annexure-1 in the original quotation dated: 22/10/2022.

Extension of date:-

The last date for submission of the below mentioned quotation is extended till 01.11.2022 at 03.00 PM.

S. No.	Inquiry No.	Item Description	Previous last date of Submission	New last date of Submission
1	Admin/Gen/46-10(ii)/2022-AIIMS.JDH	Incise Drape	31 st October, 2022	01 st November, 2022

Revision in Annexure-1 in the quotation dated: 11/10/2022.

For: -

Sr. No	Item Name	Specification	Qty.	Make
1.	Incise Drape	Incise Drape which is Antimicrobial, Strong, Comfortable and breathable Iodophor impregnated Polyester film coated with hypoallergenic adhesive for superior adhesion. Compatible with Povidone Iodine & CHG Preps. Iodine from the drape should be Present in the deepest Skin Layer (1000 micron). Should be Latex Free. Study/Evidence to Show reduction is SSI 34 cm * 35 cm (Adhesive Area). It Should be gamma Sterilize with USFDA Approved.	150 Nos.	3M/ Lifelong/ SRS Meditech
2.		Incise Drape which is Antimicrobial, Strong, Comfortable and breathable Iodophor impregnated Polyester film coated with hypoallergenic adhesive for superior adhesion. Compatible with Povidone Iodine & CHG Preps. Iodine from the drape should be Present in the deepest Skin Layer (1000 micron). Should be Latex Free. Study/Evidence to Show reduction is SSI 56 cm * 45 cm (Adhesive Area). It Should be gamma Sterilize with USFDA Approved.	350 Nos.	

Read: -

Sr. No	Item Name	Specification	Qty.	Make
1.	Incise Drape	Incise Drape which is Antimicrobial, Strong, Comfortable and breathable Iodophor impregnated Polyester film coated with hypoallergenic adhesive for superior adhesion. Compatible with Povidone Iodine & CHG Preps. Iodine from the drape should be Present in the deepest Skin Layer (1000 micron). Should be Latex Free. Study/Evidence to Show reduction is SSI 34 cm * 35 cm (Adhesive Area). It Should be gamma Sterilize with USFDA Approved.	150 Nos.	All can apply
2.		Incise Drape which is Antimicrobial, Strong, Comfortable and breathable Iodophor impregnated Polyester film coated with hypoallergenic adhesive for superior adhesion. Compatible with Povidone Iodine & CHG Preps. Iodine from the drape should be Present in the deepest Skin Layer (1000 micron). Should be Latex Free. Study/Evidence to Show reduction is SSI 56 cm * 45 cm (Adhesive Area). It Should be gamma Sterilize with USFDA Approved.	350 Nos.	

➤ **Rest of the contents of the quotation remains unchanged.**