Invitation of quotation

for

Tyvek Packing Material for Plasma Sterilization At

All India Institute of Medical Sciences, Jodhpur

Inquiry No. : Admin/Gen/48-05/2018-AIIMS.JDH

Inquiry Issue Date : 30th May, 2018

Last Date of Submission : 04th June, 2018 at 05:00 PM.



All India Institute of Medical Sciences, Jodhpur

Basni Phase - II, Jodhpur – 342005, Rajasthan Telefax: 0291- 2012978, email: <u>procurement@aiimsjodhpur.edu.in</u> www.aiimsjodhpur.edu.in

Invitation of quotation for Tyvek Packing Material for Plasma Sterilization at AIIMS Jodhpur

Sealed Quotations are hereby invited by the undersigned on behalf of the Director, AIIMS Jodhpur for supply of Tyvek Packing Material for Plasma Sterilization for the Institute as per terms & conditions mentioned below. The filled quotations along with all the required document must reach in the office of the undersigned on or before 04.06.2018 05.00 PM. The Envelope containing the quotation would please be sealed and super scribed as under:-

"QUOTATION FOR SUPPLY OF TYVEK PACKING MATERIAL FOR PLASMA STERILIZATION AGAINST INQUIRY NO. ADMN/GEN/48-05/2018-AIIMS.JDH" DUE ON 04.06.2018 05.00 PM"

1. Terms & Conditions:

- A) The quotations received after this deadline & unsealed shall not be entertained under any circumstances whatsoever. In case of postal delay this Institute will not be responsible. The offer Submitted Fax/Email shall not be considered and no correspondence will be entertained in this matter.
- B) Quotations must be in the enclosed prescribed Performa on the letter head of the firm duly signed by the Proprietor/ Partner/ Director or their authorized representative, In case of signing of quotation by the authorized representative letter of authorization must be attached with the quotation. Quotation must be dropped in "Quotation Box" located in Administration Block of AIIMS, Jodhpur.
- C) Rates must be quoted in **Indian rupees** and as per the format specified taxes extra if any must be written separately.
- D) Rates must be quoted FOR basis (including Freight charges, Insurance, installation etc.)
- E) No overwriting or cutting is permitted in the rate. If found, the quotation shall be summarily rejected.
- F) The rates quoted must be valid for 60 days minimum from the date of opening of the quotation and silence of any tendered on this issue shall be treated as agreed with this condition.
- G) Becoming L1 will not be the criteria for awarding of purchase order unless the rates are reasonable & justified.
- H) RTGS/NEFT details need to be furnished by the supplier with the quotation on the letter head of supplier/firm/agency.
- I) The firm/agency may satisfy the following conditions and attach self-attested copy of the same with the quotation:
 - Firm shall be registered with the Government of Rajasthan / Central Government.
 - The firm shall have valid GST No. and IT PAN.
 - The firm should not be black listed by any Govt. Agency/Dept.

INQUIRY NO. Admin/Gen/48-05/2018-AIIMS.JDH

- J) Quotations qualified by such vague and indefinite expressions such as "subject to prior confirmation", "subject to immediate acceptance" etc. will be treated as vague offers and rejected accordingly. Any conditional quotation shall be rejected summarily.
- K) **Delivery Period** within 30 days from Purchase order.
- L) **Liquidated Damage:** If the supplier fails to deliver the material on or before the stipulated date, then a penalty at the rate of 0.5 % per week of the total order value shall be levied subject to maximum of 10% of the total order value.
- M) **Payment Terms:** Payment will be only after satisfactorily delivery / commissioning of material and after inspection by the AIIMS Jodhpur.
- N) **Disputes:** -In the event of any dispute or disagreement arising between the contractors and any other department of AIIMS Jodhpur with regards to the interpretation of "Terms & Conditions" of this inquiry, the same shall be referred to the Director, AIIMS Jodhpur whose decision will be final and binding upon the contractor.
- O) AIIMS, Jodhpur reserves the right to increase or decrease quantity and / or amount of work. Decision of Quantity of material in the AIIMS, Jodhpur will be final in this regard.
- P) AIIMS, Jodhpur reserves the right to reject any quotation or part or the whole of inviting quotation process without assigning any reason. Decision of the AIIMS, Jodhpur will be final in this regard.

2. **Special Terms & Conditions:**

- A) Bidder must quote the product as per specification provided in Annexure 1.
- B) Catalog must be attached with quotation for technical evaluation.

Administrative Officer

Encl.: Annexure 1 (Specification)

Annexure 2 (Format of price bid)

INQUIRY NO. Admin/Gen/48-05/2018-AIIMS.JDH

Annexure 1

Specification

| Sr. No | Description Of Item | Code | Size | Qty |
|--------|--|-------|--------|--------|
| 1 | Tyvek Roll with Sterrad Chemical Indicator, 6*228 | 12415 | Small | 2 Case |
| 2 | Tyvek Roll with Sterrad Chemical Indicator, 10*228 | 12425 | Medium | 2 Case |
| 3 | Tyvek Roll with Sterrad Chemical Indicator, 20*228 | 12450 | Large | 2 Case |
| 4 | Sterrad Chemical Indicator Strip | 14100 | - | 2 Case |

INQUIRY NO. Admin/Gen/48-05/2018-AIIMS.JDH

[On the letterhead of firm] ANNEXURE "2" PRICE BIDFORM

| T | o, | | | | | | | | | | |
|-------------------|----------------------------------|---|--|--|-----------------------------------|--|-----------------|---|--|--|--|
| | | istrative Officer, , Jodhpur. | | | | | | | | | |
| D | ear S | ir, | | | | | | | | | |
| fo P D 1 | or Er LAS OUE (ocum | nquiry No. "QUOTATION MA STERILIZATION AGA ON 04.06.2018 05.00 PM at A I/We thoroughly examined, a ent, failing which my quotatio | FOR INST T IIMS Jounderston underston will b | SUPPLY THE IN(odhpur". ood and be rejecte | Y OF TOURY In accepted dout right | YVEK PANO.Admn/ol terms & chily. | ACKIN Gen/48 | NG MATE 8-05/2018-A ons given in | RIAL FOR MIMS.JDH" | | |
| | S. No. | Particulars | Pack Size | Quoted Make | Qty | Price/Unit Exclusive of GST (INR) | GST % | Price/Unit inclusive of GST (INR) | Total Amount Inclusive of GST (INR) | | |
| | 1. | Tyvek Roll with Sterrad Chemical Indicator, 6'*228' | | | 2 Case | | | | | | |
| | 2. | Tyvek Roll with Sterrad Chemical Indicator, 10'*228' | | | 2 Case | | | | | | |
| | 3. | Tyvek Roll with Sterrad Chemical Indicator, 20'*228' | | | 2 Case | | | | | | |
| | 4. | Sterrad Chemical Indicator Strip | | | 2 Case | | | | | | |
| | | | | (3) | | | _ | | | | |
| | (Signature of Authorized Person) | | | | | | | | | | |
| | | (Name) Name of Firm/Company/Agency | | | | | | | | | |
| | | | | | | | | | | | |
| | Phone No | | | | | | | | | | |
| | | Email: | | | | | | | | | |