



# All India Institute of Medical Sciences Jodhpur

Admn//Prop/06/2017-AIIMS.JDH

Dated: - 29<sup>th</sup> May, 2017.

**Subject:** Purchase of Pneumatic Drill Set for the department of Orthopedics at AIIMS, Jodhpur on proprietary basis - **Inviting comments thereon.**

The Institute is in the purchase of Pneumatic Drill Set for the department of Orthopedics at AIIMS, Jodhpur from M/s DePuy Synthes on proprietary basis. The proposal submitted by M/s Johnson & Johnson Pvt. Ltd and PAC certification by user are attached.

The above document are being uploaded for open information to submit subjection, comments, if any from any manufacturer regarding proprietary nature of the equipment within 21days of issue giving reference Admn/Prop/06/2017-AIIMS.JDH. The comments should be received by office of Administrative Officer, Medical College at AIIMS, Jodhpur on or before 23<sup>rd</sup> June, 2017 upto 03:00 PM failing which it will be presumed that any other vendor is having no comment to offer and case will be decided on merits.

**Yours faithfully,**

**Administrative Officer**

**Enclosed: Related documents enclosed.**



# All India Institute of Medical Sciences Jodhpur

DePuy Synthes


## Proprietary Certificate & Agreement for Representation

To whom it may concern:

We hereby certify that products attached in Annexure A, are proprietary products of Synthes GmbH Eimattstrasse 3, CH-4436 Oberdorf, Switzerland

We further certify that M/s Johnson & Johnson Private Limited, India is our affiliate and is authorized to market the products in India through its regional branch offices at Chennai, Delhi, Mumbai, and Kolkata. Johnson & Johnson Private Limited, India and its regional branch offices are authorized to submit offers and commit, sign, finalize, execute orders and also to appoint distributors, dealers and liaison agents for selling the complete range of products on our behalf.

On behalf of Synthes GmbH, Switzerland

  
Ariana Strähl  
Regulatory Affairs Specialist I  
RALI - Strategy Implementation

Stefanie Stark  
Regulatory Affairs Manager  
RALI Portfolio Leader EMEA



# All India Institute of Medical Sciences Jodhpur

DePuy Synthes

Annexure A

**Power Tools**

- Air Accessories
- APD – Air Pen Drive & Attachments
- Air Power Line II and Attachments
- Battery Power Line and Attachments– BPL /II
- Colibri II & Attachments
- CAD II - Compact Air Drive II & Attachments
- Cutting Tools - Saw Blades, Burrs and Drill Bits
- Electric Pen Drive & Consoles & Attachments
- Oscillating Saw
- TRS - Trauma Recon System & Attachments
- TRS Recon Saw
- Small Electric Drive System & Attachments
- UBC II - Universal Battery Charger II

*Handwritten signatures in blue ink.*

Synthes Contact  
Intelligence 3  
15100 Hillside Drive  
Warren, NJ 07059



# All India Institute of Medical Sciences Jodhpur

## Technical Specification Pneumatic Drill Set

S.No	Name of Items	Length Approx	QTY
1	<b>The Cannulated Pneumatic Drill handpiece - Compatible with existing attachments</b>		
	Cannulation with 3.2 mm diameter		
	Operating pressure : 6 - 7 bars (maximum 10 bars)		
	Weight of handpiece 600-800grams without any attachments		
	Power 120 w		
	Variable Speed from 0-900 rpm		
	Noise Level of max 75 db		
	Separate forward and reverse triggers		
	Safety Device to cut off air supply to drill on handpiece		
	Handpiece is compatible with radiolucent drive		
	Instant change between clockwise and counterclockwise rotation		
	Offers reliable protection of soft tissues with oscillating drill attachment		
	Fully Autoclavable		
	Fully machine washable		
	All Attachments can be fitted on single handpiece		
	The reverse trigger automatically locks when the oscillating saw and the reduction drive attachments are attached to handpiece		
2	<b>Adapter for Lubrication</b>		
	For oiling of hand piece		
	Autoclavable		
	Should be made of Stainless Steel		
3	<b>Double Air Hose</b>		
	Length 5 meters.		
	Autoclavable		
	Should have cocentric inlet and outlet pipes		
4	<b>Radiolucent Drive</b>		
	Precise aiming and drilling under image intensifier control for locking intramedullary nails		
	Drill Bit diameter 2.0 to 4.5 mm, Length 100 to 150 mm, Usable length 80 to 120 mm		
	Reduced exposure to x-rays		
5	<b>Jacob's Chuck attachment</b>		
	Chuck capacity up to 0 to 6.5 mm		
	Cannulation of 3.2 mm diameter		
	Maximum Speed of 900rpm		
	Torque of 4-5 Nm		
6	<b>Quick Coupling attachment</b>		



# All India Institute of Medical Sciences Jodhpur

	Cannulation 3.2 mm		
	Maximum Speed: 900 rpm		
	Torque of 4-5 Nm		
7	<b>Reduction Drive for Intramedullary / Acetabular Reaming with reverse option</b>		
	Reaming Speed of 300-350 rpm		
	Reaming Torque of 12-14 Nm		
	Option of attachment with reverse rotation		
8	<b>Quick Coupling for K-wire</b>		
	Continous adjustment facility for wire diameter from 0.6 to 3.2 mm		
	Speed up to 900 rpm		
9	<b>Oscillating Saw attachment with key</b>		
	It can operate on an oscillating frequency of 0 to 14,000 osc/min.		
	Attachment can be locked in 8 different positions		
10	<b>Quick Coupling for drill bits</b>		
	Speed: 0-900 rpm		
	Torque: 0-4.7 Nm		
	Cannulation: 1.3 mm		
11	<b>Quick coupling for DHS / DCS triple reamers</b>		
12	<b>Oscillating Saw Blades (All Sizes)</b>		
	For Trauma		6
	For Joint replacement		6
13	<b>Aluminium case for Pneumatic Drill system, Perforated, Autoclavable</b>		1
14	<b>Aluminium Box for accessory attachment, Perforated, Autoclavable</b>		1

*(Handwritten signatures)*